



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116735

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	10/07/2021	17:00
2	CAL2	1	Water	NA	NA	21.2	7.00	10/07/2021	17:01
3	CAL3	1	Water	NA	NA	21.2	10.03	10/07/2021	17:02
4	ICV	1	Water	NA	NA	21.2	7.02	10/07/2021	17:05
5	CCV1	1	Water	NA	NA	21.3	2.01	10/07/2021	17:06
6	M4144-01	1	Water	NA	NA	22.2	9.39	10/07/2021	17:20
7	M4144-01DUP	1	Water	NA	NA	22.1	9.40	10/07/2021	17:21
8	M4144-02	1	Water	NA	NA	22.3	9.78	10/07/2021	17:24
9	M4144-03	1	Water	NA	NA	21.9	11.79	10/07/2021	17:26
10	CCV2	1	Water	NA	NA	21.5	12.02	10/07/2021	17:30