



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116736

Slope : 98.5

BalanceID: WC SC-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	10/07/2021	17:00
2	CAL2	1	Water	NA	NA	21.2	7.00	10/07/2021	17:01
3	CAL3	1	Water	NA	NA	21.2	10.03	10/07/2021	17:02
4	ICV	1	Water	NA	NA	21.2	7.02	10/07/2021	17:05
5	CCV1	1	Water	NA	NA	21.3	2.01	10/07/2021	17:06
6	M4145-01	1	Solid	20.02	20	22.8	7.52	10/07/2021	17:35
7	M4145-03	1	Solid	20.03	20	22.1	9.53	10/07/2021	17:38
8	M4145-05	1	Solid	20.04	20	22.6	6.05	10/07/2021	17:42
9	M4145-05DUP	1	Solid	20.03	20	22.7	6.06	10/07/2021	17:43
10	CCV2	1	Water	NA	NA	21.7	12.02	10/07/2021	17:45

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY M4145

WorkList ID : 153057

Department : Wet-Chemistry

Date : 10-07-2021 16:32:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/07/2021	Solid	M4145-01	Corrosivity	Cool 4 deg C	JACO05	K11	IDW-SO-DRUM-3-100721	10/07/2021	9045D
10/07/2021	Solid	M4145-03	Corrosivity	Cool 4 deg C	JACO05	K11	IDW-SO-DRUM-7-100721	10/07/2021	9045D
10/07/2021	Solid	M4145-05	Corrosivity	Cool 4 deg C	JACO05	K11	IDW-SO-DRUM-5-100721	10/07/2021	9045D

Date/Time 10/07/21 16:35
 Raw Sample Received by: JP 1601
 Raw Sample Relinquished by: CP SM

Date/Time 10/07/21 18:25
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP 1601