


LB116738

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.5174e-004

C= 1.3993e-002

Rho= 0.9999329

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b
0.0	0.000	0.011	0.011	-11	0.000	-11					
0.2	0.200	0.192	-0.008	707	0.0 %	707					
2.5	2.500	2.504	0.004	9893	0.0 %	9893					
5.0	5.000	4.977	-0.023	19714	0.0 %	19714					
7.5	7.500	7.520	0.020	29816	0.0 %	29816					
10.0	10.000	9.996	-0.004	39650	0.0 %	39650					

LB116738 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 08 Oct 2021 10:22:50					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-11	-	0.0000	7470A	PPB	08 Oct 2021 10:39:37	S	Std
0.2 - 1	50.2	707	-	0.2000	7470A	PPB	08 Oct 2021 10:41:54	S	Std
2.5 - 1	52.5	9893	-	2.5000	7470A	PPB	08 Oct 2021 10:44:11	S	Std
5.0 - 1	55	19714	-	5.0000	7470A	PPB	08 Oct 2021 10:46:28	S	Std
7.5 - 1	57.5	29816	-	7.5000	7470A	PPB	08 Oct 2021 10:48:44	S	Std
10.0 - 1	50	39650	-	10.0000	7470A	PPB	08 Oct 2021 10:51:00	S	Std
ICV35 - 1	ICV35	15170	3.8329	-7470A	PPB	08 Oct 2021 10:53:39	U	SMPL	
ICB35 - 1	ICB35	-127	-0.0180	-7470A	PPB	08 Oct 2021 10:55:53	U	SMPL	
CCV21 - 1	CCV21	19715	4.9771	-7470A	PPB	08 Oct 2021 10:58:08	U	SMPL	
CCB21 - 1	CCB21	-142	-0.0218	-7470A	PPB	08 Oct 2021 11:00:23	U	SMPL	
CRA - 1	CRA	668	0.1822	-7470A	PPB	08 Oct 2021 11:02:38	U	SMPL	
HighStd - 1	HighStd	38893	9.8050	-7470A	PPB	08 Oct 2021 11:04:54	U	SMPL	
ChkStd - 1	ChkStd	27359	6.9014	-7470A	PPB	08 Oct 2021 11:07:09	U	SMPL	
PB139723BL - 1	PBW	-221	-0.0416	-7470A	PPB	08 Oct 2021 11:09:26	U	SMPL	
PB139723BS - 1	LCSW	14979	3.7848	-7470A	PPB	08 Oct 2021 11:11:43	U	SMPL	
M4072-02 - 1	VNJ-207	-147	-0.0230	-7470A	PPB	08 Oct 2021 11:13:59	U	SMPL	
M4087-10 - 1	WC-20211005-ROCK	-64	-0.0021	-7470A	PPB	08 Oct 2021 11:16:16	U	SMPL	
M4111-02 - 1	GROWS-1	-33	0.0057	-7470A	PPB	08 Oct 2021 11:18:33	U	SMPL	
M4111-05 - 1	GROWS-2	-24	0.0080	-7470A	PPB	08 Oct 2021 11:20:50	U	SMPL	
M4111-08 - 1	GROWS-3	-16	0.0100	-7470A	PPB	08 Oct 2021 11:23:05	U	SMPL	
CCV22 - 1	CCV22	19330	4.8802	-7470A	PPB	08 Oct 2021 11:25:20	U	SMPL	
CCB22 - 1	CCB22	-142	-0.0218	-7470A	PPB	08 Oct 2021 11:27:35	U	SMPL	
M4118-02 - 1	CONCRETE-100621-01	-40	0.0039	-7470A	PPB	08 Oct 2021 11:29:50	U	SMPL	
M4118-05 - 1	CONCRETE-100621-02	-26	0.0074	-7470A	PPB	08 Oct 2021 11:32:05	U	SMPL	
M4118-05DUP - 1	CONCRETE-100621-02DUP	-35	0.0052	-7470A	PPB	08 Oct 2021 11:34:21	U	SMPL	
M4118-05MS - 1	CONCRETE-100621-02MS	14008	3.5404	-7470A	PPB	08 Oct 2021 11:36:37	U	SMPL	
M4118-05MSD - 1	CONCRETE-100621-02MSD	15859	4.0064	-7470A	PPB	08 Oct 2021 11:38:53	U	SMPL	
PB139725BL - 1	PBW	-166	-0.0278	-7470A	PPB	08 Oct 2021 11:41:09	U	SMPL	
PB139725BS - 1	LCSW	15412	3.8938	-7470A	PPB	08 Oct 2021 11:43:25	U	SMPL	
M4096-01 - 1	MWHR2-8-100521	-155	-0.0250	-7470A	PPB	08 Oct 2021 11:45:42	U	SMPL	
M4096-03 - 1	MWHR3-8-100521	-50	0.0014	-7470A	PPB	08 Oct 2021 11:47:59	U	SMPL	
M4096-03DUP - 1	MWHR3-8-100521DUP	-37	0.0047	-7470A	PPB	08 Oct 2021 11:50:15	U	SMPL	
CCV23 - 1	CCV23	19147	4.8341	-7470A	PPB	08 Oct 2021 11:52:31	U	SMPL	
CCB23 - 1	CCB23	-164	-0.0273	-7470A	PPB	08 Oct 2021 11:54:46	U	SMPL	
M4096-04 - 1	M4096-03MS	15519	3.9208	-7470A	PPB	08 Oct 2021 11:57:02	U	SMPL	
M4096-05 - 1	M4096-03MSD	14577	3.6836	-7470A	PPB	08 Oct 2021 11:59:17	U	SMPL	
M4096-06 - 1	A3SEEP-100521	25722	6.4893	-7470A	PPB	08 Oct 2021 12:01:33	U	SMPL	
M4096-07 - 1	DUP	-146	-0.0228	-7470A	PPB	08 Oct 2021 12:03:49	U	SMPL	
M4130-01 - 1	P287-DITCH	227	0.0711	-7470A	PPB	08 Oct 2021 12:06:07	U	SMPL	
M4130-02 - 1	P287-DITCH	105	0.0404	-7470A	PPB	08 Oct 2021 12:08:23	U	SMPL	
M4130-03 - 1	SW-RR-03-(SA)	-24	0.0080	-7470A	PPB	08 Oct 2021 12:10:39	U	SMPL	
M4130-04 - 1	SW-RR-03-(SA)	-20	0.0090	-7470A	PPB	08 Oct 2021 12:12:56	U	SMPL	
M4130-05 - 1	SW-RR-01-(2.0)	-29	0.0067	-7470A	PPB	08 Oct 2021 12:15:13	U	SMPL	
M4130-06 - 1	SW-RR-01-(2.0)	-22	0.0085	-7470A	PPB	08 Oct 2021 12:17:29	U	SMPL	
CCV24 - 1	CCV24	19016	4.8011	-7470A	PPB	08 Oct 2021 12:19:45	U	SMPL	
CCB24 - 1	CCB24	-62	-0.0016	-7470A	PPB	08 Oct 2021 12:22:00	U	SMPL	
M4144-01 - 1	IDW-LIQ-DRUM-2-100721	891	0.2383	-7470A	PPB	08 Oct 2021 12:24:16	U	SMPL	
M4144-02 - 1	IDW-LIQ-DRUM-1-100721	1295	0.3400	-7470A	PPB	08 Oct 2021 12:26:31	U	SMPL	
M4144-03 - 1	IDW-LIQ-COMP-100721	0	0.0140	-7470A	PPB	08 Oct 2021 12:28:48	U	SMPL	
PB139723TB - 1	PB139672TB	-19	0.0092	-7470A	PPB	08 Oct 2021 12:31:04	U	SMPL	
M4118-05LX5 - 1		-25	0.0077	-7470A	PPB	08 Oct 2021 12:33:21	U	SMPL	
M4118-05A - 1		15069	3.8075	-7470A	PPB	08 Oct 2021 12:35:38	U	SMPL	
M4096-03LX5 - 1		63	0.0299	-7470A	PPB	08 Oct 2021 12:37:55	U	SMPL	
M4096-03A - 1		14840	3.7498	-7470A	PPB	08 Oct 2021 12:40:12	U	SMPL	
CCV25 - 1	CCV25	19034	4.8056	-7470A	PPB	08 Oct 2021 12:42:29	U	SMPL	
CCB25 - 1	CCB25	-113	-0.0145	-7470A	PPB	08 Oct 2021 12:44:45	U	SMPL	