

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : mb Instrument ID : Konelab

10/8/2021 13:36

Test: Phenolics- mb

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.060	0.0	0.107	
ICB1	0.003	0.0	0.006	
CCV1	1.065	0.0	0.107	
CCB1	0.001	0.0	0.006	
PB139755BL	0.001	0.0	0.006	
PB139755BS	1.066	0.0	0.107	
M4068-07	0.028	0.0	0.008	
M4068-08	0.057	0.0	0.011	
PB139756BL	0.000	0.0	0.006	
PB139756BS	1.062	0.0	0.107	
M4068-01	0.029	0.0	0.008	
M4068-02	0.058	0.0	0.011	
CCV2	1.059	0.0	0.107	
CCB2	0.000	0.0	0.006	

N	14
Mean	0.392
SD	0.5189
CV%	132.34

Aquakem v. 7.2AQ1

Results from time period:

Fri Oct 08 13:20:18 2021

Fri Oct 08 13:35:17 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.0011	mg/l	10/8/2021 10:49:33	
0.05PPM	A	Phenolics-	P	0.0511	mg/l	10/8/2021 10:49:34	
0.1PPM	A	Phenolics-	P	0.1027	mg/l	10/8/2021 10:49:35	
0.25PPM	A	Phenolics-	P	0.2512	mg/l	10/8/2021 10:49:36	
0.50PPM	A	Phenolics-	P	0.5001	mg/l	10/8/2021 10:49:37	
1.0PPM	A	Phenolics-	P	0.9927	mg/l	10/8/2021 10:49:38	
2.0PPM	A	Phenolics-	P	2.0033	mg/l	10/8/2021 10:49:39	
ICV1	S	Phenolics-	P	1.0596	mg/l	10/8/2021 13:20:19	
ICB1	S	Phenolics-	P	0.0028	mg/l	10/8/2021 13:20:20	
CCV1	S	Phenolics-	P	1.065	mg/l	10/8/2021 13:20:23	
CCB1	S	Phenolics-	P	0.0008	mg/l	10/8/2021 13:20:24	
PB139755BL	S	Phenolics-	P	0.001	mg/l	10/8/2021 13:20:26	
PB139755BS	S	Phenolics-	P	1.0662	mg/l	10/8/2021 13:20:28	
M4068-07	S	Phenolics-	P	0.0279	mg/l	10/8/2021 13:30:20	
M4068-08	S	Phenolics-	P	0.0567	mg/l	10/8/2021 13:30:23	
PB139756BL	S	Phenolics-	P	0.0004	mg/l	10/8/2021 13:30:25	
PB139756BS	S	Phenolics-	P	1.0625	mg/l	10/8/2021 13:30:27	
M4068-01	S	Phenolics-	P	0.0289	mg/l	10/8/2021 13:30:29	
M4068-02	S	Phenolics-	P	0.0582	mg/l	10/8/2021 13:30:30	
CCV2	S	Phenolics-	P	1.0592	mg/l	10/8/2021 13:35:15	
CCB2	S	Phenolics-	P	0.0002	mg/l	10/8/2021 13:35:16	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AS Instrument ID : Konelab

10/8/2021 10:50


Test Phenolics-

Accepted 10/8/2021 10:50

Factor 10.49
Bias 0.006

Coeff. of det. 0.999976

Errors



	Calibrator	Response	Calc. con.	Conc.	Rel. Errors
1	0.00PPM	0.005	-0.0011	0.0000	2.2
2	Phenol-2	0.010	0.0511	0.0500	7.2
3	Phenol-2	0.015	0.1027	0.1000	0.5
4	Phenol-2	0.030	0.2512	0.2500	0.0
5	Phenol-2	0.053	0.5001	0.5000	0.7
6	Phenol-2	0.100	0.9927	1.0000	0.2
7	Phenol-2	0.197	2.0033	2.0000	

AS
10/8/21