



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116753

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	79.9	10/07/2021	16:15
2	M4047-06		1	10	100.00	>212.0	>212.0	10/07/2021	16:40
3	M4047-06DUP		1	10	100.00	>212.0	>212.0	10/07/2021	17:05
4	M4049-01		1	12	100.00	>212.0	>212.0	10/07/2021	17:25
5	M4059-01		1	15	100.00	>212.0	>212.0	10/07/2021	17:45
6	M4088-02		1	12	100.00	>212.0	>212.0	10/07/2021	18:05
7	M4104-01		1	14	100.00	>212.0	>212.0	10/07/2021	18:25
8	M4142-05		1	15	100.00	>212.0	>212.0	10/07/2021	18:40
9	M4144-01		1	12	100.00	>212.0	>212.0	10/07/2021	19:05
10	M4144-02		1	15	100.00	>212.0	>212.0	10/07/2021	19:20
11	M4144-03		1	14	100.00	>212.0	>212.0	10/07/2021	19:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 116753

WorkList Name : FP10421 WorkList ID : 152881 Department : Wet-Chemistry Date : 10-04-2021 10:48:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/15/2021	Water	M4047-06	Flash Point	Cool 4 deg C	JPCL01	J61	MW-1	10/01/2021	1010B
10/15/2021	Water	M4049-01	Flash Point	Cool 4 deg C	JPCL01		MW-1	10/01/2021	1010B
10/18/2021	Water	M4059-01	Flash Point	Cool 4 deg C	PSEG03	J61	SH-SUMP	10/04/2021	1010B
10/12/2021	Water	M4088-02	Flash Point	Cool 4 deg C	PARS02		WC-20211005-WATER	10/05/2021	1010B
10/19/2021	Water	M4104-01	Flash Point	Cool 4 deg C	DIST01	K11	55TH-ST-33	10/06/2021	1010B
10/14/2021	Water	M4142-05	Flash Point	Cool 4 deg C	PSEG03	K11	2087	10/07/2021	1010B
10/11/2021	Water	M4144-01	Flash Point	Cool 4 deg C	JACO05	K11	IDW-LIQ-DRUM-2-10072	10/07/2021	1010B
10/11/2021	Water	M4144-02	Flash Point	Cool 4 deg C	JACO05	K11	IDW-LIQ-DRUM-1-10072	10/07/2021	1010B
10/11/2021	Water	M4144-03	Flash Point	Cool 4 deg C	JACO05	K11	IDW-LIQ-COMP-100721	10/07/2021	1010B

Date/Time 10/7/21 15:55
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ C Ext-1060

Date/Time 10/7/21 19:45
 Raw Sample Received by: RJ C Ext-1060
 Raw Sample Relinquished by: KS (WC)