

Analytical Summary Report

Analysis Method: 9040C

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116759

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/08/2021	15:25
2	CAL2	1	Water	NA	NA	20.3	7.00	10/08/2021	15:26
3	CAL3	1	Water	NA	NA	20.2	10.02	10/08/2021	15:27
4	ICV	1	Water	NA	NA	20.3	7.01	10/08/2021	15:29
5	CCV1	1	Water	NA	NA	20.4	2.00	10/08/2021	15:30
6	M4149-01	1	Water	NA	NA	21.3	4.89	10/08/2021	15:31
7	M4149-02	1	Water	NA	NA	21.3	4.92	10/08/2021	15:32
8	M4151-01	1	Water	NA	NA	22.2	5.80	10/08/2021	15:33
9	M4151-02	1	Water	NA	NA	22.0	5.40	10/08/2021	15:34
10	M4151-03	1	Water	NA	NA	21.9	5.80	10/08/2021	15:35
11	M4151-04	1	Water	NA	NA	22.0	5.92	10/08/2021	15:36
12	M4151-05	1	Water	NA	NA	22.0	5.95	10/08/2021	15:37
13	M4151-06	1	Water	NA	NA	22.1	5.88	10/08/2021	15:38
14	M4151-07	1	Water	NA	NA	22.2	5.86	10/08/2021	15:39
15	M4151-08	1	Water	NA	NA	22.2	5.83	10/08/2021	15:40
16	CCV2	1	Water	NA	NA	20.3	12.02	10/08/2021	15:41
17	M4156-01	1	Water	NA	NA	20.2	7.58	10/08/2021	15:42
18	M4157-01	1	Water	NA	NA	20.4	7.15	10/08/2021	15:45
19	M4157-01DUP	1	Water	NA	NA	20.4	7.15	10/08/2021	15:46
20	M4162-01	1	Water	NA	NA	21.5	7.36	10/08/2021	15:47

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	M4162-03	1	Water	NA	NA	21.4	7.16	10/08/2021	15:48
22	CCV3	1	Water	NA	NA	20.3	2.01	10/08/2021	15:50

LB 116759

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M4147 WorkList ID : 153100 Department : Wet-Chemistry Date : 10-08-2021 15:14:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/07/2021	Water	M4149-01	pH	Cool 4 deg C	STER16	D11	MAIN-CELL	10/07/2021	9040C
10/07/2021	Water	M4149-02	pH	Cool 4 deg C	STER16	D11	EXCELL	10/07/2021	9040C
10/08/2021	Water	M4151-01	pH	Cool 4 deg C	PSEG03	K11	33481	10/08/2021	9040C
10/08/2021	Water	M4151-02	pH	Cool 4 deg C	PSEG03	K11	33482	10/08/2021	9040C
10/08/2021	Water	M4151-03	pH	Cool 4 deg C	PSEG03	K11	33483	10/08/2021	9040C
10/08/2021	Water	M4151-04	pH	Cool 4 deg C	PSEG03	K11	33484	10/08/2021	9040C
10/08/2021	Water	M4151-05	pH	Cool 4 deg C	PSEG03	K11	33485	10/08/2021	9040C
10/08/2021	Water	M4151-06	pH	Cool 4 deg C	PSEG03	K11	33486	10/08/2021	9040C
10/08/2021	Water	M4151-07	pH	Cool 4 deg C	PSEG03	K11	33487	10/08/2021	9040C
10/08/2021	Water	M4151-08	pH	Cool 4 deg C	PSEG03	K11	33488	10/08/2021	9040C
10/08/2021	Water	M4156-01	pH	Cool 4 deg C	PARS02	K12	20-2011008-WC	10/08/2021	9040C
10/08/2021	Water	M4157-01	pH	Cool 4 deg C	PARS02	K12	FDR-20211008-WC	10/08/2021	9040C
10/08/2021	Water	M4162-01	pH	Cool 4 deg C	PSEG03	K12	33477	10/08/2021	9040C
10/08/2021	Water	M4162-03	pH	Cool 4 deg C	PSEG03	K12	33478	10/08/2021	9040C

Date/Time 10/08/21 15:20
Raw Sample Received by: KS(WC)
Raw Sample Relinquished by: JH(WC)

Date/Time 10/08/21 15:53
Raw Sample Received by: JH(WC)
Raw Sample Relinquished by: KS(WC)