



Analytical Summary Report

Analysis Method: 9040C

Analyst By : ketankumar

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116763

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/08/2021	15:25
2	CAL2	1	Water	NA	NA	20.3	7.00	10/08/2021	15:26
3	CAL3	1	Water	NA	NA	20.2	10.02	10/08/2021	15:27
4	ICV	1	Water	NA	NA	20.3	7.01	10/08/2021	15:29
5	CCV1	1	Water	NA	NA	20.4	2.00	10/08/2021	15:30
6	M4156-02	1	Water	NA	NA	20.2	7.58	10/08/2021	15:56
7	M4157-02	1	Water	NA	NA	20.4	7.15	10/08/2021	15:57
8	M4157-02DUP	1	Water	NA	NA	20.4	7.15	10/08/2021	15:58
9	CCV2	1	Water	NA	NA	20.4	12.02	10/08/2021	15:59

LB 116763

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M4147 WorkList ID : 153101 Department : Wet-Chemistry Date : 10-08-2021 15:15:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/08/2021	Water	M4156-02	Corrosivity	Cool 4 deg C	PARS02	K12	20-2011008-WC	10/08/2021	9040C
10/08/2021	Water	M4157-02	Corrosivity	Cool 4 deg C	PARS02	K12	FDR-20211008-WC	10/08/2021	9040C

Date/Time 10/08/21 15:20
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: SA (WC)

Date/Time 10/08/21 16:00
Raw Sample Received by: SA (WC)
Raw Sample Relinquished by: KS (WC)