



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116764

Slope : 98.6

BalanceID:

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/08/2021	15:25
2	CAL2	1	Water	NA	NA	20.3	7.00	10/08/2021	15:26
3	CAL3	1	Water	NA	NA	20.2	10.02	10/08/2021	15:27
4	ICV	1	Water	NA	NA	20.3	7.01	10/08/2021	15:29
5	CCV1	1	Water	NA	NA	20.3	2.00	10/08/2021	16:02
6	M4147-02	1	Solid	20	20	23.0	9.76	10/08/2021	16:03
7	M4147-02DUP	1	Solid	20	20	23.0	9.75	10/08/2021	16:04
8	M4150-02	1	Solid	20	20	23.2	9.05	10/08/2021	16:05
9	CCV2	1	Water	NA	NA	20.3	12.02	10/08/2021	16:06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M4158 S WorkList ID : 153103 Department : Wet-Chemistry Date : 10-08-2021 15:16:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/07/2021	Solid	M4147-02	Corrosivity	Cool 4 deg C	PARS02	K11	20-20211007-WC	10/07/2021	9045D
10/07/2021	Solid	M4150-02	Corrosivity	Cool 4 deg C	PARS02	K11	FDR-20211007-WC	10/07/2021	9045D

Date/Time 10/08/21 15:20
 Raw Sample Received by: KS(WC)
 Raw Sample Relinquished by: JA(WC)

Date/Time 10/08/21 16:20
 Raw Sample Received by: JA(WC)
 Raw Sample Relinquished by: KS(WC)

LB 116764