

LB116780


7470A

INSTRUMENT ID: CVZ

Linear

 μ Abs.:

36615



A= 0.0000e+000

B= 2.6923e-004

C= -1.8891e-002

Rho= 0.9995562

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.027	-0.027	-29	0.000	-29					
0.2	0.200	0.178	-0.022	730	0.0 %	730					-11
2.5	2.500	2.515	0.015	9412	0.0 %	9412					1
5.0	5.000	4.983	-0.017	18579	0.0 %	18579					0
7.5	7.500	7.712	0.212	28714	0.0 %	28714					3
10.0	10.000	9.839	-0.161	36615	0.0 %	36615					-2

LB116780 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 11 Oct 2021 14:56:49							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-29	-	0.0000	7470A	PPB	11 Oct 2021 14:57:37	S	Std
0.2 - 1	50.2	730	-	0.2000	7470A	PPB	11 Oct 2021 14:59:54	S	Std
2.5 - 1	52.5	9412	-	2.5000	7470A	PPB	11 Oct 2021 15:02:11	S	Std
5.0 - 1	55	18579	-	5.0000	7470A	PPB	11 Oct 2021 15:04:28	S	Std
7.5 - 1	57.5	28714	-	7.5000	7470A	PPB	11 Oct 2021 15:06:44	S	Std
10.0 - 1	59	36615	-	10.0000	7470A	PPB	11 Oct 2021 15:09:00	S	Std
ICV37 - 1	ICV37	15966	4.2796	-	7470A	PPB	11 Oct 2021 15:11:44	U	SMPL
ICB37 - 1	ICB37	11	-0.0159	-	7470A	PPB	11 Oct 2021 15:16:31	U	SMPL
CCV29 - 1	CCV29	18373	4.9277	-	7470A	PPB	11 Oct 2021 15:21:44	U	SMPL
CCB29 - 1	CCB29	-110	-0.0485	-	7470A	PPB	11 Oct 2021 15:23:59	U	SMPL
CRA - 1	CRA	695	0.1682	-	7470A	PPB	11 Oct 2021 15:26:13	U	SMPL
HighStd - 1	HighStd	36852	9.9028	-	7470A	PPB	11 Oct 2021 15:28:28	U	SMPL
ChkStd - 1	ChkStd	25023	6.7181	-	7470A	PPB	11 Oct 2021 15:30:43	U	SMPL
PB139763BL - 1	PBW	-88	-0.0426	-	7470A	PPB	11 Oct 2021 15:33:00	U	SMPL
PB139763BS - 1	LCSW	15255	4.0882	-	7470A	PPB	11 Oct 2021 15:38:16	U	SMPL
M4127-07 - 1	CHRT-20426	-150	-0.0593	-	7470A	PPB	11 Oct 2021 15:40:32	U	SMPL
M4136-02 - 1	PSC124055	-44	-0.0307	-	7470A	PPB	11 Oct 2021 15:42:48	U	SMPL
M4136-04 - 1	PSC124027	-33	-0.0278	-	7470A	PPB	11 Oct 2021 15:45:05	U	SMPL
M4136-06 - 1	PSC124042	-1	-0.0192	-	7470A	PPB	11 Oct 2021 15:47:23	U	SMPL
M4145-02 - 1	IDW-SO-DRUM-3-100721	-12	-0.0221	-	7470A	PPB	11 Oct 2021 15:49:39	U	SMPL
CCV30 - 1	CCV30	19382	5.1993	-	7470A	PPB	11 Oct 2021 15:51:54	U	SMPL
CCB30 - 1	CCB30	2	-0.0184	-	7470A	PPB	11 Oct 2021 15:56:40	U	SMPL
M4145-04 - 1	IDW-SO-DRUM-7-100721	-9	-0.0213	-	7470A	PPB	11 Oct 2021 15:58:55	U	SMPL
M4145-06 - 1	IDW-SO-DRUM-5-100721	-5	-0.0202	-	7470A	PPB	11 Oct 2021 16:01:10	U	SMPL
M4145-06DUP - 1	IDW-SO-DRUM-5-100721DUP	-5	-0.0202	-	7470A	PPB	11 Oct 2021 16:03:26	U	SMPL
M4145-06MS - 1	IDW-SO-DRUM-5-100721MS	16181	4.3375	-	7470A	PPB	11 Oct 2021 16:05:42	U	SMPL
M4145-06MSD - 1	IDW-SO-DRUM-5-100721MSD	15517	4.1588	-	7470A	PPB	11 Oct 2021 16:07:57	U	SMPL
PB139765BL - 1	PBW	-44	-0.0307	-	7470A	PPB	11 Oct 2021 16:12:51	U	SMPL
PB139765BS - 1	LCSW	14903	3.9935	-	7470A	PPB	11 Oct 2021 16:21:26	U	SMPL
M4154-01 - 1	SW-CB-31	-136	-0.0555	-	7470A	PPB	11 Oct 2021 16:23:42	U	SMPL
M4154-01DUP - 1	SW-CB-31DUP	-36	-0.0286	-	7470A	PPB	11 Oct 2021 16:26:00	U	SMPL
M4154-01MS - 1	SW-CB-31MS	14964	4.0099	-	7470A	PPB	11 Oct 2021 16:28:16	U	SMPL
CCV31 - 1	CCV31	18420	4.9403	-	7470A	PPB	11 Oct 2021 16:30:31	U	SMPL
CCB31 - 1	CCB31	-20	-0.0243	-	7470A	PPB	11 Oct 2021 16:32:47	U	SMPL
M4154-01MSD - 1	SW-CB-31MSD	13990	3.7476	-	7470A	PPB	11 Oct 2021 16:35:02	U	SMPL
M4154-02 - 1	SW-CB-31	-127	-0.0531	-	7470A	PPB	11 Oct 2021 16:37:18	U	SMPL
M4154-05 - 1	SW-CB-32	2712	0.7113	-	7470A	PPB	11 Oct 2021 16:39:33	U	SMPL
M4154-06 - 1	SW-CB-32	648	0.1556	-	7470A	PPB	11 Oct 2021 16:41:48	U	SMPL
PB139763TB - 1	PB139730TB	-30	-0.0270	-	7470A	PPB	11 Oct 2021 16:44:05	U	SMPL
M4145-06LX5 - 1		-19	-0.0240	-	7470A	PPB	11 Oct 2021 16:46:21	U	SMPL
M4145-06A - 1		15276	4.0939	-	7470A	PPB	11 Oct 2021 16:48:37	U	SMPL
M4154-01LX5 - 1		-90	-0.0431	-	7470A	PPB	11 Oct 2021 16:50:54	U	SMPL
M4154-01A - 1		15175	4.0667	-	7470A	PPB	11 Oct 2021 16:53:11	U	SMPL
CCV32 - 1	CCV32	18940	5.0803	-	7470A	PPB	11 Oct 2021 16:55:27	U	SMPL
CCB32 - 1	CCB32	-75	-0.0391	-	7470A	PPB	11 Oct 2021 16:57:43	U	SMPL