

LB116790


7470A

INSTRUMENT ID: CYJ

Linear

 μ Abs.:

35150



A= 0.0000e+000

B= 2.8270e-004

C= -2.7948e-003

Rho= 0.9998238

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	-15	0.000	-15				
0.2	0.200	0.182	-0.018	652	0.0 %	652				
2.5	2.500	2.439	-0.061	8638	0.0 %	8638				
5.0	5.000	5.142	0.142	18200	0.0 %	18200				
7.5	7.500	7.510	0.010	26575	0.0 %	26575				
10.0	10.000	9.934	-0.066	35150	0.0 %	35150				

LB116790 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 12 Oct 2021 07:54:07					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-15	-	0.0000	7470A	PPB	12 Oct 2021 07:59:04	S	Std
0.2 - 1	50.2	652	-	0.2000	7470A	PPB	12 Oct 2021 08:01:21	S	Std
2.5 - 1	52.5	8638	-	2.5000	7470A	PPB	12 Oct 2021 08:03:38	S	Std
5.0 - 1	55	18200	-	5.0000	7470A	PPB	12 Oct 2021 08:08:29	S	Std
7.5 - 1	57.5	26575	-	7.5000	7470A	PPB	12 Oct 2021 08:13:18	S	Std
10.0 - 1	510	35150	-	10.0000	7470A	PPB	12 Oct 2021 08:19:36	S	Std
ICV38 - 1	ICV38	15174	4.2869	-	7470A	PPB	12 Oct 2021 08:22:20	U	SMPL
ICB38 - 1	ICB38	-194	-0.0576	-	7470A	PPB	12 Oct 2021 08:24:36	U	SMPL
CCV33 - 1	CCV33	18308	5.1729	-	7470A	PPB	12 Oct 2021 08:26:51	U	SMPL
CCB33 - 1	CCB33	-168	-0.0503	-	7470A	PPB	12 Oct 2021 08:29:07	U	SMPL
CRA - 1	CRA	578	0.1606	-	7470A	PPB	12 Oct 2021 08:31:22	U	SMPL
HighStd - 1	HighStd	35041	9.9033	-	7470A	PPB	12 Oct 2021 08:33:37	U	SMPL
ChkStd - 1	ChkStd	24544	6.9358	-	7470A	PPB	12 Oct 2021 08:35:52	U	SMPL
PB139764BL - 1	PBW	-199	-0.0591	-	7470A	PPB	12 Oct 2021 08:38:09	U	SMPL
PB139764BS - 1	LCSW	14605	4.1260	-	7470A	PPB	12 Oct 2021 08:40:26	U	SMPL
M4142-04 - 1	2186-2187	-161	-0.0483	-	7470A	PPB	12 Oct 2021 08:42:43	U	SMPL
M4142-04DUP - 1	2186-2187DUP	-62	-0.0203	-	7470A	PPB	12 Oct 2021 08:45:00	U	SMPL
M4142-04MS - 1	2186-2187MS	14641	4.1362	-	7470A	PPB	12 Oct 2021 08:47:17	U	SMPL
M4142-04MSD - 1	2186-2187MSD	15425	4.3578	-	7470A	PPB	12 Oct 2021 08:49:34	U	SMPL
M4142-05 - 1	2087	-139	-0.0421	-	7470A	PPB	12 Oct 2021 08:51:49	U	SMPL
CCV34 - 1	CCV34	18318	5.1757	-	7470A	PPB	12 Oct 2021 08:54:05	U	SMPL
CCB34 - 1	CCB34	-148	-0.0446	-	7470A	PPB	12 Oct 2021 08:56:21	U	SMPL
M4142-07 - 1	2158	-46	-0.0158	-	7470A	PPB	12 Oct 2021 08:58:38	U	SMPL
PB139764TB - 1	PB139749TB	-16	-0.0073	-	7470A	PPB	12 Oct 2021 09:00:53	U	SMPL
M4142-04LX5 - 1		-6	-0.0045	-	7470A	PPB	12 Oct 2021 09:03:08	U	SMPL
M4142-04A - 1		15025	4.2448	-	7470A	PPB	12 Oct 2021 09:05:24	U	SMPL
CCV35 - 1	CCV35	18246	5.1553	-	7470A	PPB	12 Oct 2021 09:07:40	U	SMPL
CCB35 - 1	CCB35	-122	-0.0373	-	7470A	PPB	12 Oct 2021 09:09:56	U	SMPL