



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116791

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/12/2021	10:20
2	M4151-09		1	10	100.00	>212.0	>212.0	10/12/2021	10:45
3	M4151-09DUP		1	10	100.00	>212.0	>212.0	10/12/2021	11:10
4	M4151-11		1	12	100.00	>212.0	>212.0	10/12/2021	11:35
5	M4156-02		1	15	100.00	>212.0	>212.0	10/12/2021	12:05
6	M4157-02		1	14	100.00	>212.0	>212.0	10/12/2021	12:32
7	M4162-01		1	15	100.00	>212.0	>212.0	10/12/2021	12:58
8	M4162-03		1	13	100.00	>212.0	>212.0	10/12/2021	13:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP101121 WorkList ID : 153166 Department : Wet-Chemistry Date : 10-11-2021 17:12:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/15/2021	Water	M4151-09	Flash Point	Cool 4 deg C	PSEG03	K11	81-85	10/08/2021	1010B
10/15/2021	Water	M4151-11	Flash Point	Cool 4 deg C	PSEG03	K11	86-88	10/08/2021	1010B
10/15/2021	Water	M4156-02	Flash Point	Cool 4 deg C	PARS02	K12	20-20211008-WC	10/08/2021	1010B
10/15/2021	Water	M4157-02	Flash Point	Cool 4 deg C	PARS02	K12	FDR-20211008-WC	10/08/2021	1010B
10/15/2021	Water	M4162-01	Flash Point	Cool 4 deg C	PSEG03	K12	33477	10/08/2021	1010B
10/15/2021	Water	M4162-03	Flash Point	Cool 4 deg C	PSEG03	K12	33478	10/08/2021	1010B

Date/Time 10/12/21 08:45
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ (WC-196)

Date/Time 10/12/21 13:30
 Raw Sample Received by: RJ (WC-196)
 Raw Sample Relinquished by: KS (WC)