

LB116795


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

39839



A= 0.0000e+000

B= 2.5328e-004

C= 2.6549e-002

Rho= 0.9996735

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.025	0.025	-8	0.000	-8					-
0.2	0.200	0.233	0.033	815	0.0 %	815					17
2.5	2.500	2.567	0.067	10032	0.0 %	10032					3
5.0	5.000	4.834	-0.166	18979	0.0 %	18979					-3
7.5	7.500	7.424	-0.076	29208	0.0 %	29208					-1
10.0	10.000	10.117	0.117	39839	0.0 %	39839					1

LB116795 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 12 Oct 2021 10:42:04					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-8	-	0.0000	7470A	PPB	12 Oct 2021 10:43:51	S	Std
0.2 - 1	502	815	-	0.2000	7470A	PPB	12 Oct 2021 10:46:08	S	Std
2.5 - 1	52.5	10032	-	2.5000	7470A	PPB	12 Oct 2021 10:48:25	S	Std
5.0 - 1	55	18979	-	5.0000	7470A	PPB	12 Oct 2021 10:50:42	S	Std
7.5 - 1	57.5	29208	-	7.5000	7470A	PPB	12 Oct 2021 10:52:58	S	Std
10.0 - 1	510	39839	-	10.0000	7470A	PPB	12 Oct 2021 11:00:26	S	Std
ICV39 - 1	ICV39	14986	3.8222	-	7470A	PPB	12 Oct 2021 11:05:38	U	SMPL
ICB39 - 1	ICB39	-150	-0.0114	-	7470A	PPB	12 Oct 2021 11:07:53	U	SMPL
CCV36 - 1	CCV36	18705	4.7642	-	7470A	PPB	12 Oct 2021 11:10:08	U	SMPL
CCB36 - 1	CCB36	-165	-0.0152	-	7470A	PPB	12 Oct 2021 11:12:23	U	SMPL
CRA - 1	CRA	758	0.2185	-	7470A	PPB	12 Oct 2021 11:14:37	U	SMPL
HighStd - 1	HighStd	39794	10.1056	-	7470A	PPB	12 Oct 2021 11:16:52	U	SMPL
ChkStd - 1	ChkStd	27258	6.9305	-	7470A	PPB	12 Oct 2021 11:19:08	U	SMPL
PB139834BL - 1	PBW	-215	-0.0279	-	7470A	PPB	12 Oct 2021 11:21:25	U	SMPL
PB139834BS - 1	LCSW	14521	3.7045	-	7470A	PPB	12 Oct 2021 11:23:41	U	SMPL
M4156-01 - 1	20-20211008-WC	17513	4.4623	-	7470A	PPB	12 Oct 2021 11:25:57	U	SMPL
M4157-01 - 1	FDR-20211008-WC	834	0.2378	-	7470A	PPB	12 Oct 2021 11:28:13	U	SMPL
M4169-01 - 1	P287-DITCH	487	0.1499	-	7470A	PPB	12 Oct 2021 11:35:20	U	SMPL
M4169-01DUP - 1	P287-DITCHDUP	368	0.1198	-	7470A	PPB	12 Oct 2021 11:37:36	U	SMPL
M4169-01MS - 1	P287-DITCHMS	14528	3.7062	-	7470A	PPB	12 Oct 2021 11:39:51	U	SMPL
CCV37 - 1	CCV37	18324	4.6677	-	7470A	PPB	12 Oct 2021 11:42:07	U	SMPL
CCB37 - 1	CCB37	-85	0.0050	-	7470A	PPB	12 Oct 2021 11:44:23	U	SMPL
M4169-01MSD - 1	P287-DITCHMSD	14627	3.7313	-	7470A	PPB	12 Oct 2021 11:46:38	U	SMPL
M4169-02 - 1	P287-DITCH	143	0.0628	-	7470A	PPB	12 Oct 2021 11:48:55	U	SMPL
M4169-03 - 1	SW-RR-03-(SA)	13	0.0298	-	7470A	PPB	12 Oct 2021 11:51:11	U	SMPL
M4169-04 - 1	SW-RR-03-(SA)	-24	0.0205	-	7470A	PPB	12 Oct 2021 11:53:27	U	SMPL
M4169-05 - 1	SW-RR-01-(2.0)	-22	0.0210	-	7470A	PPB	12 Oct 2021 11:55:43	U	SMPL
M4169-06 - 1	SW-RR-01-(2.0)	-19	0.0217	-	7470A	PPB	12 Oct 2021 11:58:00	U	SMPL
M4170-01 - 1	P287-DITCH	315	0.1063	-	7470A	PPB	12 Oct 2021 12:00:16	U	SMPL
M4170-02 - 1	P287-DITCH	71	0.0445	-	7470A	PPB	12 Oct 2021 12:02:33	U	SMPL
M4170-03 - 1	SW-RR-03-(SA)	-7	0.0248	-	7470A	PPB	12 Oct 2021 12:04:50	U	SMPL
M4170-04 - 1	SW-RR-03-(SA)	-23	0.0207	-	7470A	PPB	12 Oct 2021 12:07:05	U	SMPL
CCV38 - 1	CCV38	19241	4.8999	-	7470A	PPB	12 Oct 2021 12:09:21	U	SMPL
CCB38 - 1	CCB38	-73	0.0081	-	7470A	PPB	12 Oct 2021 12:11:36	U	SMPL
M4170-05 - 1	SW-RR-01-(2.0)	-33	0.0182	-	7470A	PPB	12 Oct 2021 12:13:52	U	SMPL
M4170-06 - 1	SW-RR-01-(2.0)	-27	0.0197	-	7470A	PPB	12 Oct 2021 12:16:07	U	SMPL
PB139835BL - 1	PBW	49	0.0390	-	7470A	PPB	12 Oct 2021 12:18:23	U	SMPL
PB139835BS - 1	LCSW	14751	3.7627	-	7470A	PPB	12 Oct 2021 12:20:39	U	SMPL
M4156-02 - 1	20-20211008-WC	-54	0.0129	-	7470A	PPB	12 Oct 2021 12:22:54	U	SMPL
M4157-02 - 1	FDR-20211008-WC	-18	0.0220	-	7470A	PPB	12 Oct 2021 12:25:11	U	SMPL
M4157-02DUP - 1	FDR-20211008-WCDUP	-8	0.0245	-	7470A	PPB	12 Oct 2021 12:27:27	U	SMPL
M4157-02MS - 1	FDR-20211008-WCMS	15521	3.9577	-	7470A	PPB	12 Oct 2021 12:29:44	U	SMPL
M4157-02MSD - 1	FDR-20211008-WCMSD	15102	3.8516	-	7470A	PPB	12 Oct 2021 12:32:01	U	SMPL
CCV39 - 1	CCV39	19654	5.0045	-	7470A	PPB	12 Oct 2021 12:34:17	U	SMPL
CCB39 - 1	CCB39	-146	-0.0104	-	7470A	PPB	12 Oct 2021 12:36:32	U	SMPL
M4169-01LX5 - 1		102	0.0524	-	7470A	PPB	12 Oct 2021 12:38:48	U	SMPL
M4169-01A - 1		14147	3.6097	-	7470A	PPB	12 Oct 2021 12:41:04	U	SMPL
M4157-02LX5 - 1		-85	0.0050	-	7470A	PPB	12 Oct 2021 12:43:19	U	SMPL
M4157-02A - 1		15429	3.9344	-	7470A	PPB	12 Oct 2021 12:45:35	U	SMPL
PB139835TB - 1	PB139749TB	-90	0.0038	-	7470A	PPB	12 Oct 2021 12:47:51	U	SMPL
CCV40 - 1	CCV40	19447	4.9521	-	7470A	PPB	12 Oct 2021 12:50:06	U	SMPL
CCB40 - 1	CCB40	-142	-0.0094	-	7470A	PPB	12 Oct 2021 12:52:23	U	SMPL