


Batch Folder: D:\Agilent\ICPMH\1\DATA\P8101221 MS.b\
Analysis File: P8101221 MS.batch.bin
DA Date-Time: 2021-10-12 11:56:16
Calibration Title:
Calibration Method: External Calibration
VIS Interpolation Fit:

Level	Standard Data File	Sample Name	Acq. Date-Time
1	005CALB.d	S00	2021-10-12 10:55:58
2	007CAL.S.d	S02	2021-10-12 11:02:07
3	008CAL.S.d	S03	2021-10-12 11:05:13
4	009CAL.S.d	S04	2021-10-12 11:08:16
5	010CAL.S.d	S05	2021-10-12 11:11:22
6	013CAL.S.d	S06	2021-10-12 11:20:34
7	014CAL.S.d	S07	2021-10-12 11:23:39
8	016CAL.S.d	S08	2021-10-12 11:46:27

9 Be [No Gas] ISTD:6 Li [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	37.32	0.0000	P	18.7
2	☐	1.000	0.962	2393.73	0.0007	P	7.5
3	☐	50.000	49.294	118552.55	0.0333	P	2.3
4	☐	125.000	123.677	289740.36	0.0835	P	0.2
5	☐	250.000	248.288	573014.16	0.1676	P	0.7
6	☐	500.000	502.215	1142549.60	0.3390	A	1.0
7	☐	1000.000	999.521	2323750.92	0.6746	A	2.7
8	☐			447.25	0.0001	P	6.4

$$y = 6.7490E-004 * x + 1.1225E-005$$


$$R = 1.0000$$

$$DL = 0.009341$$

Weight: <None>

Min Conc: 0

10 B [No Gas] ISTD:6 Li [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	631.22	0.0002	P	8.3
2	☐	25.000	21.166	6148.09	0.0017	P	7.5
3	☐	50.000	43.935	11811.78	0.0033	P	2.6
4	☐	125.000	114.480	28913.42	0.0083	P	5.3
5	☐	250.000	231.892	57025.85	0.0167	P	4.1
6	☐	500.000	542.871	130749.09	0.0388	P	1.5
7	☐	1000.000	984.805	241998.10	0.0702	P	1.4
8	☐			5920.04	0.0018	P	0.6


$$y = 7.1099E-005 * x + 1.8996E-004$$

$$R = 0.9984$$

$$DL = 0.6633$$

Weight: <None>

Min Conc: 0

11 B [No Gas] ISTD:6 Li [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	3147.72	0.0009	P	3.6
2	☐	25.000	21.785	28968.24	0.0080	P	6.4
3	☐	50.000	47.128	57598.64	0.0162	P	3.6
4	☐	125.000	120.183	137938.80	0.0397	P	3.8
5	☐	250.000	246.913	275754.44	0.0806	P	4.8
6	☐	500.000	573.830	627569.62	0.1862	P	4.5
7	☐	1000.000	964.683	1076717.21	0.3123	A	2.5
8	☐			28337.65	0.0085	P	1.6


$$y = 3.2278\text{E-}004 * x + 9.4658\text{E-}004$$

$$R = 0.9958$$


$$DL = 0.3191$$

Weight: <None>


Min Conc: 0

12 C [No Gas] No available calibration points

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐			5321138.66		A	1.1
2	☐			5400697.39		A	2.1
3	☐			6005161.03		A	0.9
4	☐			6064221.36		A	1.0
5	☐			6067968.45		A	1.8
6	☐			6513401.62		A	1.8
7	☐			7032756.32		A	2.2
8	☐			8381444.20		A	0.5

12 C [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			61501.87		P	1.7
2	☐			64257.88		P	1.5
3	☐			72827.97		P	1.1
4	☐			74235.58		P	1.2
5	☐			73431.13		P	0.5
6	☐			79199.44		P	0.7
7	☐			89755.03		P	0.4
8	☐			108768.25		P	1.9

23 Na [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	33088.06	0.1159	P	1.4
2	☐	500.000	495.821	269411.85	0.9316	P	0.9
3	☐	5000.000	5426.931	2584750.54	9.0442	A	0.5
4	☐	12500.000	13522.515	6191990.89	22.3628	A	1.1
5	☐	25000.000	26797.149	12124254.18	44.2020	A	2.3
6	☐	50000.000	50919.518	24388414.96	83.8876	A	1.2
7	☐	100000.00	104670.68	49468266.04	172.3179	A	0.8
8	☐	500000.00	498854.22	237858235.8	820.8202	A	1.6


$$y = 0.0016 * x + 0.1159$$

$$R = 1.0000$$

$$DL = 2.877$$

Weight: <None>

Min Conc: 0

24 Mg [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	652.81	0.0023	P	7.9
2	☐	500.000	493.765	122754.50	0.4244	P	2.0
3	☐	5000.000	5507.346	1346350.04	4.7107	A	2.4
4	☐	12500.000	13727.041	3250036.68	11.7380	A	1.4
5	☐	25000.000	27201.369	6378714.55	23.2577	A	3.2
6	☐	50000.000	51545.998	12813069.30	44.0707	A	1.5
7	☐	100000.00	104195.45	25572866.20	89.0825	A	0.7
8	☐	500000.00	498860.49	123598580.5	426.4952	A	0.5


$$y = 8.5493\text{E-}004 * x + 0.0023$$

$$R = 1.0000$$

$$DL = 0.6327$$

Weight: <None>

Min Conc: 0

27 Al [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	202.06	0.0007	P	2.8
2	☐	20.000	21.096	2039.06	0.0071	P	1.1
3	☐	1000.000	1034.241	89067.35	0.3117	P	1.2
4	☐	2500.000	2654.362	221176.18	0.7988	P	0.6
5	☐	5000.000	5218.978	430622.39	1.5699	P	2.2
6	☐	10000.000	9895.791	865206.53	2.9762	P	1.9
7	☐	20000.000	20531.371	1772194.01	6.1741	A	2.3
8	☐	100000.00	99888.996	8703933.79	30.0352	A	1.1

$$y = 3.0068\text{E-}004 * x + 7.0770\text{E-}004$$


$$R = 1.0000$$

$$DL = 0.2005$$

Weight: <None>

Min Conc: 0

28 Si [No Gas] ISTD : 45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	565128.35	0.0755	P	6.7
2	☐	10.000	7.738	644856.16	0.0824	P	4.0
3	☐	50.000	52.211	959254.53	0.1225	A	0.6
4	☐	125.000	136.472	1473302.15	0.1984	A	4.1
5	☐	250.000	264.722	2330973.64	0.3139	A	1.1
6	☐	500.000	543.275	4447644.28	0.5648	A	4.6
7	☐	1000.000	973.160	7741445.31	0.9520	A	5.8
8	☐			795085.51	0.1023	P	5.2

$$y = 9.0075E-004 * x + 0.0755$$


$$R = 0.9984$$

$$DL = 16.96$$

Weight: <None>

Min Conc: 0

31 P [He] ISTD : 45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	4.751	180.56	0.0006	P	13.9
2	☐	0.000	-4.751	136.12	0.0005	P	15.7
3	☐	1000.000	1053.164	5273.33	0.0185	P	3.3
4	☐	2500.000	2666.209	12703.12	0.0459	P	5.5
5	☐	5000.000	5316.892	24944.27	0.0909	P	2.8
6	☐	10000.000	10017.652	49671.87	0.1709	P	2.2
7	☐	20000.000	19888.517	97207.00	0.3387	P	3.6
8	☐			238.90	0.0008	P	9.7


$$y = 1.7001E-005 * x + 5.5159E-004$$

$$R = 0.9998$$

$$DL = 14.27$$

Weight: <None>

Min Conc: 0

34 S [No Gas] ISTD:45 Sc [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	94.034	466196.01	0.0622	P	6.6
2	☐	0.000	-94.034	478453.17	0.0612	P	4.2
3	☐	1000.000	1146.448	534865.89	0.0683	P	1.8
4	☐	2500.000	3536.542	609598.96	0.0821	P	2.8
5	☐	5000.000	6739.708	746290.04	0.1005	P	1.0
6	☐	10000.000	10253.409	951689.74	0.1207	M	5.7
7	☐	20000.000	19301.478	1406217.94	0.1728	A	1.9
8	☐			421368.24	0.0542	P	1.2


$$y = 5.7584\text{E-}006 * x + 0.0617$$

R = 0.9945


DL = 1746

Weight: <None>


Min Conc: 0

35 Cl [No Gas] No available calibration points

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐			1269046.57		A	2.2
2	☐			1136956.40		M	2.0
3	☐			1159890.05		A	1.9
4	☐			1087673.56		P	0.6
5	☐			1071570.66		P	1.0
6	☐			1019694.44		P	1.4
7	☐			1015531.84		P	1.4
8	☐			1219534.78		A	2.7

35 Cl [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			9734.10		P	3.5
2	☐			8764.02		P	2.8
3	☐			8978.01		P	1.9
4	☐			8358.20		P	3.3
5	☐			8502.75		P	5.2
6	☐			7896.80		P	3.5
7	☐			7393.76		P	4.4
8	☐			8972.43		P	3.7

39 K [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	25297.75	0.0886	P	4.0
2	☐	500.000	464.602	124214.98	0.4295	P	1.7
3	☐	2500.000	2521.396	554024.97	1.9388	P	1.3
4	☐	6250.000	6633.795	1372355.33	4.9564	A	0.6
5	☐	12500.000	12916.800	2623761.02	9.5668	A	3.9
6	☐	25000.000	24811.140	5318375.47	18.2947	A	2.6
7	☐	50000.000	49973.710	10551886.70	36.7587	A	1.3
8	☐	250000.00	249993.56	53184179.66	183.5310	A	1.5


$$y = 7.3379\text{E-}004 * x + 0.0886$$

$$R = 1.0000$$

$$DL = 14.54$$

Weight: <None>

Min Conc: 0

43 Ca [No Gas] ISTD:45 Sc [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	608.69	0.0001	P	11.5
2	☐	500.000	482.060	17510.11	0.0022	P	4.2
3	☐	5000.000	5245.212	184426.06	0.0235	P	0.5
4	☐	12500.000	13768.197	458202.83	0.0617	P	2.7
5	☐	25000.000	27618.506	918118.86	0.1237	P	1.0
6	☐	50000.000	55023.898	1940681.92	0.2463	A	0.6
7	☐	100000.00	105909.51	3855157.83	0.4739	A	3.1
8	☐	500000.00	498150.64	17337542.21	2.2289	A	0.3


$$y = 4.4743\text{E-}006 * x + 8.1431\text{E-}005$$

$$R = 0.9999$$

$$DL = 6.29$$

Weight: <None>

Min Conc: 0

44 Ca [No Gas] ISTD:45 Sc [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	57712.78	0.0077	P	5.9
2	☐	500.000	465.384	332345.61	0.0425	P	4.5
3	☐	5000.000	5231.717	3122572.03	0.3987	A	0.5
4	☐	12500.000	13223.154	7398439.06	0.9960	A	3.3
5	☐	25000.000	26878.221	14972893.62	2.0166	A	1.3
6	☐	50000.000	51939.489	30642482.66	3.8898	A	2.1
7	☐	100000.00	101544.85	61821077.88	7.5974	A	0.7
8	☐	500000.00	499382.80	290338481.4	37.3325	A	1.5


$$y = 7.4742\text{E-}005 * x + 0.0077$$

$$R = 1.0000$$

$$DL = 18.19$$

Weight: <None>

Min Conc: 0

47 Ti [No Gas] ISTD :45 Sc [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	341.68	0.0000	P	25.9
2	☐	5.000	4.828	8705.66	0.0011	P	6.5
3	☐	250.000	254.158	440278.34	0.0562	P	0.9
4	☐	625.000	661.056	1085907.64	0.1462	P	1.4
5	☐	1250.000	1326.696	2177526.96	0.2933	A	2.4
6	☐	2500.000	2530.641	4406848.93	0.5594	A	1.8
7	☐	5000.000	4960.791	8922365.82	1.0965	A	0.9
8	☐			2116.88	0.0003	P	4.9


$$y = 2.2103E-004 * x + 4.5620E-005$$

$$R = 0.9998$$

$$DL = 0.1601$$

Weight: <None>

Min Conc: 0

51 V [He] ISTD :45 Sc [He]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	28.89	0.0001	P	13.1
2	☐	5.000	4.799	13853.46	0.0479	P	0.6
3	☐	50.000	49.689	141438.89	0.4950	P	2.3
4	☐	125.000	126.852	349851.90	1.2635	P	0.5
5	☐	250.000	253.910	693682.33	2.5290	P	2.4
6	☐	500.000	493.957	1430229.08	4.9199	A	2.3
7	☐	1000.000	1001.829	2864459.23	9.9782	A	0.9
8	☐			792.25	0.0027	P	5.8


$$y = 0.0100 * x + 1.0117E-004$$

$$R = 1.0000$$

$$DL = 0.003977$$

Weight: <None>

Min Conc: 0

52 Cr [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1021.16	0.0036	P	4.0
2	☐	2.000	1.925	7984.52	0.0276	P	1.4
3	☐	50.000	49.646	178093.28	0.6233	P	2.8
4	☐	125.000	126.826	439342.02	1.5867	P	0.4
5	☐	250.000	252.679	866176.81	3.1578	P	1.9
6	☐	500.000	501.463	1820925.36	6.2633	A	1.1
7	☐	1000.000	998.388	3578631.20	12.4664	A	1.7
8	☐			12517.77	0.0432	P	1.9


$$y = 0.0125 * x + 0.0036$$

$$R = 1.0000$$

$$DL = 0.03473$$

Weight: <None>

Min Conc: 0

54 Fe [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	948.20	0.0033	P	10.0
2	☐	50.000	53.356	8928.87	0.0309	P	3.7
3	☐	2500.000	2679.978	396486.51	1.3875	P	1.6
4	☐	6250.000	6876.436	984300.50	3.5549	P	0.8
5	☐	12500.000	14077.372	1995029.92	7.2742	A	3.2
6	☐	25000.000	26603.888	3995501.39	13.7440	A	2.3
7	☐	50000.000	53739.852	7968629.46	27.7595	A	1.4
8	☐	250000.00	248995.31	37266539.26	128.6073	A	2.3


$$y = 5.1649E-004 * x + 0.0033$$

$$R = 0.9999$$

$$DL = 1.933$$

Weight: <None>

Min Conc: 0

⁵⁵Mn [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	162.60	0.0006	P	28.4
2	☐	1.000	0.890	1811.73	0.0063	P	6.3
3	☐	500.000	496.597	908541.61	3.1794	P	1.3
4	☐	1250.000	1324.134	2346985.19	8.4765	A	0.4
5	☐	2500.000	2570.870	4513950.62	16.4571	A	2.6
6	☐	5000.000	4905.972	9130088.25	31.4044	A	1.2
7	☐	10000.000	10020.200	18413462.70	64.1413	A	0.6
8	☐			26735.78	0.0923	P	2.1


$$y = 0.0064 * x + 5.6985E-004$$

$$R = 0.9999$$

$$DL = 0.07581$$

Weight: <None>

Min Conc: 0

⁵⁶Fe [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	7852.30	0.0275	P	2.3
2	☐	50.000	52.107	148230.37	0.5125	P	0.6
3	☐	2500.000	2758.399	7344689.69	25.7044	A	2.2
4	☐	6250.000	7033.563	18135762.67	65.5002	A	0.3
5	☐	12500.000	13993.010	35738739.69	130.2830	A	1.5
6	☐	25000.000	26636.211	72089036.80	247.9736	A	1.9
7	☐	50000.000	53815.796	143816855.0	500.9778	A	1.1
8	☐	250000.00	248976.39	671608155.7	2,317.651	A	1.6


$$y = 0.0093 * x + 0.0275$$

$$R = 0.9999$$

$$DL = 0.2047$$

Weight: <None>

Min Conc: 0

⁵⁷Fe [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	311.12	0.0011	P	13.5
2	☐	50.000	50.634	3772.79	0.0130	P	1.5
3	☐	2500.000	2648.117	179004.78	0.6264	P	1.1
4	☐	6250.000	6757.194	442095.63	1.5967	P	0.5
5	☐	12500.000	13502.819	874968.74	3.1896	P	1.3
6	☐	25000.000	26869.354	1844601.18	6.3459	A	3.8
7	☐	50000.000	53522.773	3628635.01	12.6397	A	0.8
8	☐	250000.00	249044.20	17042410.68	58.8091	A	1.1


$$y = 2.3613\text{E-}004 * x + 0.0011$$

$$R = 0.9999$$

$$DL = 1.872$$

Weight: <None>

Min Conc: 0

⁵⁹Co [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	92.22	0.0003	P	25.2
2	☐	1.000	0.979	6254.76	0.0216	P	2.4
3	☐	50.000	50.149	311931.38	1.0916	P	2.0
4	☐	125.000	128.897	776744.37	2.8053	P	0.2
5	☐	250.000	259.594	1549708.47	5.6495	A	1.9
6	☐	500.000	496.058	3138098.11	10.7953	A	3.2
7	☐	1000.000	999.078	6240858.03	21.7418	A	2.0
8	☐			14439.64	0.0498	P	1.1


$$y = 0.0218 * x + 3.2296\text{E-}004$$

$$R = 0.9999$$

$$DL = 0.01123$$

Weight: <None>

Min Conc: 0

60 Ni [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	132.22	0.0005	P	18.8
2	☐	1.000	1.007	1866.81	0.0065	P	2.4
3	☐	50.000	50.845	86577.20	0.3029	P	1.0
4	☐	125.000	126.658	208756.59	0.7540	P	1.3
5	☐	250.000	254.193	414906.29	1.5127	P	2.4
6	☐	500.000	478.943	828460.33	2.8497	P	1.6
7	☐	1000.000	1009.231	1723765.29	6.0044	A	0.6
8	☐			6283.68	0.0217	P	2.6


$$y = 0.0059 * x + 4.6295E-004$$

$$R = 0.9997$$

$$DL = 0.04395$$

Weight: <None>

Min Conc: 0

63 Cu [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1106.72	0.0039	P	8.0
2	☐	2.000	1.874	9978.05	0.0345	P	1.4
3	☐	500.000	528.919	2471368.10	8.6498	A	3.2
4	☐	1250.000	1337.082	6052797.76	21.8604	A	0.8
5	☐	2500.000	2656.468	11910450.24	43.4276	A	3.3
6	☐	5000.000	5021.028	23863002.14	82.0798	A	1.3
7	☐	10000.000	9938.038	46633455.42	162.4554	A	1.5
8	☐			16485.27	0.0569	P	3.8


$$y = 0.0163 * x + 0.0039$$

$$R = 0.9998$$

$$DL = 0.05721$$

Weight: <None>

Min Conc: 0

66 Zn [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	737.81	0.0026	P	8.6
2	☐	2.000	2.344	2002.38	0.0069	P	0.5
3	☐	500.000	524.834	278340.59	0.9742	P	2.6
4	☐	1250.000	1332.869	683893.92	2.4700	P	1.4
5	☐	2500.000	2606.515	1324200.10	4.8277	P	2.3
6	☐	5000.000	5057.262	2722369.86	9.3645	A	2.1
7	☐	10000.000	9933.140	5279584.05	18.3907	A	0.7
8	☐			10307.17	0.0356	P	1.7


$$y = 0.0019 * x + 0.0026$$

$$R = 0.9999$$

$$DL = 0.3606$$

Weight: <None>

Min Conc: 0

75 As [He] ISTD:89 Y [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	15.00	0.0000	P	17.6
2	☐	1.000	0.991	333.37	0.0001	P	2.8
3	☐	50.000	50.497	16233.93	0.0050	P	1.4
4	☐	125.000	130.974	40580.67	0.0130	P	1.2
5	☐	250.000	253.725	80027.64	0.0253	P	1.6
6	☐	500.000	498.538	163072.36	0.0496	P	0.8
7	☐	1000.000	999.028	322279.14	0.0995	P	1.9
8	☐			139.68	0.0000	P	6.4

$$y = 9.9559E-005 * x + 4.6094E-006$$


$$R = 1.0000$$

$$DL = 0.02451$$

Weight: <None>

Min Conc: 0

77 Se [He] ISTD:89 Y [He]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	2.22	0.0000	P	173.
2	☐	5.000	5.315	37.78	0.0000	P	8.3
3	☐	50.000	48.762	327.79	0.0001	P	3.5
4	☐	125.000	135.305	874.48	0.0003	P	11.5
5	☐	250.000	262.192	1721.23	0.0005	P	6.0
6	☐	500.000	493.881	3360.44	0.0010	P	4.5
7	☐	1000.000	998.784	6699.43	0.0021	P	2.6
8	☐			5.56	0.0000	P	91.3

$$y = 2.0697E-006 * x + 6.8314E-007$$


R = 0.9998

DL = 1.715

Weight: <None>

Min Conc: 0

78 Se [He] ISTD:89 Y [He]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	85.34	0.0000	P	10.7
2	☐	5.000	4.598	189.35	0.0001	P	5.7
3	☐	50.000	49.099	1199.84	0.0004	P	4.3
4	☐	125.000	133.929	3014.30	0.0010	P	4.2
5	☐	250.000	257.061	5816.13	0.0018	P	1.5
6	☐	500.000	502.873	11717.72	0.0036	P	0.5
7	☐	1000.000	995.729	22801.34	0.0070	P	1.6
8	☐			110.01	0.0000	P	8.0


$$y = 7.0407E-006 * x + 2.6227E-005$$

R = 0.9999


DL = 1.199

Weight: <None>

Min Conc: 0

81 Br [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			22274.06		P	3.3
2	☐			22160.43		P	1.6
3	☐			21665.77		P	4.9
4	☐			21495.40		P	3.2
5	☐			21047.49		P	1.2
6	☐			22989.35		P	2.7
7	☐			23089.77		P	2.5
8	☐			22538.08		P	2.7

82 Se [No Gas] ISTD: 89 Y [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	4.21	0.0000	P	925.
2	☐	5.000	5.081	936.31	0.0000	P	2.9
3	☐	50.000	52.707	9369.10	0.0004	P	1.1
4	☐	125.000	132.875	23076.01	0.0009	P	1.5
5	☐	250.000	261.428	45409.31	0.0018	P	1.4
6	☐	500.000	502.849	91917.12	0.0035	P	1.7
7	☐	1000.000	994.598	184429.08	0.0070	P	1.4
8	☐			161.11	0.0000	P	25.0


$$y = 7.0210\text{E-}006 * x + 2.3972\text{E-}007$$

$$R = 0.9999$$


$$DL = 0.9477$$

Weight: <None>

Min Conc: 0

83 Kr [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			973.38		P	5.5
2	☐			998.94		P	2.0
3	☐			1001.16		P	5.9
4	☐			990.04		P	7.0
5	☐			997.83		P	5.7
6	☐			1008.94		P	4.2
7	☐			1006.71		P	10.8
8	☐			1030.05		P	2.4

86 Sr [No Gas] ISTD : 89 Y [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	1572.35	0.0001	P	1.6
2	☐	1.000	1.098	5629.08	0.0002	P	3.2
3	☐	500.000	512.628	1797412.04	0.0710	A	2.2
4	☐	1250.000	1284.895	4401632.96	0.1780	A	0.6
5	☐	2500.000	2592.799	8881851.41	0.3591	A	1.7
6	☐	5000.000	4992.929	17999617.24	0.6914	A	1.2
7	☐	10000.000	9975.342	36478761.92	1.3812	A	1.2
8	☐			30311.17	0.0012	P	0.8


$$y = 1.3846E-004 * x + 6.3841E-005$$

$$R = 1.0000$$

$$DL = 0.02226$$

Weight: <None>

Min Conc: 0

88 Sr [No Gas] ISTD:89 Y [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	805.60	0.0000	P	4.7
2	☐	1.000	1.057	33950.41	0.0013	P	3.5
3	☐	500.000	506.052	15377522.57	0.6078	A	1.5
4	☐	1250.000	1275.579	37887324.80	1.5319	A	1.7
5	☐	2500.000	2579.564	76634284.35	3.0980	A	1.2
6	☐	5000.000	4970.930	155400767.3	5.9699	A	2.6
7	☐	10000.000	9991.144	316913370.9	11.9989	A	0.7
8	☐			250338.75	0.0099	P	1.4


$$y = 0.0012 * x + 3.2716E-005$$

$$R = 1.0000$$

$$DL = 0.003871$$

Weight: <None>

Min Conc: 0

90 Zr [No Gas] ISTD:89 Y [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	1044.52	0.0000	P	9.6
2	☐	1.000	9.127	19019.03	0.0007	P	3.4
3	☐	500.000	496.731	947348.73	0.0374	P	0.2
4	☐	1250.000	1288.606	2400293.09	0.0971	A	1.9
5	☐	2500.000	2542.836	4737209.92	0.1915	A	0.3
6	☐	5000.000	4982.635	9766874.87	0.3752	A	2.1
7	☐	10000.000	9993.311	19873426.83	0.7524	A	0.9
8	☐			11591.27	0.0005	P	1.6


$$y = 7.5290E-005 * x + 4.2409E-005$$

$$R = 1.0000$$

$$DL = 0.1622$$

Weight: <None>

Min Conc: 0

91 Zr [No Gas] ISTD: 89 Y [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	194.45	0.0000	P	21.7
2	☐	1.000	8.203	3839.52	0.0001	P	6.6
3	☐	500.000	497.753	214201.40	0.0085	P	0.6
4	☐	1250.000	1266.512	532401.89	0.0215	P	1.1
5	☐	2500.000	2528.301	1062815.51	0.0430	P	2.1
6	☐	5000.000	5020.776	2221142.74	0.0853	A	2.0
7	☐	10000.000	9980.584	4478640.64	0.1696	A	2.9
8	☐			2742.01	0.0001	P	3.2


$$y = 1.6992\text{E-}005 * x + 7.9421\text{E-}006$$

$$R = 1.0000$$

$$DL = 0.3038$$

Weight: <None>

Min Conc: 0

94 Mo [No Gas] ISTD: 89 Y [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	366.68	0.0000	P	25.8
2	☐	5.000	4.706	22310.21	0.0009	P	2.5
3	☐	250.000	254.468	1150869.97	0.0455	P	0.9
4	☐	625.000	644.338	2847724.15	0.1152	A	1.3
5	☐	1250.000	1283.362	5672646.65	0.2293	A	1.9
6	☐	2500.000	2459.894	11442670.22	0.4396	A	2.3
7	☐	5000.000	5009.072	23636700.70	0.8951	A	2.3
8	☐			10148.44	0.0004	P	2.7

$$y = 1.7869\text{E-}004 * x + 1.4764\text{E-}005$$


$$R = 0.9999$$

$$DL = 0.06395$$

Weight: <None>

Min Conc: 0

95 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	97.23	0.0000	P	9.7
2	☐	5.000	4.699	26670.62	0.0010	P	3.4
3	☐	250.000	254.443	1396424.06	0.0552	M	0.7
4	☐	625.000	645.954	3464580.11	0.1401	A	1.6
5	☐	1250.000	1275.271	6841615.08	0.2766	A	1.7
6	☐	2500.000	2480.977	14009437.64	0.5381	A	0.9
7	☐	5000.000	5000.352	28640557.24	1.0845	A	1.8
8	☐			9028.17	0.0004	P	3.2

$$y = 2.1689E-004 * x + 3.9374E-006$$


$$R = 1.0000$$

$$DL = 0.005256$$

Weight: <None>

Min Conc: 0

96 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	280.57	0.0000	P	3.3
2	☐	5.000	4.739	29960.73	0.0011	P	3.2
3	☐	250.000	257.570	1564905.83	0.0619	A	2.5
4	☐	625.000	643.008	3817707.46	0.1544	A	2.2
5	☐	1250.000	1271.688	7551880.99	0.3053	A	2.3
6	☐	2500.000	2483.178	15521031.77	0.5962	A	1.6
7	☐	5000.000	5000.360	31707975.36	1.2005	A	0.7
8	☐			21659.16	0.0009	P	2.3

$$y = 2.4009E-004 * x + 1.1384E-005$$


$$R = 1.0000$$

$$DL = 0.004736$$

Weight: <None>

Min Conc: 0

97 Mo [No Gas] ISTD:89 Y [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	75.00	0.0000	P	21.1
2	☐	5.000	4.960	17202.78	0.0007	P	1.6
3	☐	250.000	253.200	848195.75	0.0335	P	1.3
4	☐	625.000	646.313	2115768.75	0.0856	A	2.7
5	☐	1250.000	1295.176	4241848.10	0.1715	A	0.3
6	☐	2500.000	2500.206	8617306.70	0.3310	A	1.3
7	☐	5000.000	4985.779	17432716.54	0.6601	A	1.3
8	☐			5637.44	0.0002	P	8.4

$$y = 1.3239\text{E-}004 * x + 3.0540\text{E-}006$$


$$R = 1.0000$$

$$DL = 0.01462$$

Weight: <None>

Min Conc: 0

98 Mo [No Gas] ISTD:89 Y [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	175.01	0.0000	P	30.8
2	☐	5.000	4.749	42449.17	0.0016	P	3.6
3	☐	250.000	251.297	2170693.67	0.0858	A	2.7
4	☐	625.000	643.032	5427863.10	0.2195	A	3.5
5	☐	1250.000	1288.408	10881110.40	0.4399	A	0.6
6	☐	2500.000	2510.078	22309820.12	0.8569	A	1.5
7	☐	5000.000	4983.041	44927525.66	1.7012	A	1.4
8	☐			13387.41	0.0005	P	1.6


$$y = 3.4139\text{E-}004 * x + 7.0567\text{E-}006$$

$$R = 1.0000$$

$$DL = 0.01907$$

Weight: <None>

Min Conc: 0

106 Cd [No Gas] ISTD :159 Tb [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	8463.96	0.0007	P	1.9
2	☐	1.000	-0.428	8569.56	0.0007	P	4.4
3	☐	50.000	49.858	20971.96	0.0018	P	2.1
4	☐	125.000	132.396	40853.03	0.0035	P	1.6
5	☐	250.000	261.400	72813.02	0.0063	P	1.9
6	☐	500.000	518.652	141977.86	0.0117	P	1.6
7	☐	1000.000	986.908	273026.37	0.0216	P	2.2
8	☐			6754.62	0.0006	P	3.0


$$y = 2.1163\text{E-}005 * x + 7.4044\text{E-}004$$

$$R = 0.9996$$

$$DL = 2.016$$

Weight: <None>

Min Conc: 0

107 Ag [No Gas] ISTD :159 Tb [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	227.79	0.0000	P	13.9
2	☐	1.000	0.960	12058.38	0.0010	P	4.8
3	☐	50.000	52.599	646155.15	0.0554	P	2.4
4	☐	125.000	136.458	1655752.99	0.1436	A	1.7
5	☐	250.000	268.048	3273479.91	0.2820	A	2.0
6	☐	500.000	507.035	6463225.19	0.5334	A	2.6
7	☐	1000.000	990.408	13154391.09	1.0419	A	1.8
8	☐			2761.47	0.0002	P	1.5


$$y = 0.0011 * x + 1.9801\text{E-}005$$

$$R = 0.9997$$

$$DL = 0.007865$$

Weight: <None>

Min Conc: 0

108 Cd [No Gas] ISTD :159 Tb [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	61.11	0.0000	P	37.2
2	☐	1.000	0.962	229.17	0.0000	P	25.2
3	☐	50.000	54.256	9408.87	0.0008	P	0.4
4	☐	125.000	135.784	23164.00	0.0020	P	1.3
5	☐	250.000	263.193	45130.60	0.0039	P	1.7
6	☐	500.000	514.853	92092.79	0.0076	P	1.4
7	☐	1000.000	987.714	184003.02	0.0146	P	2.0
8	☐			176.39	0.0000	P	19.2


$$y = 1.4750\text{E-}005 * x + 5.4310\text{E-}006$$

$$R = 0.9997$$

$$DL = 0.4106$$

Weight: <None>

Min Conc: 0

109 Ag [No Gas] ISTD :159 Tb [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	161.12	0.0000	P	39.5
2	☐	1.000	0.994	11710.89	0.0010	P	4.0
3	☐	50.000	53.181	615838.70	0.0528	P	2.4
4	☐	125.000	136.249	1558476.25	0.1351	A	2.1
5	☐	250.000	268.809	3094700.44	0.2666	A	2.4
6	☐	500.000	521.674	6269913.90	0.5174	A	0.5
7	☐	1000.000	982.895	12307565.81	0.9748	A	2.5
8	☐			2455.84	0.0002	P	8.3


$$y = 9.9174\text{E-}004 * x + 1.3982\text{E-}005$$

$$R = 0.9994$$

$$DL = 0.01669$$

Weight: <None>

Min Conc: 0



	R _{ct}	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	859.17	0.0001	P	1.8
2	☐	1.000	0.985	3045.38	0.0003	P	6.8
3	☐	50.000	53.125	117380.21	0.0101	P	1.5
4	☐	125.000	134.030	291220.80	0.0253	P	1.8
5	☐	250.000	265.260	579339.70	0.0499	P	1.9
6	☐	500.000	513.536	1169825.10	0.0965	P	1.4
7	☐	1000.000	988.132	2344508.02	0.1857	A	1.3
8	☐			1464.53	0.0001	P	3.7


$$y = 1.8784\text{E-}004 * x + 7.5175\text{E-}005$$

$$R = 0.9997$$

$$DL = 0.0216$$

Weight: <None>

Min Conc: 0



	R _{ct}	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	2692.01	0.0002	P	7.5
2	☐	5.000	4.893	41148.84	0.0035	P	1.7
3	☐	50.000	51.860	408004.42	0.0350	P	3.3
4	☐	125.000	130.964	1013911.73	0.0879	P	1.2
5	☐	250.000	262.151	2040163.72	0.1757	A	1.8
6	☐	500.000	514.357	4175874.23	0.3446	A	1.7
7	☐	1000.000	988.946	8361617.72	0.6623	A	2.1
8	☐			4653.72	0.0004	P	5.8


$$y = 6.6949\text{E-}004 * x + 2.3603\text{E-}004$$

$$R = 0.9998$$

$$DL = 0.0796$$

Weight: <None>

Min Conc: 0



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	75.00	0.0000	P	18.0
2	☐	2.000	1.907	17269.75	0.0015	P	2.9
3	☐	50.000	50.894	457372.01	0.0392	P	2.4
4	☐	125.000	130.132	1155250.65	0.1002	P	1.6
5	☐	250.000	264.372	2362105.02	0.2035	A	3.1
6	☐	500.000	508.583	4743766.73	0.3915	A	1.4
7	☐	1000.000	991.430	9633323.13	0.7631	A	2.8
8	☐			4375.85	0.0004	P	6.6


$$y = 7.6972\text{E-}004 * x + 6.6143\text{E-}006$$

$$R = 0.9998$$

$$DL = 0.004639$$

Weight: <None>

Min Conc: 0



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	41.67	0.0000	P	66.1
2	☐	10.000	9.619	22224.27	0.0019	P	3.4
3	☐	250.000	252.466	580375.49	0.0497	P	2.2
4	☐	625.000	631.218	1433404.30	0.1243	P	2.3
5	☐	1250.000	1299.680	2970810.46	0.2559	A	2.1
6	☐	2500.000	2553.182	6092149.70	0.5027	A	1.6
7	☐	5000.000	4960.089	12328919.63	0.9767	A	3.0
8	☐			5754.17	0.0005	P	1.8


$$y = 1.9691\text{E-}004 * x + 3.7498\text{E-}006$$

$$R = 0.9999$$

$$DL = 0.03776$$

Weight: <None>

Min Conc: 0

137 Ba [No Gas] ISTD : 159 Tb [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	66.67	0.0000	P	50.8
2	☐	10.000	9.908	39547.40	0.0034	P	5.3
3	☐	250.000	256.388	1018847.23	0.0873	P	2.0
4	☐	625.000	650.444	2553494.05	0.2214	A	0.9
5	☐	1250.000	1296.845	5124432.75	0.4414	A	1.3
6	☐	2500.000	2551.062	10522849.20	0.8683	A	0.5
7	☐	5000.000	4959.258	21309518.18	1.6880	A	2.6
8	☐			10496.00	0.0009	P	3.5


$$y = 3.4038E-004 * x + 5.6992E-006$$

$$R = 0.9999$$


$$DL = 0.02553$$

Weight: <None>


Min Conc: 0

150 Nd [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐			8.89		P	43.3
2	☐			6.67		P	100.
3	☐			42.22		P	39.7
4	☐			77.78		P	30.1
5	☐			155.56		P	17.3
6	☐			251.12		P	10.7
7	☐			466.69		P	27.6
8	☐			97.78		P	3.9

150 Nd [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			1.11		P	173.
2	☐			1.11		P	173.
3	☐			5.55		P	69.3
4	☐			4.44		P	43.4
5	☐			7.78		P	107.
6	☐			26.67		P	54.5
7	☐			47.78		P	10.7
8	☐			34.44		P	40.3

156 Dy [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			22.22		P	21.6
2	☐			8.33		P	100.
3	☐			27.78		P	45.8
4	☐			88.89		P	5.4
5	☐			119.45		P	46.5
6	☐			241.68		P	6.0
7	☐			425.02		P	16.8
8	☐			1600.15		P	2.6

156 Dy [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			3.33		P	100.
2	☐			0.00		P	
3	☐			17.78		P	57.3
4	☐			32.22		P	12.0
5	☐			58.89		P	19.9
6	☐			98.89		P	24.4
7	☐			196.67		P	14.8
8	☐			745.58		P	5.6

160 Gd [No Gas] Noavailable calibration poin...


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			141.67		P	23.5
2	☐			141.67		P	30.6
3	☐			152.78		P	31.0
4	☐			116.67		P	14.3
5	☐			183.34		P	12.0
6	☐			280.57		P	6.2
7	☐			419.46		P	17.4
8	☐			1614.03		P	8.4

160 Gd [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			98.89		P	12.8
2	☐			66.67		P	36.1
3	☐			107.78		P	12.9
4	☐			91.11		P	4.2
5	☐			131.11		P	5.3
6	☐			166.67		P	6.9
7	☐			223.34		P	10.4
8	☐			877.82		P	2.2

164 Er [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			63.89		P	64.3
2	☐			69.45		P	36.7
3	☐			63.89		P	39.8
4	☐			100.00		P	28.9
5	☐			105.56		P	4.6
6	☐			91.67		P	48.1
7	☐			119.45		P	34.4
8	☐			308.35		P	10.8

164 Er [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			38.89		P	17.8
2	☐			24.44		P	47.9
3	☐			32.22		P	33.3
4	☐			31.11		P	24.7
5	☐			41.11		P	40.8
6	☐			37.78		P	33.4
7	☐			57.78		P	14.5
8	☐			145.56		P	3.5

172 Yb [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			25.00		P	57.7
2	☐			38.89		P	32.7
3	☐			33.34		P	43.3
4	☐			16.67		P	86.6
5	☐			44.45		P	47.2
6	☐			44.45		P	47.2
7	☐			50.00		P	16.7
8	☐			169.45		P	30.1

172 Yb [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			11.11		P	50.0
2	☐			18.52		P	17.3
3	☐			14.82		P	57.3
4	☐			20.37		P	15.7
5	☐			14.82		P	57.3
6	☐			14.82		P	57.3
7	☐			31.48		P	44.4
8	☐			105.56		P	27.8

176 Yb [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			1939.09		P	6.9
2	☐			2002.98		P	1.3
3	☐			3425.55		P	4.2
4	☐			6037.62		P	3.5
5	☐			10173.69		P	6.3
6	☐			18432.76		P	1.5
7	☐			35804.85		P	0.4
8	☐			1600.14		P	5.6

176 Yb [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			4476.76		P	2.9
2	☐			3900.65		P	3.3
3	☐			4539.77		P	5.1
4	☐			5745.80		P	5.2
5	☐			7934.02		P	5.2
6	☐			12435.65		P	3.0
7	☐			21232.14		P	1.4
8	☐			4541.61		P	1.4

203 Tl [No Gas] ISTD :209 Bi [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	619.48	0.0001	P	6.3
2	☐	1.000	0.901	7961.01	0.0012	P	4.1
3	☐	50.000	48.404	382405.28	0.0596	P	2.0
4	☐	125.000	124.756	961329.67	0.1535	P	0.7
5	☐	250.000	259.613	2007070.40	0.3194	A	4.6
6	☐	500.000	506.179	4021381.73	0.6227	A	1.1
7	☐	1000.000	994.618	8057791.15	1.2234	A	1.7
8	☐			1950.21	0.0003	P	3.9

$$y = 0.0012 * x + 9.9595E-005$$


$$R = 0.9999$$

$$DL = 0.01529$$

Weight: <None>

Min Conc: 0

205 Tl [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	1452.91	0.0002	P	6.7
2	☐	1.000	0.900	18649.98	0.0028	P	5.3
3	☐	50.000	49.858	923278.23	0.1440	P	4.3
4	☐	125.000	129.951	2348063.44	0.3750	A	0.4
5	☐	250.000	262.537	4761432.48	0.7574	A	0.9
6	☐	500.000	502.836	9367605.76	1.4505	A	1.0
7	☐	1000.000	994.836	18898939.31	2.8695	A	2.3
8	☐			5065.02	0.0009	P	6.0

$$y = 0.0029 * x + 2.3395E-004$$


$$R = 0.9999$$

$$DL = 0.01625$$

Weight: <None>

Min Conc: 0

206 [Pb] [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	1972.42	0.0003	P	8.4
2	☐	1.000	0.877	7898.83	0.0012	P	4.4
3	☐	250.000	239.048	1539437.94	0.2402	P	4.3
4	☐	625.000	643.836	4046101.68	0.6463	A	1.6
5	☐	1250.000	1297.013	8181174.49	1.3016	A	1.8
6	☐	2500.000	2551.172	16531706.85	2.5599	A	2.0
7	☐	5000.000	4960.854	32781431.06	4.9775	A	2.6
8	☐			14106.10	0.0024	P	1.3

$$y = 0.0010 * x + 3.1905E-004$$


$$R = 0.9999$$

$$DL = 0.07995$$

Weight: <None>

Min Conc: 0

207 [Pb] [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1672.36	0.0003	P	13.6
2	☐	1.000	0.881	6842.68	0.0010	P	4.2
3	☐	250.000	240.599	1344414.83	0.2097	P	2.2
4	☐	625.000	650.022	3543383.62	0.5660	A	2.8
5	☐	1250.000	1269.931	6949138.58	1.1055	A	1.0
6	☐	2500.000	2543.632	14297880.76	2.2140	A	1.6
7	☐	5000.000	4970.544	28492432.35	4.3261	A	1.8
8	☐			11868.62	0.0021	P	0.9

$$y = 8.7030E-004 * x + 2.7116E-004$$


$$R = 0.9999$$

$$DL = 0.1271$$

Weight: <None>

Min Conc: 0

208 Pb [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	7721.43	0.0012	P	9.1
2	☐	1.000	0.891	31816.00	0.0048	P	2.8
3	☐	250.000	248.007	6396643.18	0.9976	A	2.2
4	☐	625.000	644.328	16213082.71	2.5897	A	1.8
5	☐	1250.000	1273.745	32175487.06	5.1182	A	0.5
6	☐	2500.000	2534.206	65753121.26	10.1818	A	1.8
7	☐	5000.000	4974.645	131631832.5	19.9857	A	1.3
8	☐			55783.46	0.0097	P	0.2


$$y = 0.0040 * x + 0.0012$$

$$R = 0.9999$$

$$DL = 0.08454$$

Weight: <None>

Min Conc: 0

238 U [No Gas] ISTD :209 Bi [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	25.00	0.0000	P	63.8
2	☐	1.000	0.896	21888.83	0.0033	P	1.8
3	☐	50.000	48.529	1150900.28	0.1795	P	2.6
4	☐	125.000	129.234	2992596.94	0.4780	A	1.6
5	☐	250.000	254.940	5927470.10	0.9429	A	0.9
6	☐	500.000	496.567	11861799.22	1.8366	A	2.6
7	☐	1000.000	1000.026	24362670.34	3.6987	A	0.6
8	☐			4400.90	0.0008	P	2.7


$$y = 0.0037 * x + 4.0257E-006$$

$$R = 1.0000$$


$$DL = 0.002083$$

Weight: <None>


Min Conc: 0

6 Li (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		3329047.65		A	4.9
2	☐	1.000		3637123.23		A	6.4
3	☐	1.000		3563285.64		A	1.9
4	☐	1.000		3470768.71		A	0.4
5	☐	1.000		3419244.84		A	0.9
6	☐	1.000		3370833.60		A	0.2
7	☐	1.000		3446048.39		A	2.4
8	☐	1.000		3335412.04		A	1.0

45 Sc (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		7510446.21		A	6.2
2	☐	1.000		7831368.63		A	3.9
3	☐	1.000		7831508.98		A	1.0
4	☐	1.000		7431309.19		A	2.4
5	☐	1.000		7425232.18		A	0.9
6	☐	1.000		7880527.66		A	2.7
7	☐	1.000		8137989.60		A	2.4
8	☐	1.000		7778209.74		A	1.6

45 Sc (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		285502.59		P	0.3
2	☐	1.000		289205.30		P	0.8
3	☐	1.000		285791.18		P	1.3
4	☐	1.000		276882.19		P	0.3
5	☐	1.000		274359.62		P	1.7
6	☐	1.000		290747.62		P	1.0
7	☐	1.000		287079.95		P	1.0
8	☐	1.000		289811.23		P	1.0

89 Y (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		24641599.08		A	4.7
2	☐	1.000		26083723.22		A	2.5
3	☐	1.000		25302451.84		A	0.5
4	☐	1.000		24732238.52		A	1.3
5	☐	1.000		24738976.85		A	1.4
6	☐	1.000		26037764.61		A	1.6
7	☐	1.000		26413337.11		A	1.4
8	☐	1.000		25280215.18		A	1.3

89 Y (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		3254280.92		A	0.1
2	☐	1.000		3229816.27		A	2.0
3	☐	1.000		3226176.27		A	0.2
4	☐	1.000		3111405.16		A	1.5
5	☐	1.000		3167947.32		A	1.3
6	☐	1.000		3285204.32		A	0.1
7	☐	1.000		3241058.15		A	2.4
8	☐	1.000		3322386.72		A	1.7

103 Rh (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		9796529.90		A	6.6
2	☐	1.000		10329337.30		A	2.3
3	☐	1.000		9929793.10		A	0.6
4	☐	1.000		9678317.14		A	0.8
5	☐	1.000		9667597.09		A	0.5
6	☐	1.000		10037799.10		A	2.0
7	☐	1.000		10132032.60		A	1.2
8	☐	1.000		9230672.25		A	0.9

103 Rh (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		5130037.11		A	1.4
2	☐	1.000		5081869.40		A	1.7
3	☐	1.000		5018725.13		A	0.5
4	☐	1.000		4851754.23		A	0.7
5	☐	1.000		4900067.74		A	0.8
6	☐	1.000		4977739.40		A	1.8
7	☐	1.000		4858854.51		A	2.0
8	☐	1.000		4551703.68		A	2.4

115 In (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		9874590.62		A	5.9
2	☐	1.000		10332692.94		A	3.5
3	☐	1.000		10137271.36		A	0.6
4	☐	1.000		9877239.50		A	1.4
5	☐	1.000		9789821.85		A	0.5
6	☐	1.000		10212111.03		A	2.8
7	☐	1.000		10459306.50		A	0.3
8	☐	1.000		9823824.57		A	1.9

115 In (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		1881073.44		A	0.6
2	☐	1.000		1880325.48		A	1.2
3	☐	1.000		1839585.07		A	2.7
4	☐	1.000		1791192.58		A	1.7
5	☐	1.000		1816614.84		A	0.8
6	☐	1.000		1864180.65		A	1.9
7	☐	1.000		1793434.56		A	0.2
8	☐	1.000		1806469.25		A	3.3

159 Tb (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		11430676.84		A	6.9
2	☐	1.000		11722126.08		A	4.2
3	☐	1.000		11676770.66		A	1.9
4	☐	1.000		11533638.86		A	0.8
5	☐	1.000		11610073.30		A	1.2
6	☐	1.000		12118894.40		A	1.1
7	☐	1.000		12628380.51		A	2.1
8	☐	1.000		11920615.93		A	0.3

159 Tb (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		5662754.15		A	1.1
2	☐	1.000		5715407.35		A	0.7
3	☐	1.000		5597320.89		A	3.4
4	☐	1.000		5519965.13		A	1.3
5	☐	1.000		5491570.41		A	1.0
6	☐	1.000		5774161.58		A	1.7
7	☐	1.000		5815056.65		A	1.2
8	☐	1.000		5686732.76		A	2.2

165 Ho (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		10788415.17		A	6.0
2	☐	1.000		11143105.02		A	0.8
3	☐	1.000		11052925.16		A	1.4
4	☐	1.000		10918219.25		A	2.0
5	☐	1.000		10822709.53		A	3.1
6	☐	1.000		11464398.78		A	2.2
7	☐	1.000		11683641.06		A	0.5
8	☐	1.000		11132029.33		A	1.5

165 Ho (ISTD) [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		5637713.46		A	1.1
2	☐	1.000		5776392.14		A	1.6
3	☐	1.000		5657853.66		A	2.9
4	☐	1.000		5548027.77		A	1.6
5	☐	1.000		5626969.15		A	1.2
6	☐	1.000		5849472.07		A	3.1
7	☐	1.000		5906253.73		A	1.0
8	☐	1.000		5694950.96		A	0.6

209 Bi (ISTD) [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		6205609.21		A	6.8
2	☐	1.000		6596573.23		A	4.0
3	☐	1.000		6414352.20		A	2.2
4	☐	1.000		6260861.85		A	0.4
5	☐	1.000		6286667.34		A	1.6
6	☐	1.000		6458549.14		A	0.9
7	☐	1.000		6586716.22		A	0.8
8	☐	1.000		5772624.63		A	1.1

209 Bi (ISTD) [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		3841000.40		A	2.5
2	☐	1.000		3910890.36		A	2.7
3	☐	1.000		3815697.55		A	0.2
4	☐	1.000		3717850.26		A	0.3
5	☐	1.000		3754789.35		A	2.3
6	☐	1.000		3857295.15		A	2.1
7	☐	1.000		3748245.15		A	0.3
8	☐	1.000		3463290.75		A	1.4

US EPA Tune Check Report

Operator Name Jaswal
Acq/Data Batch D:\Agilent\ICPMH\1\DATA\P8101221 MS.b
Acq. Date-Time 2021-10-12 10:16:02
Report Comment ---
Instrument Name G8403A SG19224459

[No Gas]

Sensitivity

Mass	Conc. [ug/l]	Count	CPS	Resp (Required) [cps/ug/l]	Resp (Flag)	RSD%	RSD% (Required)
9		14733	147330.45			1.116	5.000
24		127005	1270049.59			0.842	5.000
25		16878	168776.79			0.347	5.000
26		19784	197837.84			0.583	5.000
59		36670	366702.40			0.479	5.000
113		15659	156594.23			0.662	5.000
115		49971	499710.15			0.864	5.000
206		19559	195586.89			0.647	5.000
207		16183	161828.19			0.757	5.000
208		39946	399457.76			0.600	5.000
220		1	9.70			12.938	

Mass	RSD% (Flag)
9	
24	
25	
26	
59	
113	
115	
206	
207	
208	
220	


Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
9	14879	14728	14888	14488	14682
24	125683	127621	127038	126288	128395
25	16840	16894	16878	16811	16965
26	19768	19939	19806	19616	19790
59	36478	36711	36660	36562	36940
113	15760	15700	15547	15550	15741
115	49367	49812	49899	50362	50416
206	19555	19555	19356	19646	19682
207	16162	16264	15984	16296	16208
208	40014	39863	39603	39988	40261

US EPA Tune Check Report

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
220	1	1	1	1	1

Integration Time [sec] 0.1

Resolution/Axis



Mass	Peak Height	Axis	Axis (Required)	Axis (Flag)
9	25228.71	9.05	8.90 - 9.10	
24	225081.67	24.00	23.90 - 24.10	
25	29603.42	24.95	24.90 - 25.10	
26	34029.90	25.95	25.90 - 26.10	
59	66244.72	58.95	58.90 - 59.10	
113	29965.79	113.00	112.90 - 113.10	
115	96083.08	115.05	114.90 - 115.10	
206	35841.07	206.00	205.90 - 206.10	
207	29837.91	207.00	206.90 - 207.10	
208	74661.17	207.95	207.90 - 208.10	
220	1.45	219.85	-	

Mass	W-50%	W-5%	W-5% (Required)	W-5% (Flag)
9	0.61	0.781	0.900	
24	0.60	0.784	0.900	
25	0.60	0.780	0.900	
26	0.61	0.743	0.900	
59	0.59	0.769	0.900	
113	0.55	0.732	0.900	
115	0.54	0.766	0.900	
206	0.57	0.816	0.900	
207	0.57	0.816	0.900	
208	0.56	0.817	0.900	
220	0.36	1.770		

Integration Time [sec] 0.1

Acquisition Time [sec] 256.770000000002

Y Axis Linear

Tune Parameters

Plasma Parameters

Plasma Mode --- Nebulizer Gas 0.73 L/min Dilution Gas 0.40 L/min

US EPA Tune Check Report

RF Power	1600 W	Option Gas	---	Auxiliary Gas	0.90 L/min
RF Matching	1.40 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	10.0 mm	S/C Temp	2 °C		

Lens Parameters

Extract 1	-6.0 V	Omega Lens	9.7 V	Deflect	16.0 V
Extract 2	-205.0 V	Cell Entrance	-30 V	Plate Bias	-50 V
Omega Bias	-105 V	Cell Exit	-50 V		

Cell Parameters

Use Gas	No	3rd Gas Flow	---	Energy Discrimination	4.0 V
He Flow	0.0 mL/min	OctP Bias	-8.0 V		
H2 Flow	---	OctP RF	200 V		

QP Parameters

Mass Gain	123	Axis Gain	0.9993	QP Bias	-4.0 V
Mass Offset	125	Axis Offset	-0.02		

Hardware Settings

Torch

Torch H	0.9 mm	Torch V	-0.9 mm
---------	--------	---------	---------

EM

Discriminator	3.9 mV	Analog HV	2264 V	Pulse HV	1546 V
---------------	--------	-----------	--------	----------	--------

[He]

Sensitivity

Mass	Conc. [ug/l]	Count	CPS	Resp (Required) [cps/ug/l]	Resp (Flag)	RSD%	RSD% (Required)
59		8627	86265.40			3.825	
89		26928	269280.92			3.856	
205		17371	173706.41			4.169	

Mass	RSD% (Flag)
59	
89	
205	

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
59	8899	8411	8680	8962	8180
89	27666	26317	27187	28010	25460
205	18168	16933	17458	17910	16385

Integration Time [sec] 0.1

Tune Parameters

Plasma Parameters

Plasma Mode	---	Nebulizer Gas	0.75 L/min	Dilution Gas	0.40 L/min
RF Power	1600 W	Option Gas	---	Auxiliary Gas	0.90 L/min
RF Matching	1.40 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	10.0 mm	S/C Temp	2 °C		

Lens Parameters

Extract 1	-6.0 V	Omega Lens	9.7 V	Deflect	4.2 V
-----------	--------	------------	-------	---------	-------

US EPA Tune Check Report

Extract 2	-205.0 V	Cell Entrance	-50 V	Plate Bias	-60 V
Omega Bias	-105 V	Cell Exit	-70 V		
Cell Parameters					
Use Gas	Yes	3rd Gas Flow	---	Energy Discrimination	5.0 V
He Flow	4.5 mL/min	OctP Bias	-18.0 V		
H2 Flow	---	OctP RF	190 V		
QP Parameters					
Mass Gain	123	Axis Gain	0.9993	QP Bias	-13.0 V
Mass Offset	125	Axis Offset	-0.02		
Hardware Settings					
Torch					
Torch H	0.9 mm	Torch V	-0.9 mm		
EM					
Discriminator	3.9 mV	Analog HV	2264 V	Pulse HV	1546 V