



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116810

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/12/2021	16:10
2	CAL2	1	Water	NA	NA	20.3	7.00	10/12/2021	16:11
3	CAL3	1	Water	NA	NA	20.3	10.03	10/12/2021	16:12
4	ICV	1	Water	NA	NA	20.4	6.99	10/12/2021	16:15
5	CCV1	1	Water	NA	NA	20.3	2.01	10/12/2021	16:16
6	M4175-04	1	Water	NA	NA	20.3	7.69	10/12/2021	16:19
7	M4175-04DUP	1	Water	NA	NA	20.4	7.70	10/12/2021	16:20
8	CCV2	1	Water	NA	NA	20.4	12.02	10/12/2021	16:21

WORKLIST(Hardcopy Internal Chain)

116810

WorkList Name : PH M 4175 W WorkList ID : 153188 Department : Wet-Chemistry Date : 10-12-2021 15:57:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/12/2021	Water	M4175-04	pH	Cool 4 deg C	PSEG03	K12	41042	10/12/2021	9040C

Date/Time 10/12/21 16:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 10/12/21 17:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]