

LIB 116815


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

4428E



A= 0.0000e+000

B= 2.2530e-004 slope

C= 2.3678e-003 y-intercept

Rho= 0.9999376

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	90D
0.00	0.000	0.037	0.037	155	0.000	155					=
0.05	0.050					205					1
0.20	0.200	0.202	0.002	888	0.0 %	888					-2
2.50	2.500	2.453	-0.047	10878	0.0 %	10878					-1
5.00	5.000	4.959	-0.041	22001	0.0 %	22001					1
7.50	7.500	7.567	0.067	33578	0.0 %	33578					0
10.0	10.000	9.980	-0.020	44288	0.0 %	44288					

LB116815
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	155	-		0 SFAM01.1	PPB	10/13/2021 13:49	Std
	0.2 S01	888	-		0.2 SFAM01.1	PPB	10/13/2021 13:51	Std
	2.5 S02	10878	-		2.5 SFAM01.1	PPB	10/13/2021 13:53	Std
	5 S03	22001	-		5 SFAM01.1	PPB	10/13/2021 13:56	Std
	7.5 S04	33578	-		7.5 SFAM01.1	PPB	10/13/2021 13:58	Std
	10 S05	44288	-		10 SFAM01.1	PPB	10/13/2021 14:00	Std
ICV076	ICV076	17497	3.9444 -		SFAM01.1	PPB	10/13/2021 14:03	SMPL
ICB076	ICB076	90	0.0226 -		SFAM01.1	PPB	10/13/2021 14:06	SMPL
CCV044	CCV044	22394	5.0477 -		SFAM01.1	PPB	10/13/2021 14:08	SMPL
CCB044	CCB044	77	0.0197 -		SFAM01.1	PPB	10/13/2021 14:10	SMPL
PB139870BL	PBS870	45	0.0125 -		SFAM01.1	PPB	10/13/2021 14:15	SMPL
M4034-01	C00Q4	2376	0.5377 -		SFAM01.1	PPB	10/13/2021 14:17	SMPL
M4034-02	C00Q9	1830	0.4147 -		SFAM01.1	PPB	10/13/2021 14:19	SMPL
M4034-03	C00R0	1805	0.409 -		SFAM01.1	PPB	10/13/2021 14:22	SMPL
M4034-04	C00R0D	1805	0.409 -		SFAM01.1	PPB	10/13/2021 14:24	SMPL
M4034-05	C00R0S	13659	3.0797 -		SFAM01.1	PPB	10/13/2021 14:26	SMPL
M4034-06	C00R1	2601	0.5884 -		SFAM01.1	PPB	10/13/2021 14:28	SMPL
M4034-07	C00R2	818	0.1867 -		SFAM01.1	PPB	10/13/2021 14:31	SMPL
M4034-08	C00R3	1062	0.2416 -		SFAM01.1	PPB	10/13/2021 14:33	SMPL
M4034-09	C00R4	902	0.2056 -		SFAM01.1	PPB	10/13/2021 14:35	SMPL
M4034-10	C00R5	1068	0.243 -		SFAM01.1	PPB	10/13/2021 14:37	SMPL
M4034-11	C00R6	1349	0.3063 -		SFAM01.1	PPB	10/13/2021 14:40	SMPL
M4034-12	C00R7	1558	0.3534 -		SFAM01.1	PPB	10/13/2021 14:42	SMPL
M4034-13	C00R8	1354	0.3074 -		SFAM01.1	PPB	10/13/2021 14:44	SMPL
M4034-14	C00S5	980	0.2232 -		SFAM01.1	PPB	10/13/2021 14:46	SMPL
M4034-15	C00S6	567	0.1301 -		SFAM01.1	PPB	10/13/2021 14:49	SMPL
M4034-16	C00S7	1435	0.3257 -		SFAM01.1	PPB	10/13/2021 14:51	SMPL
M4034-17	C00S8	672	0.1538 -		SFAM01.1	PPB	10/13/2021 14:53	SMPL
M4034-18	C00S9	1406	0.3191 -		SFAM01.1	PPB	10/13/2021 14:56	SMPL
M4034-19	C00T0	1527	0.3464 -		SFAM01.1	PPB	10/13/2021 14:58	SMPL
CCV045	CCV045	22849	5.1502 -		SFAM01.1	PPB	10/13/2021 15:00	SMPL
CCB045	CCB045	110	0.0272 -		SFAM01.1	PPB	10/13/2021 15:02	SMPL
M4034-20	C00T1	340	0.079 -		SFAM01.1	PPB	10/13/2021 15:05	SMPL
PB139871BL	PBS871	137	0.0332 -		SFAM01.1	PPB	10/13/2021 15:07	SMPL
M4038-01	C00Q5	845	0.1927 -		SFAM01.1	PPB	10/13/2021 15:09	SMPL
M4038-02	C00Q5D	812	0.1853 -		SFAM01.1	PPB	10/13/2021 15:11	SMPL
M4038-03	C00Q5S	12903	2.9094 -		SFAM01.1	PPB	10/13/2021 15:14	SMPL
M4038-04	C00T9	525	0.1207 -		SFAM01.1	PPB	10/13/2021 15:16	SMPL
M4038-05	C00U0	468	0.1078 -		SFAM01.1	PPB	10/13/2021 15:18	SMPL
M4038-06	C00U5	4367	0.9863 -		SFAM01.1	PPB	10/13/2021 15:20	SMPL
M4038-07	C00U6	6629	1.4959 -		SFAM01.1	PPB	10/13/2021 15:23	SMPL
M4038-08	C00U7	449	0.1035 -		SFAM01.1	PPB	10/13/2021 15:25	SMPL
CCV046	CCV046	22755	5.1291 -		SFAM01.1	PPB	10/13/2021 15:28	SMPL
CCB046	CCB046	94	0.0235 -		SFAM01.1	PPB	10/13/2021 15:30	SMPL