



Analytical Summary Report

Analysis Method: 1010B  
Parameter: Flash Point  
Run Number: LB116819  
Thermometer ID: flashpoint

Reviewed By: ketankumar  
Supervisor Review By: apatel  
Ambient Barometric Pressure (mmHg): 760.00  
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/13/2021	10:45
2	M4175-04		1	10	100.00	>212.0	>212.0	10/13/2021	11:10
3	M4175-04DUP		1	10	100.00	>212.0	>212.0	10/13/2021	11:40

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

WORKLIST(Hardcopy Internal Chain)

LB116879

WorkList Name : fp101321      WorkList ID : 153208      Department : Wet-Chemistry      Date : 10-13-2021 09:01:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/19/2021	Water	M4175-04	Flash Point	Cool 4 deg C	PSEG03	K12	41042	10/12/2021	1010B

Date/Time 10/13/21 10:15  
 Raw Sample Received by: KS (WC)  
 Raw Sample Relinquished by: JO (WC)

Date/Time 10/13/21 11:45  
 Raw Sample Received by: JO (WC)  
 Raw Sample Relinquished by: KS (WC)