


LB 116823

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.7581e-004

C= 8.7438e-003

Rho= 0.9997477

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Y/D
0.0	0.000	0.023	0.023	52	0.000	52					-
0.2	0.200	0.202	0.002	699	0.0 %	699					1
2.5	2.500	2.479	-0.021	8957	0.0 %	8957					-1
5.0	5.000	4.909	-0.091	17767	0.0 %	17767					-2
7.5	7.500	7.662	0.162	27749	0.0 %	27749					2
10.0	10.000	9.925	-0.075	35952	0.0 %	35952					-1

LB116823 INSTRUMENT ID : CV1

Date of Analysis: 13 Oct 2021 15:41:36

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	52	-	0.00007470A	PPB		13 Oct 2021 15:42:24	S	Std
0.2 - 1	502	699	-	0.20007470A	PPB		13 Oct 2021 15:44:41	S	Std
2.5 - 1	525	8957	-	2.50007470A	PPB		13 Oct 2021 15:46:58	S	Std
5.0 - 1	55	17767	-	5.00007470A	PPB		13 Oct 2021 15:49:14	S	Std
7.5 - 1	575	27749	-	7.50007470A	PPB		13 Oct 2021 15:51:31	S	Std
10.0 - 1	510	35952	-	10.00007470A	PPB		13 Oct 2021 15:53:46	S	Std
ICV42 - 1	ICV42	13364	3.6947	-7470A	PPB		13 Oct 2021 15:56:32	U	SMPL
ICB42 - 1	ICB42	0	0.0087	-7470A	PPB		13 Oct 2021 15:58:47	U	SMPL
CCV47 - 1	CCV47	17674	4.8835	-7470A	PPB		13 Oct 2021 16:01:02	U	SMPL
CCB47 - 1	CCB47	-11	0.0057	-7470A	PPB		13 Oct 2021 16:03:17	U	SMPL
CRA - 1	CRA	656	0.1897	-7470A	PPB		13 Oct 2021 16:05:32	U	SMPL
HighStd - 1	HighStd	35525	9.8070	-7470A	PPB		13 Oct 2021 16:07:48	U	SMPL
ChkStd - 1	ChkStd	25349	7.0003	-7470A	PPB		13 Oct 2021 16:10:03	U	SMPL
PB139802BL - 1	PBW	-67	-0.0097	-7470A	PPB		13 Oct 2021 16:12:20	U	SMPL
PB139802BS - 1	LCSW	14498	4.0075	-7470A	PPB		13 Oct 2021 16:14:37	U	SMPL
M4158-21 - 1	T3-COMP-(1-4)	-31	0.0002	-7470A	PPB		13 Oct 2021 16:16:53	U	SMPL
M4158-22 - 1	T8-COMP-(1-4)	17	0.0134	-7470A	PPB		13 Oct 2021 16:19:09	U	SMPL
M4159-21 - 1	T9-COMP-(1-4)	21	0.0145	-7470A	PPB		13 Oct 2021 16:21:26	U	SMPL
M4159-22 - 1	T10-COMP-(1-4)	38	0.0192	-7470A	PPB		13 Oct 2021 16:23:43	U	SMPL
M4159-22DUP - 1	T10-COMP-(1-4)DUP	24	0.0154	-7470A	PPB		13 Oct 2021 16:25:59	U	SMPL
CCV48 - 1	CCV48	17447	4.8209	-7470A	PPB		13 Oct 2021 16:28:14	U	SMPL
CCB48 - 1	CCB48	-17	0.0041	-7470A	PPB		13 Oct 2021 16:30:29	U	SMPL
M4159-22MS - 1	T10-COMP-(1-4)MS	14062	3.8872	-7470A	PPB		13 Oct 2021 16:32:45	U	SMPL
M4159-22MSD - 1	T10-COMP-(1-4)MSD	12821	3.5449	-7470A	PPB		13 Oct 2021 16:35:00	U	SMPL
PB139802TB - 1	PB1239766TB	-10	0.0060	-7470A	PPB		13 Oct 2021 16:37:15	U	SMPL
M4159-22LX5 - 1		26	0.0159	-7470A	PPB		13 Oct 2021 16:39:31	U	SMPL
M4159-22A - 1		14357	3.9686	-7470A	PPB		13 Oct 2021 16:41:47	U	SMPL
CCV49 - 1	CCV49	17625	4.8700	-7470A	PPB		13 Oct 2021 16:44:04	U	SMPL
CCB49 - 1	CCB49	-14	0.0049	-7470A	PPB		13 Oct 2021 16:46:20	U	SMPL