

LB116825


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43383



A= 0.0000e+000

B= 2.3083e-004 slope

C= -3.4391e-003 y-intercept

Rho= 0.9999744

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.031	0.031	150	0.000	150					=
0.05	0.050					205					
0.20	0.200	0.212	0.012	934	0.0 %	934					6
2.50	2.500	2.450	-0.050	10629	0.0 %	10629					-2
5.00	5.000	4.983	-0.017	21603	0.0 %	21603					0
7.50	7.500	7.513	0.013	32563	0.0 %	32563					0
10.0	10.000	10.011	0.011	43383	0.0 %	43383					0

LB116825
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	150	-		0 SFAM01.1	PPB	10/13/2021 16:56	Std
	0.2 S01	934	-		0.2 SFAM01.1	PPB	10/13/2021 16:58	Std
	2.5 S02	10629	-		2.5 SFAM01.1	PPB	10/13/2021 17:01	Std
	5 S03	21603	-		5 SFAM01.1	PPB	10/13/2021 17:05	Std
	7.5 S04	32563	-		7.5 SFAM01.1	PPB	10/13/2021 17:07	Std
	10 S05	43383	-		10 SFAM01.1	PPB	10/13/2021 17:10	Std
ICV077	ICV077	17084	3.94 -		SFAM01.1	PPB	10/13/2021 17:12	SMPL
ICB077	ICB077	69	0.0125 -		SFAM01.1	PPB	10/13/2021 17:15	SMPL
CCV050	CCV050	21749	5.0168 -		SFAM01.1	PPB	10/13/2021 17:17	SMPL
CCB050	CCB050	61	0.0106 -		SFAM01.1	PPB	10/13/2021 17:19	SMPL
PB139872BL	PBS872	183	0.0388 -		SFAM01.1	PPB	10/13/2021 17:21	SMPL
M4074-01	JCF20	201489	46.5057 -		SFAM01.1	PPB	10/13/2021 17:24	SMPL
M4074-02	JCF21	21253	4.9023 -		SFAM01.1	PPB	10/13/2021 17:26	SMPL
M4074-03	JCF22	10199	2.3508 -		SFAM01.1	PPB	10/13/2021 17:28	SMPL
M4074-04	JCF23	7662	1.7652 -		SFAM01.1	PPB	10/13/2021 17:30	SMPL
M4074-05	JCF24	7899	1.8199 -		SFAM01.1	PPB	10/13/2021 17:33	SMPL
M4074-06	JCF25	6817	1.5701 -		SFAM01.1	PPB	10/13/2021 17:35	SMPL
M4074-07	JCF26	5213	1.1999 -		SFAM01.1	PPB	10/13/2021 17:37	SMPL
M4074-08	JCF27	8906	2.0523 -		SFAM01.1	PPB	10/13/2021 17:40	SMPL
M4074-09	JCF27D	8889	2.0484 -		SFAM01.1	PPB	10/13/2021 17:42	SMPL
M4074-10	JCF27S	22173	5.1147 -		SFAM01.1	PPB	10/13/2021 17:44	SMPL
M4074-11	JCF28	8065	1.8582 -		SFAM01.1	PPB	10/13/2021 17:46	SMPL
M4074-12	JCF29	73944	17.0648 -		SFAM01.1	PPB	10/13/2021 17:49	SMPL
M4074-13	JCF30	8207	1.891 -		SFAM01.1	PPB	10/13/2021 17:51	SMPL
M4074-14	JCF40	1984	0.4545 -		SFAM01.1	PPB	10/13/2021 17:53	SMPL
M4074-15	JCF41	2882	0.6618 -		SFAM01.1	PPB	10/13/2021 17:55	SMPL
M4074-16	JCF42	63067	14.5541 -		SFAM01.1	PPB	10/13/2021 17:58	SMPL
M4074-17	JCF61	74	0.0136 -		SFAM01.1	PPB	10/13/2021 18:00	SMPL
M4074-18	JCF62	160	0.0335 -		SFAM01.1	PPB	10/13/2021 18:02	SMPL
CCV051	CCV051	22247	5.1318 -		SFAM01.1	PPB	10/13/2021 18:07	SMPL
CCB051	CCB051	71	0.0129 -		SFAM01.1	PPB	10/13/2021 18:09	SMPL