



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116842

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	10/14/2021	15:25
2	CAL2	1	Water	NA	NA	20.6	7.00	10/14/2021	15:26
3	CAL3	1	Water	NA	NA	20.7	10.03	10/14/2021	15:27
4	ICV	1	Water	NA	NA	20.6	6.99	10/14/2021	15:30
5	CCV1	1	Water	NA	NA	20.6	2.01	10/14/2021	15:31
6	M4226-01	1	Water	NA	NA	22.7	7.82	10/14/2021	15:33
7	M4226-01DUP	1	Water	NA	NA	22.8	7.83	10/14/2021	15:34
8	M4226-03	1	Water	NA	NA	22.6	7.40	10/14/2021	15:36
9	M4230-01	1	Water	NA	NA	22.5	4.85	10/14/2021	15:40
10	CCV2	1	Water	NA	NA	20.9	12.02	10/14/2021	15:41