



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116845

Slope : 98.6

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	10/14/2021	15:30
2	CAL2	1	Water	NA	NA	20.6	7.00	10/14/2021	15:31
3	CAL3	1	Water	NA	NA	20.6	10.03	10/14/2021	15:32
4	ICV	1	Water	NA	NA	20.7	7.01	10/14/2021	15:35
5	CCV1	1	Water	NA	NA	20.6	2.01	10/14/2021	15:36
6	M4221-02	1	Solid	20.03	20	22.9	5.76	10/14/2021	15:40
7	M4221-02DUP	1	Solid	20.04	20	22.8	5.77	10/14/2021	15:41
8	M4224-01	1	Solid	20.03	20	23.1	7.61	10/14/2021	15:44
9	CCV2	1	Water	NA	NA	20.7	12.03	10/14/2021	15:45

VB116843

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph s m4221 WorkList ID : 153263 Department : Wet-Chemistry Date : 10-14-2021 15:19:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/14/2021	Solid	M4221-02	pH	Cool 4 deg C	PSEG03	K13	337	10/14/2021	9045D
10/14/2021	Solid	M4224-01	pH	Cool 4 deg C	ENV160	D11	515-2	10/14/2021	9045D

Date/Time 10/14/21 15:20
 Raw Sample Received by: JO WCC
 Raw Sample Relinquished by: JO WCC

Date/Time 10/14/21 16:30
 Raw Sample Received by: JO WCC
 Raw Sample Relinquished by: JO WCC