

LB116847


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45631



A= 0.0000e+000

B= 2.1928e-004

C= -4.8167e-002

Rho= 0.9998594

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/00
0.0	0.000	0.002	0.002	228	0.000	228					-16
0.2	0.200	0.168	-0.032	988	0.0 %	988					-2
2.5	2.500	2.459	-0.041	11436	0.0 %	11436					3
5.0	5.000	5.134	0.134	23635	0.0 %	23635					0
7.5	7.500	7.478	-0.022	34324	0.0 %	34324					0
10.0	10.000	9.958	-0.042	45631	0.0 %	45631					

LB116847

Method: 7471B

Operator: Admin

Date of Analysis: 15 Oct 2021 08:18:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	228	-	0.0000	7471B	PPB	15 Oct 2021 08:20:45	S	Std
0.2 - 1	50.2	988	-	0.2000	7471B	PPB	15 Oct 2021 08:23:02	S	Std
2.5 - 1	52.5	11436	-	2.5000	7471B	PPB	15 Oct 2021 08:25:19	S	Std
5.0 - 1	55	23635	-	5.0000	7471B	PPB	15 Oct 2021 08:27:36	S	Std
7.5 - 1	57.5	34324	-	7.5000	7471B	PPB	15 Oct 2021 08:29:52	S	Std
10.0 - 1	510	45631	-	10.0000	7471B	PPB	15 Oct 2021 08:34:42	S	Std
ICV44 - 1	ICV44	17658	3.8238	-	7471B	PPB	15 Oct 2021 08:37:26	U	SMPL
ICB44 - 1	ICB44	63	-0.0344	-	7471B	PPB	15 Oct 2021 08:39:40	U	SMPL
CCV60 - 1	CCV60	22987	4.9923	-	7471B	PPB	15 Oct 2021 08:41:56	U	SMPL
CCB60 - 1	CCB60	20	-0.0438	-	7471B	PPB	15 Oct 2021 08:44:11	U	SMPL
CRA - 1	CRA	1063	0.1849	-	7471B	PPB	15 Oct 2021 08:46:27	U	SMPL
HighStd - 1	HighStd	46013	10.0413	-	7471B	PPB	15 Oct 2021 08:48:42	U	SMPL
ChkStd - 1	ChkStd	32076	6.9853	-	7471B	PPB	15 Oct 2021 08:50:57	U	SMPL
PB139908BL - 1	PBS	-6	-0.0495	-	7471B	PPB	15 Oct 2021 08:58:49	U	SMPL
PB139908BS - 1	LCSS	17120	3.7058	-	7471B	PPB	15 Oct 2021 09:01:05	U	SMPL
M4175-01 - 1	30784	52	-0.0368	-	7471B	PPB	15 Oct 2021 09:03:22	U	SMPL
M4175-02 - 1	41173	712	0.1080	-	7471B	PPB	15 Oct 2021 09:05:39	U	SMPL
M4175-03 - 1	31123	446	0.0496	-	7471B	PPB	15 Oct 2021 09:24:03	U	SMPL
M4178-01 - 1	CL-02-101221	10873	2.3360	-	7471B	PPB	15 Oct 2021 09:26:20	U	SMPL
M4178-01DUP - 1	CL-02-101221DUP	10259	2.2014	-	7471B	PPB	15 Oct 2021 09:28:34	U	SMPL
CCV61 - 1	CCV61	24566	5.3386	-	7471B	PPB	15 Oct 2021 09:30:50	U	SMPL
CCB61 - 1	CCB61	18	-0.0442	-	7471B	PPB	15 Oct 2021 09:33:05	U	SMPL
M4178-01MS - 1	CL-02-101221MS	27485	5.9786	-	7471B	PPB	15 Oct 2021 09:39:44	U	SMPL
M4178-01MSD - 1	CL-02-101221MSD	28016	6.0951	-	7471B	PPB	15 Oct 2021 09:41:59	U	SMPL
M4180-01 - 1	NYPD-69TH-WC-1	16858	3.6484	-	7471B	PPB	15 Oct 2021 09:44:14	U	SMPL
M4191-01 - 1	COMP-4	2695	0.5428	-	7471B	PPB	15 Oct 2021 09:46:31	U	SMPL
M4191-02 - 1	COMP-5	2574	0.5162	-	7471B	PPB	15 Oct 2021 09:48:47	U	SMPL
M4191-03 - 1	COMP-6	1802	0.3470	-	7471B	PPB	15 Oct 2021 09:51:03	U	SMPL
M4193-01 - 1	351	106	-0.0249	-	7471B	PPB	15 Oct 2021 09:53:19	U	SMPL
M4194-01 - 1	VNJ-239	12884	2.7770	-	7471B	PPB	15 Oct 2021 09:55:36	U	SMPL
M4194-03 - 1	VNJ-209	12190	2.6248	-	7471B	PPB	15 Oct 2021 09:57:53	U	SMPL
M4195-01 - 1	HAR-SOILPILE	14854	3.2089	-	7471B	PPB	15 Oct 2021 10:00:09	U	SMPL
CCV62 - 1	CCV62	24849	5.4006	-	7471B	PPB	15 Oct 2021 10:02:24	U	SMPL
CCB62 - 1	CCB62	87	-0.0291	-	7471B	PPB	15 Oct 2021 10:04:40	U	SMPL
M4178-01LX5 - 1		2104	0.4132	-	7471B	PPB	15 Oct 2021 10:06:56	U	SMPL
M4178-01A - 1		29551	6.4316	-	7471B	PPB	15 Oct 2021 10:09:12	U	SMPL
CCV63 - 1	CCV63	25137	5.4638	-	7471B	PPB	15 Oct 2021 10:11:27	U	SMPL
CCB63 - 1	CCB63	121	-0.0216	-	7471B	PPB	15 Oct 2021 10:13:44	U	SMPL