

LB116851


7470A

INSTRUMENT ID: C12

Linear

 μ Abs.:

36945



A= 0.0000e+000

B= 2.6882e-004

C= -7.1996e-002

Rho= 0.9996853

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.066	-0.066	21	0.000	21					-5
0.2	0.200	0.190	-0.010	974	0.0 %	974					-1
2.5	2.500	2.487	-0.013	9521	0.0 %	9521					3
5.0	5.000	5.134	0.134	19366	0.0 %	19366					1
7.5	7.500	7.595	0.095	28522	0.0 %	28522					-1
10.0	10.000	9.860	-0.140	36945	0.0 %	36945					

LB116851 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 15 Oct 2021 12:46:23					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	21	-	0.0000	7470A	PPB	15 Oct 2021 12:47:42	S	Std
0.2 - 1	50.2	974	-	0.2000	7470A	PPB	15 Oct 2021 12:49:59	S	Std
2.5 - 1	52.5	9521	-	2.5000	7470A	PPB	15 Oct 2021 12:52:16	S	Std
5.0 - 1	55	19366	-	5.0000	7470A	PPB	15 Oct 2021 12:54:33	S	Std
7.5 - 1	57.5	28522	-	7.5000	7470A	PPB	15 Oct 2021 12:56:49	S	Std
10.0 - 1	510	36945	-	10.0000	7470A	PPB	15 Oct 2021 12:59:05	S	Std
ICV45 - 1	ICV45	14703	3.8805	-	7470A	PPB	15 Oct 2021 13:01:53	U	SMPL
ICB45 - 1	ICB45	-28	-0.0795	-	7470A	PPB	15 Oct 2021 13:04:08	U	SMPL
CCV64 - 1	CCV64	17858	4.7286	-	7470A	PPB	15 Oct 2021 13:06:23	U	SMPL
CCB64 - 1	CCB64	-23	-0.0782	-	7470A	PPB	15 Oct 2021 13:08:38	U	SMPL
CRA - 1	CRA	775	0.1363	-	7470A	PPB	15 Oct 2021 13:10:53	U	SMPL
HighStd - 1	HighStd	36021	9.6113	-	7470A	PPB	15 Oct 2021 13:13:08	U	SMPL
ChkStd - 1	ChkStd	23332	6.2002	-	7470A	PPB	15 Oct 2021 13:15:23	U	SMPL
PB139935BL - 1	PBW	-105	-0.1002	-	7470A	PPB	15 Oct 2021 13:17:40	U	SMPL
PB139935BS - 1	LCSW	12478	3.2824	-	7470A	PPB	15 Oct 2021 13:19:56	U	SMPL
M4226-03 - 1	TRE-1247	-71	-0.0911	-	7470A	PPB	15 Oct 2021 13:22:12	U	SMPL
M4226-03DUP - 1	TRE-1247DUP	-4	-0.0731	-	7470A	PPB	15 Oct 2021 13:24:28	U	SMPL
M4226-03MS - 1	TRE-1247MS	12840	3.3797	-	7470A	PPB	15 Oct 2021 13:26:46	U	SMPL
M4226-03MSD - 1	TRE-1247MSD	12364	3.2517	-	7470A	PPB	15 Oct 2021 13:29:02	U	SMPL
CCV65 - 1	CCV65	18496	4.9001	-	7470A	PPB	15 Oct 2021 13:31:18	U	SMPL
CCB65 - 1	CCB65	-41	-0.0830	-	7470A	PPB	15 Oct 2021 13:33:33	U	SMPL
PB139935TB - 1	PB139910TB	-4	-0.0731	-	7470A	PPB	15 Oct 2021 13:35:49	U	SMPL
M4226-03LX5 - 1		15	-0.0680	-	7470A	PPB	15 Oct 2021 13:38:04	U	SMPL
M4226-03A - 1		14807	3.9085	-	7470A	PPB	15 Oct 2021 13:40:19	U	SMPL
CCV66 - 1	CCV66	19175	5.0827	-	7470A	PPB	15 Oct 2021 13:42:35	U	SMPL
CCB66 - 1	CCB66	-121	-0.1045	-	7470A	PPB	15 Oct 2021 13:44:51	U	SMPL