


LB116857

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.3677e-004 *slope*C= -3.8887e-002 *y-intercept*

Rho= 0.9997563

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.031	-0.031	33	0.000	33					-
0.05	0.050					205					-2
0.20	0.200	0.196	-0.004	993	0.0 %	993					2
2.50	2.500	2.552	0.052	10943	0.0 %	10943					2
5.00	5.000	5.080	0.080	21618	0.0 %	21618					-2
7.50	7.500	7.340	-0.160	31163	0.0 %	31163					-2
10.0	10.000	10.064	0.064	42668	0.0 %	42668					1

LB116857
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	33	-		0 SFAM01.1	PPB	10/15/2021 14:54	Std
	0.2 S01	993	-		0.2 SFAM01.1	PPB	10/15/2021 14:56	Std
	2.5 S02	10943	-		2.5 SFAM01.1	PPB	10/15/2021 14:58	Std
	5 S03	21618	-		5 SFAM01.1	PPB	10/15/2021 15:01	Std
	7.5 S04	31163	-		7.5 SFAM01.1	PPB	10/15/2021 15:03	Std
	10 S05	42668	-		10 SFAM01.1	PPB	10/15/2021 15:08	Std
ICV079	ICV079	16511	3.8704	-	SFAM01.1	PPB	10/15/2021 15:11	SMPL
ICB079	ICB079	-71	-0.0557	-	SFAM01.1	PPB	10/15/2021 15:13	SMPL
CCV067	CCV067	21150	4.9688	-	SFAM01.1	PPB	10/15/2021 15:15	SMPL
CCB067	CCB067	-89	-0.06	-	SFAM01.1	PPB	10/15/2021 15:18	SMPL
PB139942BL	PBW942	-4	-0.0398	-	SFAM01.1	PPB	10/15/2021 15:20	SMPL
M4004-01	C00Q0	588	0.1003	-	SFAM01.1	PPB	10/15/2021 15:22	SMPL
M4004-02	C00Q2	-35	-0.0472	-	SFAM01.1	PPB	10/15/2021 15:24	SMPL
M4004-05	C00Q6	-21	-0.0439	-	SFAM01.1	PPB	10/15/2021 15:27	SMPL
M4004-06	C00Q8	-9	-0.041	-	SFAM01.1	PPB	10/15/2021 15:29	SMPL
M4004-07	C00R9	-19	-0.0434	-	SFAM01.1	PPB	10/15/2021 15:31	SMPL
M4004-08	C00S1	-19	-0.0434	-	SFAM01.1	PPB	10/15/2021 15:33	SMPL
M4004-09	C00S3	-8	-0.0408	-	SFAM01.1	PPB	10/15/2021 15:36	SMPL
M4004-10	C00T2	-20	-0.0436	-	SFAM01.1	PPB	10/15/2021 15:38	SMPL
M4004-11	C00S3D	-3	-0.0396	-	SFAM01.1	PPB	10/15/2021 15:40	SMPL
M4004-12	C00S3S	4367	0.9951	-	SFAM01.1	PPB	10/15/2021 15:43	SMPL
PB139943BL	PBW943	0	-0.0389	-	SFAM01.1	PPB	10/15/2021 15:45	SMPL
M4006-01	C00Q1	327	0.0385	-	SFAM01.1	PPB	10/15/2021 15:47	SMPL
M4006-02	C00Q3	-46	-0.0498	-	SFAM01.1	PPB	10/15/2021 15:49	SMPL
M4006-03	C00Q7	-11	-0.0415	-	SFAM01.1	PPB	10/15/2021 15:52	SMPL
M4006-04	C00S0	4	-0.0379	-	SFAM01.1	PPB	10/15/2021 15:54	SMPL
M4006-05	C00S2	0	-0.0389	-	SFAM01.1	PPB	10/15/2021 15:56	SMPL
M4006-06	C00S4	6	-0.0375	-	SFAM01.1	PPB	10/15/2021 15:58	SMPL
M4006-07	C00S2D	-12	-0.0417	-	SFAM01.1	PPB	10/15/2021 16:01	SMPL
M4006-08	C00S2S	4810	1.1	-	SFAM01.1	PPB	10/15/2021 16:03	SMPL
CCV068	CCV068	20565	4.8303	-	SFAM01.1	PPB	10/15/2021 16:05	SMPL
CCB068	CCB068	-62	-0.0536	-	SFAM01.1	PPB	10/15/2021 16:07	SMPL
PB139944BL	PBW944	-2	-0.0394	-	SFAM01.1	PPB	10/15/2021 16:10	SMPL
M4016-01	MH1Q40	-19	-0.0434	-	SFAM01.1	PPB	10/15/2021 16:12	SMPL
M4016-02	MH1Q41	71	-0.0221	-	SFAM01.1	PPB	10/15/2021 16:14	SMPL
M4016-03	MH1Q42	32	-0.0313	-	SFAM01.1	PPB	10/15/2021 16:17	SMPL
M4016-04	MH1Q43	-25	-0.0448	-	SFAM01.1	PPB	10/15/2021 16:19	SMPL
M4016-05	MH1Q44	-2	-0.0394	-	SFAM01.1	PPB	10/15/2021 16:21	SMPL
M4016-06	MH1Q45	28	-0.0323	-	SFAM01.1	PPB	10/15/2021 16:23	SMPL
M4016-07	MH1Q46	13	-0.0358	-	SFAM01.1	PPB	10/15/2021 16:26	SMPL
M4016-08	MH1Q47	-5	-0.0401	-	SFAM01.1	PPB	10/15/2021 16:28	SMPL
M4016-09	MH1Q48	-18	-0.0431	-	SFAM01.1	PPB	10/15/2021 16:30	SMPL
M4016-10	MH1Q49	10	-0.0365	-	SFAM01.1	PPB	10/15/2021 16:32	SMPL
M4016-11	MH1Q50	0	-0.0389	-	SFAM01.1	PPB	10/15/2021 16:35	SMPL
M4016-12	MH1Q51	-14	-0.0422	-	SFAM01.1	PPB	10/15/2021 16:37	SMPL
M4016-13	MH1Q52	-2	-0.0394	-	SFAM01.1	PPB	10/15/2021 16:39	SMPL

LB116857

INSTRUMENT ID : CV1

CCV069	CCV069	20255	4.7569 -	SFAM01.1	PPB	10/15/2021 16:52 SMPL
CCB069	CCB069	-62	-0.0536 -	SFAM01.1	PPB	10/15/2021 16:54 SMPL