



Analytical Summary Report

Analysis Method: 9040C

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116860

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.02	10/15/2021	17:00
2	CAL2	1	Water	NA	NA	20.4	7.00	10/15/2021	17:01
3	CAL3	1	Water	NA	NA	20.6	10.02	10/15/2021	17:02
4	ICV	1	Water	NA	NA	20.5	6.99	10/15/2021	17:03
5	CCV1	1	Water	NA	NA	20.5	2.01	10/15/2021	17:04
6	M4253-03	1	Water	NA	NA	24.0	7.92	10/15/2021	17:05
7	M4253-03DUP	1	Water	NA	NA	24.0	7.90	10/15/2021	17:06
8	M4253-04	1	Water	NA	NA	25.2	8.04	10/15/2021	17:07
9	M4253-05	1	Water	NA	NA	24.3	6.54	10/15/2021	17:08
10	M4253-07	1	Water	NA	NA	23.7	6.56	10/15/2021	17:09
11	M4253-11	1	Water	NA	NA	23.8	6.47	10/15/2021	17:10
12	CCV2	1	Water	NA	NA	20.8	12.01	10/15/2021	17:11

WORKLIST(Hardcopy Internal Chain)

LB116860

WorkList Name : PH101521

WorkList ID : 153311

Department : Wet-Chemistry

Date : 10-15-2021 16:54:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/15/2021	Water	M4253-03	pH	Cool 4 deg C	PSEG03	K31	143	10/15/2021	9040C
10/15/2021	Water	M4253-04	pH	Cool 4 deg C	PSEG03	K31	129	10/15/2021	9040C
10/15/2021	Water	M4253-05	pH	Cool 4 deg C	PSEG03	K31	135-136	10/15/2021	9040C
10/15/2021	Water	M4253-07	pH	Cool 4 deg C	PSEG03	K31	133-134	10/15/2021	9040C
10/15/2021	Water	M4253-11	pH	Cool 4 deg C	PSEG03	K31	137-138	10/15/2021	9040C

Date/Time 10/15/21 16:55
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JA (WC)

Date/Time 10/15/21 17:12
 Raw Sample Received by: JA (WC)
 Raw Sample Relinquished by: KS (WC)