



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116861

Slope : 98.8

BalanceID:

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.02	10/15/2021	17:00
2	CAL2	1	Water	NA	NA	20.4	7.00	10/15/2021	17:01
3	CAL3	1	Water	NA	NA	20.6	10.02	10/15/2021	17:02
4	ICV	1	Water	NA	NA	20.5	6.99	10/15/2021	17:03
5	CCV1	1	Water	NA	NA	20.4	2.02	10/15/2021	17:13
6	M4253-01	1	Solid	20	20	23.3	6.52	10/15/2021	17:14
7	M4253-01DUP	1	Solid	20	20	23.3	6.50	10/15/2021	17:15
8	M4256-01	1	Solid	20	20	24.0	6.24	10/15/2021	17:16
9	CCV2	1	Water	NA	NA	20.5	12.02	10/15/2021	17:17

WORKLIST(Hardcopy Internal Chain)

LB 116861

WorkList Name : PH101521-1 WorkList ID : 153312 Department : Wet-Chemistry Date : 10-15-2021 16:55:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/15/2021	Solid	M4253-01	pH	Cool 4 deg C	PSEG03	K31	131	10/15/2021	9045D
10/14/2021	Solid	M4256-01	pH	Cool 4 deg C	LOCK01	K31	454BASIN-20211014	10/14/2021	9045D

Date/Time 10/15/21 17:00
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 10/15/21 17:20
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)