



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116862

Slope : 98.8

BalanceID:

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.02	10/15/2021	17:00
2	CAL2	1	Water	NA	NA	20.4	7.00	10/15/2021	17:01
3	CAL3	1	Water	NA	NA	20.6	10.02	10/15/2021	17:02
4	ICV	1	Water	NA	NA	20.5	6.99	10/15/2021	17:03
5	CCV1	1	Water	NA	NA	20.5	2.01	10/15/2021	17:21
6	M4259-02	1	Solid	20	20	22.5	8.21	10/15/2021	17:22
7	M4259-02DUP	1	Solid	20	20	22.5	8.20	10/15/2021	17:23
8	M4259-05	1	Solid	20	20	22.8	8.39	10/15/2021	17:24
9	M4259-08	1	Solid	20	20	22.9	7.06	10/15/2021	17:25
10	M4259-11	1	Solid	20	20	22.6	8.86	10/15/2021	17:26
11	M4260-02	1	Solid	20	20	22.8	8.07	10/15/2021	17:27
12	M4260-05	1	Solid	20	20	22.9	7.46	10/15/2021	17:28
13	CCV2	1	Water	NA	NA	20.4	12.02	10/15/2021	17:30

WORKLIST(Hardcopy Internal Chain)

LB116862

WorkList Name : CORRO101521 WorkList ID : 153313 Department : Wet-Chemistry Date : 10-15-2021 16:56:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/15/2021	Solid	M4259-02	Corrosivity	Cool 4 deg C	PSEG03	K21	WC-1-CS	10/15/2021	9045D
10/15/2021	Solid	M4259-05	Corrosivity	Cool 4 deg C	PSEG03	K21	WC-1-BS	10/15/2021	9045D
10/15/2021	Solid	M4259-08	Corrosivity	Cool 4 deg C	PSEG03	K21	WC-1-BD	10/15/2021	9045D
10/15/2021	Solid	M4259-11	Corrosivity	Cool 4 deg C	PSEG03	K21	WC-1-AS	10/15/2021	9045D
10/15/2021	Solid	M4260-02	Corrosivity	Cool 4 deg C	PSEG03	K31	TEST-DA-(2)	10/15/2021	9045D
10/15/2021	Solid	M4260-05	Corrosivity	Cool 4 deg C	PSEG03	K31	TEST-DB-(2)	10/15/2021	9045D

Date/Time 10/15/21 17:00
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 10/15/21 17:35
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)