

LB 116866


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38999



A= 0.0000e+000

B= 2.5731e-004

C= -1.1193e-001

Rho= 0.9997863

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.005	0.005	455	0.000	455					1
0.2	0.200	0.242	0.042	1375	0.0 %	1375					21
2.5	2.500	2.402	-0.098	9769	0.0 %	9769					-4
5.0	5.000	4.996	-0.004	19850	0.0 %	19850					0
7.5	7.500	7.632	0.132	30097	0.0 %	30097					2
10.0	10.000	9.923	-0.077	38999	0.0 %	38999					-1

LB116866 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Oct 2021 11:00:59

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	455	-	0.0000	7470A	PPB	16 Oct 2021 11:08:54	S	Std
0.2 - 1	50.2	1375	-	0.2000	7470A	PPB	16 Oct 2021 11:11:11	S	Std
2.5 - 1	52.5	9769	-	2.5000	7470A	PPB	16 Oct 2021 11:13:28	S	Std
5.0 - 1	55	19850	-	5.0000	7470A	PPB	16 Oct 2021 11:15:45	S	Std
7.5 - 1	57.5	30097	-	7.5000	7470A	PPB	16 Oct 2021 11:18:01	S	Std
10.0 - 1	510	38999	-	10.0000	7470A	PPB	16 Oct 2021 11:20:17	S	Std
ICV46 - 1	ICV46	14667	3.6621	-	7470A	PPB	16 Oct 2021 11:25:09	U	SMPL
ICB46 - 1	ICB46	321	-0.0293	-	7470A	PPB	16 Oct 2021 11:27:24	U	SMPL
CCV70 - 1	CCV70	19826	4.9896	-	7470A	PPB	16 Oct 2021 11:29:39	U	SMPL
CCB70 - 1	CCB70	311	-0.0319	-	7470A	PPB	16 Oct 2021 11:31:54	U	SMPL
CRA - 1	CRA	1293	0.2208	-	7470A	PPB	16 Oct 2021 11:34:09	U	SMPL
HighStd - 1	HighStd	36723	9.3374	-	7470A	PPB	16 Oct 2021 11:36:24	U	SMPL
ChkStd - 1	ChkStd	26328	6.6626	-	7470A	PPB	16 Oct 2021 11:38:39	U	SMPL
PB139951BL - 1	PBW	78	-0.0919	-	7470A	PPB	16 Oct 2021 11:46:53	U	SMPL
PB139951BS - 1	LCSW	17585	4.4129	-	7470A	PPB	16 Oct 2021 11:49:08	U	SMPL
M4251-01 - 1	P287-DITCH	1664	0.3162	-	7470A	PPB	16 Oct 2021 11:58:53	U	SMPL
M4251-01DUP - 1	P287-DITCHDUP	1496	0.2730	-	7470A	PPB	16 Oct 2021 12:01:09	U	SMPL
M4251-01MS - 1	P287-DITCHMS	14640	3.6551	-	7470A	PPB	16 Oct 2021 12:03:26	U	SMPL
M4251-01MSD - 1	P287-DITCHMSD	15899	3.9791	-	7470A	PPB	16 Oct 2021 12:08:21	U	SMPL
M4251-02 - 1	P287-DITCH	585	0.0386	-	7470A	PPB	16 Oct 2021 12:10:37	U	SMPL
CCV71 - 1	CCV71	19697	4.9564	-	7470A	PPB	16 Oct 2021 12:12:53	U	SMPL
CCB71 - 1	CCB71	376	-0.0152	-	7470A	PPB	16 Oct 2021 12:15:08	U	SMPL
M4251-03 - 1	SW-RR-03-(SA)	1228	0.2041	-	7470A	PPB	16 Oct 2021 12:20:37	U	SMPL
M4251-04 - 1	SW-RR-03-(SA)	354	-0.0208	-	7470A	PPB	16 Oct 2021 12:25:30	U	SMPL
M4251-05 - 1	SW-RR-01-(2.0)	820	0.0991	-	7470A	PPB	16 Oct 2021 12:30:26	U	SMPL
M4251-06 - 1	SW-RR-01-(2.0)	230	-0.0527	-	7470A	PPB	16 Oct 2021 12:35:09	U	SMPL
M4251-07 - 1	RR-54	390	-0.0116	-	7470A	PPB	16 Oct 2021 12:37:25	U	SMPL
M4251-08 - 1	RR-54	13	-0.1086	-	7470A	PPB	16 Oct 2021 12:39:41	U	SMPL
M4251-01LX5 - 1		695	0.0669	-	7470A	PPB	16 Oct 2021 12:41:57	U	SMPL
M4251-01A - 1		17699	4.4423	-	7470A	PPB	16 Oct 2021 12:44:14	U	SMPL
CCV72 - 1	CCV72	19775	4.9764	-	7470A	PPB	16 Oct 2021 12:46:31	U	SMPL
CCB72 - 1	CCB72	362	-0.0188	-	7470A	PPB	16 Oct 2021 12:48:48	U	SMPL