


LB116867

SFAM01.1

INSTRUMENT ID: CV4

Linear



A= 0.0000e+000
B= 2.5277e-004 *slope*
C= -6.8363e-003 *y-intercept*
Rho= 0.9997485
Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.005	-0.005	8	0.000	8					=
0.05	0.050					205	0				14
0.20	0.200	0.228	0.028	930	0.0 %	930					1
2.50	2.500	2.537	0.037	10064	0.0 %	10064					0
5.00	5.000	4.987	-0.013	19758	0.0 %	19758					-2
7.50	7.500	7.339	-0.161	29061	0.0 %	29061					1
10.0	10.000	10.113	0.113	40036	0.0 %	40036					

LB116867
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	8	-		0 SFAM01.1	PPB	10/16/2021 14:15	Std
	0.2 S01	930	-		0.2 SFAM01.1	PPB	10/16/2021 14:18	Std
	2.5 S02	10064	-		2.5 SFAM01.1	PPB	10/16/2021 14:20	Std
	5 S03	19758	-		5 SFAM01.1	PPB	10/16/2021 14:22	Std
	7.5 S04	29061	-		7.5 SFAM01.1	PPB	10/16/2021 14:25	Std
	10 S05	40036	-		10 SFAM01.1	PPB	10/16/2021 14:27	Std
ICV080	ICV080	15818	3.9915 -		SFAM01.1	PPB	10/16/2021 14:30	SMPL
ICB080	ICB080	-112	-0.0351 -		SFAM01.1	PPB	10/16/2021 14:32	SMPL
CCV070	CCV070	19962	5.039 -		SFAM01.1	PPB	10/16/2021 14:34	SMPL
CCB070	CCB070	-108	-0.0341 -		SFAM01.1	PPB	10/16/2021 14:36	SMPL
PB139953BL	PBW953	33	0.0015 -		SFAM01.1	PPB	10/16/2021 14:39	SMPL
M4015-01	MH1Q25	160	0.0336 -		SFAM01.1	PPB	10/16/2021 14:41	SMPL
M4015-02	MH1Q26	210	0.0462 -		SFAM01.1	PPB	10/16/2021 14:43	SMPL
M4015-03	MH1Q27	92	0.0164 -		SFAM01.1	PPB	10/16/2021 14:45	SMPL
M4015-04	MH1Q28	102	0.0189 -		SFAM01.1	PPB	10/16/2021 14:48	SMPL
M4015-05	MH1Q29	131	0.0263 -		SFAM01.1	PPB	10/16/2021 14:50	SMPL
M4015-06	MH1Q30	69	0.0106 -		SFAM01.1	PPB	10/16/2021 14:52	SMPL
M4015-07	MH1Q31	112	0.0215 -		SFAM01.1	PPB	10/16/2021 14:55	SMPL
M4015-08	MH1Q32	222	0.0493 -		SFAM01.1	PPB	10/16/2021 14:57	SMPL
M4015-09	MH1Q33	133	0.0268 -		SFAM01.1	PPB	10/16/2021 14:59	SMPL
M4015-10	MH1Q33D	120	0.0235 -		SFAM01.1	PPB	10/16/2021 15:01	SMPL
M4015-11	MH1Q33S	3453	0.866 -		SFAM01.1	PPB	10/16/2021 15:04	SMPL
M4015-12	MH1Q34	105	0.0197 -		SFAM01.1	PPB	10/16/2021 15:06	SMPL
M4015-13	MH1Q35	58	0.0078 -		SFAM01.1	PPB	10/16/2021 15:08	SMPL
M4015-14	MH1Q36	63	0.0091 -		SFAM01.1	PPB	10/16/2021 15:10	SMPL
M4015-15	MH1Q37	200	0.0437 -		SFAM01.1	PPB	10/16/2021 15:13	SMPL
M4015-16	MH1Q38	682	0.1656 -		SFAM01.1	PPB	10/16/2021 15:15	SMPL
PB139954BL	PBW954	71	0.0111 -		SFAM01.1	PPB	10/16/2021 15:17	SMPL
M4101-01	YAT34	195	0.0425 -		SFAM01.1	PPB	10/16/2021 15:19	SMPL
M4101-02	YAT34D	160	0.0336 -		SFAM01.1	PPB	10/16/2021 15:22	SMPL
CCV071	CCV071	19808	5.0001 -		SFAM01.1	PPB	10/16/2021 15:24	SMPL
CCB071	CCB071	-88	-0.0291 -		SFAM01.1	PPB	10/16/2021 15:26	SMPL
M4101-03	YAT34S	4021	1.0096 -		SFAM01.1	PPB	10/16/2021 15:29	SMPL
M4101-04	YAT36	268	0.0609 -		SFAM01.1	PPB	10/16/2021 15:31	SMPL
M4101-05	YAT38	82	0.0139 -		SFAM01.1	PPB	10/16/2021 15:33	SMPL
M4101-06	YAT28	89	0.0157 -		SFAM01.1	PPB	10/16/2021 15:35	SMPL
M4101-07	YAT30	154	0.0321 -		SFAM01.1	PPB	10/16/2021 15:38	SMPL
M4101-08	YAT20	134	0.027 -		SFAM01.1	PPB	10/16/2021 15:40	SMPL
M4101-09	YAT44	138	0.028 -		SFAM01.1	PPB	10/16/2021 15:42	SMPL
M4101-10	YAT32	154	0.0321 -		SFAM01.1	PPB	10/16/2021 15:44	SMPL
M4101-11	YAT42	79	0.0131 -		SFAM01.1	PPB	10/16/2021 15:47	SMPL
M4101-12	YAT58	87	0.0152 -		SFAM01.1	PPB	10/16/2021 15:49	SMPL
PB139955BL	PBW955	110	0.021 -		SFAM01.1	PPB	10/16/2021 15:51	SMPL
M4102-01	YAT35	4161	1.0449 -		SFAM01.1	PPB	10/16/2021 15:53	SMPL
M4102-02	YAT35D	3885	0.9752 -		SFAM01.1	PPB	10/16/2021 15:56	SMPL
M4102-03	YAT35S	8128	2.0477 -		SFAM01.1	PPB	10/16/2021 15:58	SMPL

INSTRUMENT ID : CV1

M4102-04	YAT37	3123	0.7826 -	SFAM01.1	PPB	10/16/2021 16:00 SMPL
M4102-05	YAT39	55	0.0071 -	SFAM01.1	PPB	10/16/2021 16:03 SMPL
M4102-06	YAT29	80	0.0134 -	SFAM01.1	PPB	10/16/2021 16:05 SMPL
M4102-07	YAT31	76	0.0124 -	SFAM01.1	PPB	10/16/2021 16:07 SMPL
M4102-08	YAT21	102	0.0189 -	SFAM01.1	PPB	10/16/2021 16:09 SMPL
M4102-09	YAT45	120	0.0235 -	SFAM01.1	PPB	10/16/2021 16:12 SMPL
CCV072	CCV072	19927	5.0301 -	SFAM01.1	PPB	10/16/2021 16:14 SMPL
CCB072	CCB072	-67	-0.0238 -	SFAM01.1	PPB	10/16/2021 16:16 SMPL
M4102-10	YAT33	145	0.0298 -	SFAM01.1	PPB	10/16/2021 16:18 SMPL
M4102-11	YAT57	76	0.0124 -	SFAM01.1	PPB	10/16/2021 16:21 SMPL
PB139956BL	PBW956	50	0.0058 -	SFAM01.1	PPB	10/16/2021 16:23 SMPL
M4114-01	C00U1	60	0.0083 -	SFAM01.1	PPB	10/16/2021 16:25 SMPL
M4114-02	C00U3	102	0.0189 -	SFAM01.1	PPB	10/16/2021 16:27 SMPL
M4114-03	C00U8	70	0.0109 -	SFAM01.1	PPB	10/16/2021 16:30 SMPL
M4114-04	C00U3D	78	0.0129 -	SFAM01.1	PPB	10/16/2021 16:32 SMPL
M4114-05	C00U3S	3973	0.9974 -	SFAM01.1	PPB	10/16/2021 16:34 SMPL
PB139957BL	PBW957	46	0.0048 -	SFAM01.1	PPB	10/16/2021 16:36 SMPL
M4115-01	C00U2	61	0.0086 -	SFAM01.1	PPB	10/16/2021 16:39 SMPL
M4115-02	C00U4	55	0.0071 -	SFAM01.1	PPB	10/16/2021 16:41 SMPL
M4115-03	C00U4D	61	0.0086 -	SFAM01.1	PPB	10/16/2021 16:43 SMPL
M4115-04	C00U4S	4272	1.073 -	SFAM01.1	PPB	10/16/2021 16:46 SMPL
CCV073	CCV073	19662	4.9632 -	SFAM01.1	PPB	10/16/2021 16:48 SMPL
CCB073	CCB073	-79	-0.0268 -	SFAM01.1	PPB	10/16/2021 16:50 SMPL