


LB116873

7471B INSTRUMENT ID: CV1

Linear

μ Abs.:

52713



A= 0.0000e+000

B= 1.9130e-004

C= -1.2477e-001

Rho= 0.9998501

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.060	-0.060	336	0.000	336					-
0.2	0.200	0.198	-0.002	1689	0.0 %	1689					-1
2.5	2.500	2.512	0.012	13782	0.0 %	13782					0
5.0	5.000	5.132	0.132	27479	0.0 %	27479					3
7.5	7.500	7.460	-0.040	39648	0.0 %	39648					-1
10.0	10.000	9.959	-0.041	52713	0.0 %	52713					0

LB116873 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 17 Oct 2021 12:11:05					
Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	336	-	0.00007471B	PPB		17 Oct 2021 12:17:09	S	Std
0.2 - 1	50.2	1689	-	0.20007471B	PPB		17 Oct 2021 12:19:26	S	Std
2.5 - 1	52.5	13782	-	2.50007471B	PPB		17 Oct 2021 12:21:43	S	Std
5.0 - 1	55	27479	-	5.00007471B	PPB		17 Oct 2021 12:24:00	S	Std
7.5 - 1	57.5	39648	-	7.50007471B	PPB		17 Oct 2021 12:26:17	S	Std
10.0 - 1	511	52713	-	10.00007471B	PPB		17 Oct 2021 12:33:43	S	Std
ICV47 - 1	ICV47	21917	4.0678	-7471B	PPB		17 Oct 2021 12:39:24	U	SMPL
ICB47 - 1	ICB47	424	-0.0437	-7471B	PPB		17 Oct 2021 12:41:38	U	SMPL
CCV74 - 1	CCV74	27658	5.1661	-7471B	PPB		17 Oct 2021 12:43:54	U	SMPL
CCB74 - 1	CCB74	401	-0.0481	-7471B	PPB		17 Oct 2021 12:46:09	U	SMPL
CRA - 1	CRA	1695	0.1995	-7471B	PPB		17 Oct 2021 12:48:25	U	SMPL
HighStd - 1	HighStd	53940	10.1937	-7471B	PPB		17 Oct 2021 12:53:12	U	SMPL
ChkStd - 1	ChkStd	37118	6.9757	-7471B	PPB		17 Oct 2021 12:55:26	U	SMPL
PB139958BL - 1	PBS	286	-0.0701	-7471B	PPB		17 Oct 2021 12:57:43	U	SMPL
PB139958BS - 1	LCSS	20684	3.8320	-7471B	PPB		17 Oct 2021 13:00:00	U	SMPL
M4222-01 - 1	TEST-A(1)	3133	0.4746	-7471B	PPB		17 Oct 2021 13:02:16	U	SMPL
M4227-01 - 1	RBR-21502-05	1687	0.1979	-7471B	PPB		17 Oct 2021 13:04:33	U	SMPL
M4240-01 - 1	OR-03-101421	2195	0.2951	-7471B	PPB		17 Oct 2021 13:06:50	U	SMPL
M4242-10 - 1	T1-COMP-(1-4)	5885	1.0010	-7471B	PPB		17 Oct 2021 13:09:07	U	SMPL
M4242-21 - 1	T2-COMP-(1-4)	4353	0.7079	-7471B	PPB		17 Oct 2021 13:11:22	U	SMPL
CCV75 - 1	CCV75	27478	5.1316	-7471B	PPB		17 Oct 2021 13:13:38	U	SMPL
CCB75 - 1	CCB75	396	-0.0490	-7471B	PPB		17 Oct 2021 13:15:53	U	SMPL
M4243-10 - 1	T3-COMP-(1-4)	5387	0.9057	-7471B	PPB		17 Oct 2021 13:18:08	U	SMPL
M4243-21 - 1	T4-COMP-(1-4)	4291	0.6961	-7471B	PPB		17 Oct 2021 13:20:23	U	SMPL
M4244-10 - 1	T5-COMP-(1-4)	15186	2.7802	-7471B	PPB		17 Oct 2021 13:22:39	U	SMPL
M4249-01 - 1	TP1-(S)	2047	0.2668	-7471B	PPB		17 Oct 2021 13:24:55	U	SMPL
M4249-04 - 1	TP1-(D)	688	0.0068	-7471B	PPB		17 Oct 2021 13:27:11	U	SMPL
M4249-07 - 1	TP2-(S)	3064	0.4614	-7471B	PPB		17 Oct 2021 13:29:28	U	SMPL
M4249-10 - 1	TP2-(D)	603	-0.0094	-7471B	PPB		17 Oct 2021 13:31:44	U	SMPL
M4249-10DUP - 1	TP2-(D)DUP	576	-0.0146	-7471B	PPB		17 Oct 2021 13:34:01	U	SMPL
M4249-10MS - 1	TP2-(D)MS	19652	3.6346	-7471B	PPB		17 Oct 2021 13:36:18	U	SMPL
M4249-10MSD - 1	TP2-(D)MSD	19758	3.6548	-7471B	PPB		17 Oct 2021 13:38:34	U	SMPL
CCV76 - 1	CCV76	26822	5.0062	-7471B	PPB		17 Oct 2021 13:40:49	U	SMPL
CCB76 - 1	CCB76	371	-0.0538	-7471B	PPB		17 Oct 2021 13:43:05	U	SMPL
PB139959BL - 1	PBS	439	-0.0408	-7471B	PPB		17 Oct 2021 13:45:21	U	SMPL
PB139959BS - 1	LCSS	21007	3.8938	-7471B	PPB		17 Oct 2021 13:47:36	U	SMPL
M4248-01 - 1	CHVB-0513	761	0.0208	-7471B	PPB		17 Oct 2021 13:49:52	U	SMPL
M4248-03 - 1	72-12002	2939	0.4374	-7471B	PPB		17 Oct 2021 13:52:07	U	SMPL
M4252-01 - 1	TEST-BA-(1)	2382	0.3309	-7471B	PPB		17 Oct 2021 13:54:23	U	SMPL
M4252-04 - 1	TEST-BB-(1)	1351	0.1337	-7471B	PPB		17 Oct 2021 13:56:40	U	SMPL
M4252-07 - 1	MH-CA-(1)	4759	0.7856	-7471B	PPB		17 Oct 2021 13:58:56	U	SMPL
M4252-10 - 1	MH-CB-(1)	2478	0.3493	-7471B	PPB		17 Oct 2021 14:01:13	U	SMPL
M4256-01 - 1	454BASIN-20211014	1548	0.1714	-7471B	PPB		17 Oct 2021 14:03:30	U	SMPL
M4259-01 - 1	WC-1-CS	14665	2.6806	-7471B	PPB		17 Oct 2021 14:05:46	U	SMPL
CCV77 - 1	CCV77	26877	5.0167	-7471B	PPB		17 Oct 2021 14:08:01	U	SMPL
CCB77 - 1	CCB77	385	-0.0511	-7471B	PPB		17 Oct 2021 14:10:16	U	SMPL
M4259-04 - 1	WC-1-BS	9711	1.7329	-7471B	PPB		17 Oct 2021 14:12:32	U	SMPL
M4259-07 - 1	WC-1-BD	1169	0.0989	-7471B	PPB		17 Oct 2021 14:14:47	U	SMPL
M4259-10 - 1	WC-1-AS	26447	4.9344	-7471B	PPB		17 Oct 2021 14:17:03	U	SMPL
M4259-10DUP - 1	WC-1-ASDUP	26417	4.9287	-7471B	PPB		17 Oct 2021 14:21:54	U	SMPL
M4259-10MS - 1	WC-1-ASMS	45634	8.6048	-7471B	PPB		17 Oct 2021 14:24:10	U	SMPL
M4259-10MSD - 1	WC-1-ASMSD	47868	9.0322	-7471B	PPB		17 Oct 2021 14:29:22	U	SMPL
M4260-01 - 1	TEST-DA-(1)	1357	0.1348	-7471B	PPB		17 Oct 2021 14:31:38	U	SMPL
M4260-04 - 1	TEST-DB-(1)	2432	0.3405	-7471B	PPB		17 Oct 2021 14:33:56	U	SMPL
CCV78 - 1	CCV78	27283	5.0943	-7471B	PPB		17 Oct 2021 14:36:14	U	SMPL
CCB78 - 1	CCB78	374	-0.0532	-7471B	PPB		17 Oct 2021 14:38:30	U	SMPL
M4249-10LX5 - 1		442	-0.0402	-7471B	PPB		17 Oct 2021 14:40:46	U	SMPL
M4249-10A - 1		21316	3.9529	-7471B	PPB		17 Oct 2021 14:43:01	U	SMPL
M4259-10LX5 - 1		5375	0.9034	-7471B	PPB		17 Oct 2021 14:45:17	U	SMPL
M4259-10A - 1		47684	8.9970	-7471B	PPB		17 Oct 2021 14:47:32	U	SMPL
CCV79 - 1	CCV79	27188	5.0762	-7471B	PPB		17 Oct 2021 14:49:48	U	SMPL
CCB79 - 1	CCB79	372	-0.0536	-7471B	PPB		17 Oct 2021 14:52:05	U	SMPL