


LB116281

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.8087e-004

C= -4.5285e-002

Rho= 0.9998986

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Mod
0.0	0.000	-0.044	-0.044	3	0.000	3					
0.2	0.200	0.194	-0.006	852	0.0 %	852					
2.5	2.500	2.501	0.001	9066	0.0 %	9066					
5.0	5.000	5.111	0.111	18358	0.0 %	18358					
7.5	7.500	7.483	-0.017	26804	0.0 %	26804					
10.0	10.000	9.955	-0.045	35607	0.0 %	35607					

Mod

-3

0

0

0

0

LB116881 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 18 Oct 2021 11:55:03					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	3	-	0.0000	7470A	PPB	18 Oct 2021 11:59:29	S	Std
0.2 - 1	50.2	852	-	0.2000	7470A	PPB	18 Oct 2021 12:01:46	S	Std
2.5 - 1	52.5	9066	-	2.5000	7470A	PPB	18 Oct 2021 12:04:03	S	Std
5.0 - 1	55	18358	-	5.0000	7470A	PPB	18 Oct 2021 12:06:20	S	Std
7.5 - 1	57.5	26804	-	7.5000	7470A	PPB	18 Oct 2021 12:13:26	S	Std
10.0 - 1	59	35607	-	10.0000	7470A	PPB	18 Oct 2021 12:15:42	S	Std
ICV48 - 1	ICV48	14198	3.9424	-	7470A	PPB	18 Oct 2021 12:18:30	U	SMPL
ICB48 - 1	ICB48	-142	-0.0852	-	7470A	PPB	18 Oct 2021 12:20:45	U	SMPL
CCV85 - 1	CCV85	18782	5.2299	-	7470A	PPB	18 Oct 2021 12:23:00	U	SMPL
CCB85 - 1	CCB85	-150	-0.0874	-	7470A	PPB	18 Oct 2021 12:25:15	U	SMPL
CRA - 1	CRA	827	0.1870	-	7470A	PPB	18 Oct 2021 12:27:30	U	SMPL
HighStd - 1	HighStd	36007	10.0678	-	7470A	PPB	18 Oct 2021 12:29:45	U	SMPL
ChkStd - 1	ChkStd	25414	7.0926	-	7470A	PPB	18 Oct 2021 12:32:00	U	SMPL
PB139964BL - 1	PBW	-152	-0.0880	-	7470A	PPB	18 Oct 2021 12:34:17	U	SMPL
PB139964BS - 1	LCSW	13258	3.6784	-	7470A	PPB	18 Oct 2021 12:36:34	U	SMPL
M4222-02 - 1	TEST-A(2)	-114	-0.0773	-	7470A	PPB	18 Oct 2021 12:38:52	U	SMPL
M4224-01 - 1	515-2	-13	-0.0489	-	7470A	PPB	18 Oct 2021 12:41:10	U	SMPL
M4242-11 - 1	T1-COMP-(1-4)	-3	-0.0461	-	7470A	PPB	18 Oct 2021 12:43:27	U	SMPL
M4242-22 - 1	T2-COMP-(1-4)	30	-0.0369	-	7470A	PPB	18 Oct 2021 12:45:44	U	SMPL
M4243-11 - 1	T3-COMP-(1-4)	14	-0.0414	-	7470A	PPB	18 Oct 2021 12:47:59	U	SMPL
CCV86 - 1	CCV86	19045	5.3038	-	7470A	PPB	18 Oct 2021 12:50:14	U	SMPL
CCB86 - 1	CCB86	-124	-0.0801	-	7470A	PPB	18 Oct 2021 12:52:29	U	SMPL
M4243-22 - 1	T4-COMP-(1-4)	-11	-0.0484	-	7470A	PPB	18 Oct 2021 12:54:45	U	SMPL
M4244-11 - 1	T5-COMP-(1-4)	2	-0.0447	-	7470A	PPB	18 Oct 2021 12:57:00	U	SMPL
M4249-02 - 1	TP1-(S)	68	-0.0262	-	7470A	PPB	18 Oct 2021 12:59:15	U	SMPL
M4249-05 - 1	TP1-(D)	157	-0.0012	-	7470A	PPB	18 Oct 2021 13:01:31	U	SMPL
M4249-08 - 1	TP2-(S)	70	-0.0256	-	7470A	PPB	18 Oct 2021 13:03:46	U	SMPL
M4249-11 - 1	TP2-(D)	34	-0.0357	-	7470A	PPB	18 Oct 2021 13:06:03	U	SMPL
M4252-02 - 1	TEST-BA-(2)	68	-0.0262	-	7470A	PPB	18 Oct 2021 13:08:19	U	SMPL
M4252-05 - 1	TEST-BB-(2)	71	-0.0253	-	7470A	PPB	18 Oct 2021 13:10:36	U	SMPL
M4252-08 - 1	MH-CA-(2)	79	-0.0231	-	7470A	PPB	18 Oct 2021 13:12:53	U	SMPL
M4252-11 - 1	MH-CB-(2)	38	-0.0346	-	7470A	PPB	18 Oct 2021 13:15:08	U	SMPL
CCV87 - 1	CCV87	18914	5.2670	-	7470A	PPB	18 Oct 2021 13:17:24	U	SMPL
CCB87 - 1	CCB87	-114	-0.0773	-	7470A	PPB	18 Oct 2021 13:19:39	U	SMPL
M4253-01 - 1	131	-5	-0.0467	-	7470A	PPB	18 Oct 2021 13:21:55	U	SMPL
M4260-02 - 1	TEST-DA-(2)	-21	-0.0512	-	7470A	PPB	18 Oct 2021 13:24:11	U	SMPL
M4260-05 - 1	TEST-DB-(2)	-4	-0.0464	-	7470A	PPB	18 Oct 2021 13:26:26	U	SMPL
M4260-05DUP - 1	TEST-DB-(2)DUP	-10	-0.0481	-	7470A	PPB	18 Oct 2021 13:28:42	U	SMPL
M4260-05MS - 1	TEST-DB-(2)MS	13817	3.8354	-	7470A	PPB	18 Oct 2021 13:30:57	U	SMPL
M4260-05MSD - 1	TEST-DB-(2)MSD	13934	3.8683	-	7470A	PPB	18 Oct 2021 13:33:14	U	SMPL
PB139965BL - 1	PBW	-107	-0.0753	-	7470A	PPB	18 Oct 2021 13:35:30	U	SMPL
PB139965BS - 1	LCSW	13093	3.6321	-	7470A	PPB	18 Oct 2021 13:37:47	U	SMPL
M4253-03 - 1	143	-37	-0.0557	-	7470A	PPB	18 Oct 2021 13:40:04	U	SMPL
M4253-04 - 1	129	27	-0.0377	-	7470A	PPB	18 Oct 2021 13:42:19	U	SMPL
CCV88 - 1	CCV88	19188	5.3440	-	7470A	PPB	18 Oct 2021 13:44:35	U	SMPL
CCB88 - 1	CCB88	-162	-0.0908	-	7470A	PPB	18 Oct 2021 13:46:50	U	SMPL
M4253-05 - 1	135-136	-24	-0.0520	-	7470A	PPB	18 Oct 2021 13:49:06	U	SMPL
M4253-07 - 1	133-134	-18	-0.0503	-	7470A	PPB	18 Oct 2021 13:51:22	U	SMPL
M4253-11 - 1	137-138	-15	-0.0495	-	7470A	PPB	18 Oct 2021 13:53:37	U	SMPL
M4253-11DUP - 1	137-138DUP	-8	-0.0475	-	7470A	PPB	18 Oct 2021 13:55:53	U	SMPL
M4253-11MS - 1	137-138MS	13355	3.7057	-	7470A	PPB	18 Oct 2021 13:58:09	U	SMPL
M4253-11MSD - 1	137-138MSD	13951	3.8731	-	7470A	PPB	18 Oct 2021 14:00:25	U	SMPL
PB139964TB - 1	PB139930TB	-106	-0.0751	-	7470A	PPB	18 Oct 2021 14:02:41	U	SMPL
PB139965TB - 1	PB139950TB	-23	-0.0517	-	7470A	PPB	18 Oct 2021 14:04:58	U	SMPL
CCV89 - 1	CCV89	19137	5.3296	-	7470A	PPB	18 Oct 2021 14:07:15	U	SMPL
CCB89 - 1	CCB89	-125	-0.0804	-	7470A	PPB	18 Oct 2021 14:09:30	U	SMPL
M4260-05LX5 - 1		-6	-0.0470	-	7470A	PPB	18 Oct 2021 14:11:45	U	SMPL
M4260-05A - 1		14520	4.0329	-	7470A	PPB	18 Oct 2021 14:14:01	U	SMPL
M4253-11LX5 - 1		-66	-0.0638	-	7470A	PPB	18 Oct 2021 14:16:17	U	SMPL
M4253-11A - 1		14477	4.0208	-	7470A	PPB	18 Oct 2021 14:21:31	U	SMPL
CCV90 - 1	CCV90	19407	5.4055	-	7470A	PPB	18 Oct 2021 14:23:47	U	SMPL
CCB90 - 1	CCB90	-133	-0.0826	-	7470A	PPB	18 Oct 2021 14:26:02	U	SMPL