

LB 116891


SFAM01.1

INSTRUMENT ID: C12

Linear

μ Abs.:

45783



A= 0.0000e+000

B= 2.1831e-004 slope

C= -5.9844e-002 y-intercept

Rho= 0.9999333

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/p)
0.00	0.000	-0.039	-0.039	95	0.000	95					=
0.05	0.050					205	0				-6
0.20	0.200	0.188	-0.012	1135	0.0 %	1135					1
2.50	2.500	2.532	0.032	11872	0.0 %	11872					1
5.00	5.000	5.028	0.028	23306	0.0 %	23306					1
7.50	7.500	7.556	0.056	34888	0.0 %	34888					1
10.0	10.000	9.935	-0.065	45783	0.0 %	45783					-1

LB116891
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	95	-		0 SFAM01.1	PPB	10/18/2021 15:35	Std
	0.2 S01	1135	-		0.2 SFAM01.1	PPB	10/18/2021 15:37	Std
	2.5 S02	11872	-		2.5 SFAM01.1	PPB	10/18/2021 15:39	Std
	5 S03	23306	-		5 SFAM01.1	PPB	10/18/2021 15:41	Std
	7.5 S04	34888	-		7.5 SFAM01.1	PPB	10/18/2021 15:44	Std
	10 S05	45783	-		10 SFAM01.1	PPB	10/18/2021 15:46	Std
ICV082	ICV082	19166	4.1242 -		SFAM01.1	PPB	10/18/2021 15:49	SMPL
ICB082	ICB082	27	-0.0539 -		SFAM01.1	PPB	10/18/2021 15:51	SMPL
CCV091	CCV091	22712	4.8983 -		SFAM01.1	PPB	10/18/2021 15:53	SMPL
CCB091	CCB091	20	-0.0555 -		SFAM01.1	PPB	10/18/2021 15:56	SMPL
PB140000BL	PBS000	36	-0.052 -		SFAM01.1	PPB	10/18/2021 15:58	SMPL
M4216-01	JDGB7	331	0.0124 -		SFAM01.1	PPB	10/18/2021 16:00	SMPL
M4216-02	JDGB8	433	0.0347 -		SFAM01.1	PPB	10/18/2021 16:02	SMPL
M4216-03	JDGE4	358	0.0183 -		SFAM01.1	PPB	10/18/2021 16:05	SMPL
M4216-04	JDGE4D	361	0.019 -		SFAM01.1	PPB	10/18/2021 16:07	SMPL
M4216-05	JDGE4S	12736	2.7205 -		SFAM01.1	PPB	10/18/2021 16:09	SMPL
M4216-06	JDGC5	116	-0.0345 -		SFAM01.1	PPB	10/18/2021 16:11	SMPL
M4216-07	JDGC6	126	-0.0323 -		SFAM01.1	PPB	10/18/2021 16:14	SMPL
M4216-08	JDGC7	231	-0.0094 -		SFAM01.1	PPB	10/18/2021 16:16	SMPL
M4216-09	JDGC8	135	-0.0304 -		SFAM01.1	PPB	10/18/2021 16:18	SMPL
M4216-10	JDGC9	114	-0.035 -		SFAM01.1	PPB	10/18/2021 16:20	SMPL
M4216-11	JDGD8	106	-0.0367 -		SFAM01.1	PPB	10/18/2021 16:23	SMPL
M4216-12	JDGD9	201	-0.016 -		SFAM01.1	PPB	10/18/2021 16:25	SMPL
M4216-13	JDGE0	77	-0.043 -		SFAM01.1	PPB	10/18/2021 16:27	SMPL
M4216-14	JDGE8	137	-0.0299 -		SFAM01.1	PPB	10/18/2021 16:29	SMPL
M4216-15	JDGF0	116	-0.0345 -		SFAM01.1	PPB	10/18/2021 16:32	SMPL
M4216-16	JDGF3	388	0.0249 -		SFAM01.1	PPB	10/18/2021 16:34	SMPL
M4216-17	JDGF4	255	-0.0042 -		SFAM01.1	PPB	10/18/2021 16:36	SMPL
M4216-18	JDGF5	123	-0.033 -		SFAM01.1	PPB	10/18/2021 16:39	SMPL
M4216-19	JDGF6	69	-0.0448 -		SFAM01.1	PPB	10/18/2021 16:41	SMPL
CCV092	CCV092	22167	4.7793 -		SFAM01.1	PPB	10/18/2021 16:43	SMPL
CCB092	CCB092	-10	-0.062 -		SFAM01.1	PPB	10/18/2021 16:45	SMPL
M4216-20	JDGF7	64	-0.0459 -		SFAM01.1	PPB	10/18/2021 16:48	SMPL
PB140001BL	PBS001	5	-0.0588 -		SFAM01.1	PPB	10/18/2021 16:50	SMPL
M4219-01	JDGB5	169	-0.0229 -		SFAM01.1	PPB	10/18/2021 16:52	SMPL
M4219-02	JDGB6	98	-0.0384 -		SFAM01.1	PPB	10/18/2021 16:54	SMPL
M4219-03	JDGC1	249	-0.0055 -		SFAM01.1	PPB	10/18/2021 16:57	SMPL
M4219-04	JDGC2	231	-0.0094 -		SFAM01.1	PPB	10/18/2021 16:59	SMPL
M4219-05	JDGC3	258	-0.0035 -		SFAM01.1	PPB	10/18/2021 17:01	SMPL
M4219-06	JDGE3	412	0.0301 -		SFAM01.1	PPB	10/18/2021 17:03	SMPL
M4219-07	JDGE3D	428	0.0336 -		SFAM01.1	PPB	10/18/2021 17:06	SMPL
M4219-08	JDGE3S	12068	2.5747 -		SFAM01.1	PPB	10/18/2021 17:08	SMPL
M4219-09	JDGF9	439	0.036 -		SFAM01.1	PPB	10/18/2021 17:10	SMPL
M4219-10	JDGG0	443	0.0369 -		SFAM01.1	PPB	10/18/2021 17:13	SMPL
M4219-11	JDGE5	114	-0.035 -		SFAM01.1	PPB	10/18/2021 17:15	SMPL
M4219-12	JDGE6	81	-0.0422 -		SFAM01.1	PPB	10/18/2021 17:17	SMPL

LB116891
INSTRUMENT ID : CV1

M4219-13	JDGE7	77	-0.043 -	SFAM01.1	PPB	10/18/2021 17:19 SMPL
M4219-14	JDGF8	676	0.0877 -	SFAM01.1	PPB	10/18/2021 17:22 SMPL
M4219-15	JDGG1	988	0.1558 -	SFAM01.1	PPB	10/18/2021 17:24 SMPL
M4219-16	JDGD3	117	-0.0343 -	SFAM01.1	PPB	10/18/2021 17:26 SMPL
CCV093	CCV093	21858	4.7119 -	SFAM01.1	PPB	10/18/2021 17:28 SMPL
CCB093	CCB093	-14	-0.0629 -	SFAM01.1	PPB	10/18/2021 17:31 SMPL