



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116893

Slope : 98.3

BalanceID: WC SC-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/18/2021	15:30
2	CAL2	1	Water	NA	NA	20.3	7.00	10/18/2021	15:31
3	CAL3	1	Water	NA	NA	20.3	10.03	10/18/2021	15:32
4	ICV	1	Water	NA	NA	20.4	6.99	10/18/2021	15:35
5	CCV1	1	Water	NA	NA	20.4	2.01	10/18/2021	15:36
6	M4269-03	1	Solid	20.02	20	20.5	7.80	10/18/2021	16:00
7	M4269-03DUP	1	Solid	20.01	20	20.6	7.81	10/18/2021	16:01
8	M4269-05	1	Solid	20.03	20	20.3	7.30	10/18/2021	16:05
9	M4269-07	1	Solid	20.01	20	20.2	6.07	10/18/2021	16:10
10	M4269-09	1	Solid	20.03	20	20.4	6.03	10/18/2021	16:13
11	M4269-11	1	Solid	20.05	20	20.2	6.88	10/18/2021	16:15
12	M4269-13	1	Solid	20.04	20	20.4	9.12	10/18/2021	16:20
13	M4269-15	1	Solid	20.02	20	20.6	5.81	10/18/2021	16:22
14	CCV2	1	Water	NA	NA	20.4	12.02	10/18/2021	16:25

WORKLIST(Hardcopy Internal Chain)

28 116893

WorkList Name : ph m4269 WorkList ID : 153376 Department : Wet-Chemistry Date : 10-18-2021 16:00:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/18/2021	Solid	M4269-03	pH	Cool 4 deg C	PSEG03	K31	RT-2664	10/18/2021	9045D
10/18/2021	Solid	M4269-05	pH	Cool 4 deg C	PSEG03	K31	ETGI-355	10/18/2021	9045D
10/18/2021	Solid	M4269-07	pH	Cool 4 deg C	PSEG03	K31	VNJ-220	10/18/2021	9045D
10/18/2021	Solid	M4269-09	pH	Cool 4 deg C	PSEG03	K31	ETGI-365	10/18/2021	9045D
10/18/2021	Solid	M4269-11	pH	Cool 4 deg C	PSEG03	K31	CHRT-25062	10/18/2021	9045D
10/18/2021	Solid	M4269-13	pH	Cool 4 deg C	PSEG03	K31	RT-2929	10/18/2021	9045D
10/18/2021	Solid	M4269-15	pH	Cool 4 deg C	PSEG03	K31	CHRT-25259	10/18/2021	9045D

Date/Time 10/18/21 16:10
 Raw Sample Received by: SP (WC)
 Raw Sample Relinquished by: JT (SM)

Date/Time 10/18/21 17:35
 Raw Sample Received by: JT (SM)
 Raw Sample Relinquished by: SP (WC)