

LB116899


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

37924



A= 0.0000e+000

B= 2.6647e-004

C= 4.0513e-002

Rho= 0.9990156

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0D
0.0	0.000	0.045	0.045	17	0.000	17					17
0.2	0.200	0.234	0.034	727	0.0 %	727					4
2.5	2.500	2.397	-0.103	8842	0.0 %	8842					4
5.0	5.000	5.182	0.182	19295	0.0 %	19295					4
7.5	7.500	7.196	-0.304	26853	0.0 %	26853					4
10.0	10.000	10.146	0.146	37924	0.0 %	37924					1

LB116897 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Oct 2021 11:23:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	17	-	0.0000	7470A	PPB	19 Oct 2021 11:26:48	S	Std
0.2 - 1	50.2	727	-	0.2000	7470A	PPB	19 Oct 2021 11:29:05	S	Std
2.5 - 1	52.5	8842	-	2.5000	7470A	PPB	19 Oct 2021 11:31:22	S	Std
5.0 - 1	59	19295	-	5.0000	7470A	PPB	19 Oct 2021 11:33:39	S	Std
7.5 - 1	57.5	26853	-	7.5000	7470A	PPB	19 Oct 2021 11:35:56	S	Std
10.0 - 1	510	37924	-	10.0000	7470A	PPB	19 Oct 2021 11:43:19	S	Std
ICV49 - 1	ICV49	14321	3.8566	-	7470A	PPB	19 Oct 2021 11:46:09	U	SMPL
ICB49 - 1	ICB49	-98	0.0144	-	7470A	PPB	19 Oct 2021 11:48:23	U	SMPL
CCV94 - 1	CCV94	19144	5.1418	-	7470A	PPB	19 Oct 2021 11:50:38	U	SMPL
CCB94 - 1	CCB94	-106	0.0123	-	7470A	PPB	19 Oct 2021 11:52:53	U	SMPL
CRA - 1	CRA	688	0.2238	-	7470A	PPB	19 Oct 2021 11:55:09	U	SMPL
HighStd - 1	HighStd	37597	10.0589	-	7470A	PPB	19 Oct 2021 11:57:23	U	SMPL
ChkStd - 1	ChkStd	25890	6.9394	-	7470A	PPB	19 Oct 2021 11:59:38	U	SMPL
PB140009BL - 1	PBW	-169	-0.0045	-	7470A	PPB	19 Oct 2021 12:01:55	U	SMPL
PB140009BS - 1	LCSW	14989	4.0346	-	7470A	PPB	19 Oct 2021 12:04:11	U	SMPL
M4268-01 - 1	P257-DITCH	810	0.2564	-	7470A	PPB	19 Oct 2021 12:09:04	U	SMPL
M4268-01DUP - 1	P257-DITCHDUP	718	0.2318	-	7470A	PPB	19 Oct 2021 12:13:50	U	SMPL
M4268-01MS - 1	P257-DITCHMS	14337	3.8609	-	7470A	PPB	19 Oct 2021 12:16:06	U	SMPL
M4268-01MSD - 1	P257-DITCHMSD	14668	3.9491	-	7470A	PPB	19 Oct 2021 12:18:24	U	SMPL
M4268-02 - 1	P257-DITCH	68	0.0586	-	7470A	PPB	19 Oct 2021 12:20:39	U	SMPL
CCV95 - 1	CCV95	19111	5.1330	-	7470A	PPB	19 Oct 2021 12:22:54	U	SMPL
CCB95 - 1	CCB95	-32	0.0320	-	7470A	PPB	19 Oct 2021 12:25:10	U	SMPL
M4268-03 - 1	SW-RR-03-(SA)	45	0.0525	-	7470A	PPB	19 Oct 2021 12:27:25	U	SMPL
M4268-04 - 1	SW-RR-03-(SA)	-26	0.0336	-	7470A	PPB	19 Oct 2021 12:29:41	U	SMPL
M4268-05 - 1	SW-RR-01-(2.0)	-19	0.0355	-	7470A	PPB	19 Oct 2021 12:31:57	U	SMPL
M4268-06 - 1	SW-RR-01-(2.0)	-22	0.0347	-	7470A	PPB	19 Oct 2021 12:34:12	U	SMPL
M4268-01LX5 - 1		151	0.0807	-	7470A	PPB	19 Oct 2021 12:36:28	U	SMPL
M4268-01A - 1		14834	3.9933	-	7470A	PPB	19 Oct 2021 12:38:45	U	SMPL
CCV96 - 1	CCV96	18961	5.0930	-	7470A	PPB	19 Oct 2021 12:41:01	U	SMPL
CCB96 - 1	CCB96	-50	0.0272	-	7470A	PPB	19 Oct 2021 12:43:18	U	SMPL