



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116908

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	10/18/2021	09:40
2	M4253-03		1	12	100.00	>212.0	>212.0	10/18/2021	10:10
3	M4253-03DUP		1	12	100.00	>212.0	>212.0	10/18/2021	10:35
4	M4253-04		1	10	100.00	>212.0	>212.0	10/18/2021	10:58
5	M4253-05		1	13	100.00	>212.0	>212.0	10/18/2021	11:35
6	M4253-07		1	12	100.00	>212.0	>212.0	10/18/2021	12:05
7	M4253-11		1	10	100.00	>212.0	>212.0	10/18/2021	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP101921 WorkList ID : 153414 Department : Wet-Chemistry Date : 10-18-2021 08:55:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/22/2021	Water	M4253-03	Flash Point	Cool 4 deg C	PSEG03	K31	143	10/15/2021	1010B
10/22/2021	Water	M4253-04	Flash Point	Cool 4 deg C	PSEG03	K31	129	10/15/2021	1010B
10/22/2021	Water	M4253-05	Flash Point	Cool 4 deg C	PSEG03	K31	135-136	10/15/2021	1010B
10/22/2021	Water	M4253-07	Flash Point	Cool 4 deg C	PSEG03	K31	133-134	10/15/2021	1010B
10/22/2021	Water	M4253-11	Flash Point	Cool 4 deg C	PSEG03	K31	137-138	10/15/2021	1010B

Date/Time 10/18/21 09:15
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ c EXL-1ab)

Date/Time 10/18/21 12:45
 Raw Sample Received by: RJ c EXL-1ab)
 Raw Sample Relinquished by: KS (WC)