



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116909

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	10/18/2021	14:05
2	M4253-14		1	4	20.00	68.0	68.3	10/18/2021	14:32
3	M4253-14DUP		1	4	19.00	66.2	66.5	10/18/2021	14:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB116909

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP101921-1 WorkList ID : 153415 Department : Wet-Chemistry Date : 10-18 -2021 13:30:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/22/2021	Solid	M4253-14	Flash Point	Cool 4 deg C	PSEG03	K11	141	10/15/2021	1010B

Date/Time 10/18/21 13:35
Raw Sample Received by: JS (WC)
Raw Sample Relinquished by: JS (WC)

Date/Time 10/18/21 15:00
Raw Sample Received by: JS (WC)
Raw Sample Relinquished by: JS (WC)