

LB116910


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

44645



A= 0.0000e+000

B= 2.2439e-004 slope

C= -2.6536e-002 y-intercept

Rho= 0.9999960

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.013	-0.013	60	0.000	60					
0.05	0.050					205	0				
0.20	0.200	0.208	0.008	1045	0.0 %	1045					
2.50	2.500	2.508	0.008	11296	0.0 %	11296					
5.00	5.000	4.991	-0.009	22362	0.0 %	22362					
7.50	7.500	7.514	0.014	33606	0.0 %	33606					
10.0	10.000	9.991	-0.009	44645	0.0 %	44645					

LB116910

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	60	-		0 SFAM01.1	PPB	10/19/2021 15:43	Std
	0.2 S01	1045	-		0.2 SFAM01.1	PPB	10/19/2021 15:45	Std
	2.5 S02	11296	-		2.5 SFAM01.1	PPB	10/19/2021 15:48	Std
	5 S03	22362	-		5 SFAM01.1	PPB	10/19/2021 15:50	Std
	7.5 S04	33606	-		7.5 SFAM01.1	PPB	10/19/2021 15:52	Std
	10 S05	44645	-		10 SFAM01.1	PPB	10/19/2021 15:54	Std
ICV083	ICV083	18749	4.1806 -		SFAM01.1	PPB	10/19/2021 15:57	SMPL
ICB083	ICB083	-20	-0.031 -		SFAM01.1	PPB	10/19/2021 15:59	SMPL
CCV097	CCV097	22539	5.031 -		SFAM01.1	PPB	10/19/2021 16:01	SMPL
CCB097	CCB097	-31	-0.0335 -		SFAM01.1	PPB	10/19/2021 16:04	SMPL
PB140037BL	PBS037	32	-0.0194 -		SFAM01.1	PPB	10/19/2021 16:06	SMPL
M4185-01	C0B38	1956	0.4124 -		SFAM01.1	PPB	10/19/2021 16:08	SMPL
M4185-02	C0B45	1035	0.2057 -		SFAM01.1	PPB	10/19/2021 16:10	SMPL
M4185-03	C0B51	899	0.1752 -		SFAM01.1	PPB	10/19/2021 16:13	SMPL
M4185-04	C0B52	1167	0.2353 -		SFAM01.1	PPB	10/19/2021 16:15	SMPL
M4185-05	C0B03	2238	0.4756 -		SFAM01.1	PPB	10/19/2021 16:17	SMPL
M4185-06	C0B03D	2255	0.4795 -		SFAM01.1	PPB	10/19/2021 16:20	SMPL
M4185-07	C0B03S	14091	3.1353 -		SFAM01.1	PPB	10/19/2021 16:22	SMPL
M4185-08	C0AY7	1182	0.2387 -		SFAM01.1	PPB	10/19/2021 16:24	SMPL
M4185-09	C0AZ7	4681	1.0238 -		SFAM01.1	PPB	10/19/2021 16:26	SMPL
M4185-10	C0B22	4590	1.0034 -		SFAM01.1	PPB	10/19/2021 16:29	SMPL
M4185-11	C0B23	686	0.1274 -		SFAM01.1	PPB	10/19/2021 16:31	SMPL
M4185-12	C0B33	2337	0.4979 -		SFAM01.1	PPB	10/19/2021 16:33	SMPL
M4185-13	MCOAC3	1688	0.3522 -		SFAM01.1	PPB	10/19/2021 16:35	SMPL
PB140038BL	PBS038	5	-0.0254 -		SFAM01.1	PPB	10/19/2021 16:38	SMPL
M4220-01	JDGE9	99	-0.0043 -		SFAM01.1	PPB	10/19/2021 16:40	SMPL
M4220-02	JDGE9D	129	0.0024 -		SFAM01.1	PPB	10/19/2021 16:42	SMPL
M4220-03	JDGE9S	12839	2.8544 -		SFAM01.1	PPB	10/19/2021 16:44	SMPL
M4220-04	JDGF1	146	0.0062 -		SFAM01.1	PPB	10/19/2021 16:47	SMPL
M4220-05	JDGF2	268	0.0336 -		SFAM01.1	PPB	10/19/2021 16:49	SMPL
CCV098	CCV098	22926	5.1178 -		SFAM01.1	PPB	10/19/2021 16:51	SMPL
CCB098	CCB098	-7	-0.0281 -		SFAM01.1	PPB	10/19/2021 16:54	SMPL