

Analysis Method: SM2510 BANALYST: apatelParameter: ConductanceSUPERVISOR REVIEW BY: ketankumarRun Number: LB116911Meter ID: WC Conductance Meter-1Thermometer ID: WC pH meter-1

Reagent/Standard	Lot/Log #
conductivity standard	WP92563

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB116911BL		1	21.1	0.75	0.81	10/19/2021	11:40
2	LB116911BS	1413	1	21.2	1311.00	1413.60	10/19/2021	11:42
3	M4276-01		1	15.9	1.12	1.36	10/19/2021	11:45
4	M4276-01DUP		1	16.0	1.13	1.36	10/19/2021	11:47

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--

WORKLIST(Hardcopy Internal Chain)

WorkList Name : cond-101921 WorkList ID : 153406 Department : Wet-Chemistry Date : 10-19-2021 11:33:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/01/2021	Water	M4276-01	Conductance	Cool 4 deg C	STER10	K11	POOL-WATER	10/11/2021	SM2510 B

Date/Time 10/19/21 11:35
Raw Sample Received by: AP (WOL)
Raw Sample Relinquished by: SP (WOL)

Date/Time 10/19/21 11:55
Raw Sample Received by: SP (WOL)
Raw Sample Relinquished by: SP (WOL)