



Analytical Summary Report

Analysis Method: 9040C

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116915

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12 Control Limits[+/- 0.1].

True Value of CCV3 = 2 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	10/19/2021	15:10
2	CAL2	1	Water	NA	NA	20.6	7.00	10/19/2021	15:12
3	CAL3	1	Water	NA	NA	20.7	10.02	10/19/2021	15:13
4	ICV	1	Water	NA	NA	20.6	6.99	10/19/2021	15:15
5	CCV1	1	Water	NA	NA	20.5	2.01	10/19/2021	15:16
6	M4276-01	1	Water	NA	NA	20.6	5.03	10/19/2021	15:17
7	M4276-01DUP	1	Water	NA	NA	20.6	5.04	10/19/2021	15:18
8	M4279-01	1	Water	NA	NA	20.3	6.33	10/19/2021	15:19
9	M4283-06	1	Water	NA	NA	20.4	6.90	10/19/2021	15:20
10	M4284-05	1	Water	NA	NA	18.3	10.30	10/19/2021	15:22
11	M4284-06	1	Water	NA	NA	17.4	6.72	10/19/2021	15:23
12	M4284-07	1	Water	NA	NA	17.5	6.64	10/19/2021	15:25
13	M4284-08	1	Water	NA	NA	17.4	6.46	10/19/2021	15:26
14	M4284-10	1	Water	NA	NA	18.7	6.50	10/19/2021	15:27
15	M4284-11	1	Water	NA	NA	18.7	6.72	10/19/2021	15:28
16	CCV2	1	Water	NA	NA	20.8	12.02	10/19/2021	15:30
17	M4284-12	1	Water	NA	NA	18.7	6.94	10/19/2021	15:31
18	M4284-13	1	Water	NA	NA	18.1	6.83	10/19/2021	15:32
19	M4284-15	1	Water	NA	NA	17.1	4.25	10/19/2021	15:33
20	M4284-16	1	Water	NA	NA	17.7	5.46	10/19/2021	15:34

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	M4284-17	1	Water	NA	NA	17.3	6.24	10/19/2021	15:35
22	CCV3	1	Water	NA	NA	20.5	2.02	10/19/2021	15:36

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH W M4284 WorkList ID : 153419 Department : Wet-Chemistry Date : 10-19-2021 15:02:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/11/2021	Water	M4276-01	pH	Cool 4 deg C	STER10	K11	POOL-WATER	10/11/2021	9040C
10/18/2021	Water	M4279-01	pH	Cool 4 deg C	STER01	D11	POOL-WATER	10/18/2021	9040C
10/19/2021	Water	M4283-06	pH	Cool 4 deg C	PSEG03	K31	GREEN-TENT-TOTE	10/19/2021	9040C
10/19/2021	Water	M4284-05	pH	Cool 4 deg C	PSEG03	K31	30851-WATER	10/19/2021	9040C
10/19/2021	Water	M4284-06	pH	Cool 4 deg C	PSEG03	K31	20878-WATER	10/19/2021	9040C
10/19/2021	Water	M4284-07	pH	Cool 4 deg C	PSEG03	K31	20876	10/19/2021	9040C
10/19/2021	Water	M4284-08	pH	Cool 4 deg C	PSEG03	K31	20877	10/19/2021	9040C
10/19/2021	Water	M4284-10	pH	Cool 4 deg C	PSEG03	K31	30853	10/19/2021	9040C
10/19/2021	Water	M4284-11	pH	Cool 4 deg C	PSEG03	K31	30854	10/19/2021	9040C
10/19/2021	Water	M4284-12	pH	Cool 4 deg C	PSEG03	K31	30855	10/19/2021	9040C
10/19/2021	Water	M4284-13	pH	Cool 4 deg C	PSEG03	K31	30856	10/19/2021	9040C
10/19/2021	Water	M4284-15	pH	Cool 4 deg C	PSEG03	K31	30848	10/19/2021	9040C
10/19/2021	Water	M4284-16	pH	Cool 4 deg C	PSEG03	K31	30849	10/19/2021	9040C
10/19/2021	Water	M4284-17	pH	Cool 4 deg C	PSEG03	K31	20890	10/19/2021	9040C

Date/Time 10-19-21 15:05
 Raw Sample Received by: JD (sm)
 Raw Sample Relinquished by: JD (sm)

Date/Time 10-19-21 15:40
 Raw Sample Received by: JD (sm)
 Raw Sample Relinquished by: JD (sm)