



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116916

Slope : 98.6

BalanceID:

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	10/19/2021	15:10
2	CAL2	1	Water	NA	NA	20.6	7.00	10/19/2021	15:12
3	CAL3	1	Water	NA	NA	20.7	10.02	10/19/2021	15:13
4	ICV	1	Water	NA	NA	20.6	6.99	10/19/2021	15:15
5	CCV1	1	Water	NA	NA	20.5	2.01	10/19/2021	15:42
6	M4280-01	1	Solid	20	20	19.8	6.77	10/19/2021	15:43
7	M4280-01DUP	1	Solid	20	20	19.8	6.78	10/19/2021	15:44
8	M4280-03	1	Solid	20	20	19.5	6.09	10/19/2021	15:45
9	M4280-05	1	Solid	20	20	19.6	8.17	10/19/2021	15:47
10	M4284-02	1	Solid	20	20	19.1	6.57	10/19/2021	15:49
11	M4284-03	1	Solid	20	20	19.6	8.17	10/19/2021	15:50
12	M4284-04	1	Solid	20	20	19.5	6.59	10/19/2021	15:51
13	CCV2	1	Water	NA	NA	20.6	12.02	10/19/2021	15:52

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S M4284 WorkList ID : 153420 Department : Wet-Chemistry Date : 10-19-2021 15:33:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/19/2021	Solid	M4280-01	pH	Cool 4 deg C	PSEG03	K31	72-11977	10/19/2021	9045D
10/19/2021	Solid	M4280-03	pH	Cool 4 deg C	PSEG03	K31	CHRT-24051	10/19/2021	9045D
10/19/2021	Solid	M4280-05	pH	Cool 4 deg C	PSEG03	K31	CHRT-26215	10/19/2021	9045D
10/19/2021	Solid	M4284-02	pH	Cool 4 deg C	PSEG03	K31	20878-SOLID	10/19/2021	9045D
10/19/2021	Solid	M4284-03	pH	Cool 4 deg C	PSEG03	K31	30852	10/19/2021	9045D
10/19/2021	Solid	M4284-04	pH	Cool 4 deg C	PSEG03	K31	16943	10/19/2021	9045D

Date/Time 10/19/21 15:40
 Raw Sample Received by: JH (sm)
 Raw Sample Relinquished by: JT (sm)

Date/Time 10/19/21 15:55
 Raw Sample Received by: JT (sm)
 Raw Sample Relinquished by: JH (sm)