


LB116921

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.6834e-004

C= -5.9596e-003

Rho= 0.9989862

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.004	-0.004	9	0.000	9					=4
0.2	0.200	0.192	-0.008	737	0.0 %	737					1
2.5	2.500	2.523	0.023	9424	0.0 %	9424					3
5.0	5.000	4.859	-0.141	18129	0.0 %	18129					4
7.5	7.500	7.826	0.326	29188	0.0 %	29188					-2
10.0	10.000	9.804	-0.196	36557	0.0 %	36557					

LB116921 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 20 Oct 2021 10:58:31					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	90	9	-	0.0000	7470A	PPB	20 Oct 2021 10:59:44	S	Std
0.2 - 1	50.2	737	-	0.2000	7470A	PPB	20 Oct 2021 11:02:01	S	Std
2.5 - 1	52.5	9424	-	2.5000	7470A	PPB	20 Oct 2021 11:04:18	S	Std
5.0 - 1	55	18129	-	5.0000	7470A	PPB	20 Oct 2021 11:06:34	S	Std
7.5 - 1	57.5	29188	-	7.5000	7470A	PPB	20 Oct 2021 11:08:51	S	Std
10.0 - 1	510	36557	-	10.0000	7470A	PPB	20 Oct 2021 11:13:42	S	Std
ICV51 - 1	ICV51	14704	3.9397	-	7470A	PPB	20 Oct 2021 11:16:33	U	SMPL
ICB51 - 1	ICB51	-128	-0.0403	-	7470A	PPB	20 Oct 2021 11:18:48	U	SMPL
CCV03 - 1	CCV03	18570	4.9771	-	7470A	PPB	20 Oct 2021 11:21:03	U	SMPL
CCB03 - 1	CCB03	-124	-0.0392	-	7470A	PPB	20 Oct 2021 11:23:18	U	SMPL
CRA - 1	CRA	666	0.1728	-	7470A	PPB	20 Oct 2021 11:25:33	U	SMPL
HighStd - 1	HighStd	36416	9.7659	-	7470A	PPB	20 Oct 2021 11:27:48	U	SMPL
ChkStd - 1	ChkStd	25357	6.7983	-	7470A	PPB	20 Oct 2021 11:30:03	U	SMPL
PB140032BL - 1	PBW	-10	-0.0086	-	7470A	PPB	20 Oct 2021 11:35:37	U	SMPL
PB140032BS - 1	LCSW	17580	4.7115	-	7470A	PPB	20 Oct 2021 11:37:52	U	SMPL
M4259-02 - 1	WC-1-CS	-121	-0.0384	-	7470A	PPB	20 Oct 2021 11:40:08	U	SMPL
M4259-02DUP - 1	WC-1-CSDUP	-14	-0.0097	-	7470A	PPB	20 Oct 2021 11:42:25	U	SMPL
M4259-02MS - 1	WC-1-CSMS	14815	3.9695	-	7470A	PPB	20 Oct 2021 11:44:42	U	SMPL
M4259-02MSD - 1	WC-1-CSMSD	16205	4.3425	-	7470A	PPB	20 Oct 2021 11:46:59	U	SMPL
M4259-05 - 1	WC-1-BS	-116	-0.0371	-	7470A	PPB	20 Oct 2021 11:49:14	U	SMPL
CCV04 - 1	CCV04	18616	4.9895	-	7470A	PPB	20 Oct 2021 11:51:30	U	SMPL
CCB04 - 1	CCB04	-113	-0.0363	-	7470A	PPB	20 Oct 2021 11:53:46	U	SMPL
M4259-08 - 1	WC-1-BD	-29	-0.0137	-	7470A	PPB	20 Oct 2021 11:56:01	U	SMPL
M4259-11 - 1	WC-1-AS	-8	-0.0081	-	7470A	PPB	20 Oct 2021 11:58:17	U	SMPL
M4274-02 - 1	WC-2-BS	4	-0.0049	-	7470A	PPB	20 Oct 2021 12:00:32	U	SMPL
M4274-05 - 1	WC-2-CS	31	0.0024	-	7470A	PPB	20 Oct 2021 12:02:48	U	SMPL
M4274-08 - 1	WC-2-DS	9	-0.0035	-	7470A	PPB	20 Oct 2021 12:05:04	U	SMPL
PB140033BL - 1	PBW	9	-0.0035	-	7470A	PPB	20 Oct 2021 12:07:20	U	SMPL
PB140033BS - 1	LCSW	14425	3.8648	-	7470A	PPB	20 Oct 2021 12:09:36	U	SMPL
M4255-01 - 1	FMMP-1-20211014	276	0.0681	-	7470A	PPB	20 Oct 2021 12:11:53	U	SMPL
M4255-01DUP - 1	FMMP-1-20211014DUP	205	0.0491	-	7470A	PPB	20 Oct 2021 12:14:10	U	SMPL
M4255-01MS - 1	FMMP-1-20211014MS	13626	3.6504	-	7470A	PPB	20 Oct 2021 12:16:26	U	SMPL
CCV05 - 1	CCV05	17615	4.7208	-	7470A	PPB	20 Oct 2021 12:18:41	U	SMPL
CCB05 - 1	CCB05	-88	-0.0296	-	7470A	PPB	20 Oct 2021 12:20:57	U	SMPL
M4255-01MSD - 1	FMMP-1-20211014MSD	13509	3.6190	-	7470A	PPB	20 Oct 2021 12:23:13	U	SMPL
PB140062BL - 1	PBW	-69	-0.0245	-	7470A	PPB	20 Oct 2021 12:25:28	U	SMPL
M4283-06 - 1	GREEN-TENT-TOTE	-5	-0.0073	-	7470A	PPB	20 Oct 2021 12:29:59	U	SMPL
M4284-09 - 1	76-78	932	0.2441	-	7470A	PPB	20 Oct 2021 12:32:16	U	SMPL
M4284-14 - 1	53-56	9253	2.4770	-	7470A	PPB	20 Oct 2021 12:34:33	U	SMPL
M4286-01 - 1	SW-RR-01-(2.0)	51	0.0077	-	7470A	PPB	20 Oct 2021 12:36:50	U	SMPL
M4286-01DUP - 1	SW-RR-01-(2.0)DUP	97	0.0201	-	7470A	PPB	20 Oct 2021 12:39:07	U	SMPL
M4286-01MS - 1	SW-RR-01-(2.0)MS	13112	3.5125	-	7470A	PPB	20 Oct 2021 12:41:24	U	SMPL
M4286-01MSD - 1	SW-RR-01-(2.0)MSD	13728	3.6778	-	7470A	PPB	20 Oct 2021 12:43:40	U	SMPL
CCV06 - 1	CCV06	19053	5.1067	-	7470A	PPB	20 Oct 2021 12:48:51	U	SMPL
CCB06 - 1	CCB06	-117	-0.0374	-	7470A	PPB	20 Oct 2021 12:51:06	U	SMPL
M4286-02 - 1	SW-RR-01-(2.0)	87	0.0174	-	7470A	PPB	20 Oct 2021 12:53:22	U	SMPL
M4286-03 - 1	SW-RR-03-(SA)	29	0.0018	-	7470A	PPB	20 Oct 2021 12:55:37	U	SMPL
M4286-04 - 1	SW-RR-03-(SA)	-22	-0.0119	-	7470A	PPB	20 Oct 2021 12:57:53	U	SMPL
PB140032TB - 1	PB139976TB	5	-0.0046	-	7470A	PPB	20 Oct 2021 13:00:09	U	SMPL
PB140033TB - 1	PB139989TB	-34	-0.0151	-	7470A	PPB	20 Oct 2021 13:02:25	U	SMPL
CCV07 - 1	CCV07	17163	4.5996	-	7470A	PPB	20 Oct 2021 13:04:41	U	SMPL
CCB07 - 1	CCB07	-104	-0.0339	-	7470A	PPB	20 Oct 2021 13:06:58	U	SMPL
M4259-02LX5 - 1		-6	-0.0076	-	7470A	PPB	20 Oct 2021 13:09:15	U	SMPL
M4259-02A - 1		15396	4.1254	-	7470A	PPB	20 Oct 2021 13:11:32	U	SMPL
M4255-01LX5 - 1		-99	-0.0325	-	7470A	PPB	20 Oct 2021 13:13:48	U	SMPL
M4255-01A - 1		15111	4.0489	-	7470A	PPB	20 Oct 2021 13:16:03	U	SMPL
M4286-01LX5 - 1		-41	-0.0170	-	7470A	PPB	20 Oct 2021 13:18:19	U	SMPL
M4286-01A - 1		15004	4.0202	-	7470A	PPB	20 Oct 2021 13:20:34	U	SMPL
PB140062BS - 1	LCSW	15035	4.0285	-	7470A	PPB	20 Oct 2021 13:46:27	U	SMPL
CCV08 - 1	CCV08	18793	5.0370	-	7470A	PPB	20 Oct 2021 13:48:42	U	SMPL
CCB08 - 1	CCB08	-137	-0.0427	-	7470A	PPB	20 Oct 2021 13:50:58	U	SMPL