



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116934

Slope : 98.6

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	10/20/2021	16:30
2	CAL2	1	Water	NA	NA	20.9	7.00	10/20/2021	16:31
3	CAL3	1	Water	NA	NA	20.9	10.03	10/20/2021	16:32
4	ICV	1	Water	NA	NA	20.8	7.01	10/20/2021	16:35
5	CCV1	1	Water	NA	NA	20.8	2.01	10/20/2021	16:36
6	M4299-01	1	Solid	20.03	20	22.7	6.24	10/20/2021	17:10
7	M4299-01DUP	1	Solid	20.04	20	22.8	6.25	10/20/2021	17:11
8	CCV2	1	Water	NA	NA	20.6	12.03	10/20/2021	17:13

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m4299 s WorkList ID : 153482 Department : Wet-Chemistry Date : 10-20-2021 16:59:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/20/2021	Solid	M4299-01	pH	Cool 4 deg C	PSEG03	K41	NWB-1697	10/20/2021	9045D

Date/Time 10/20/21 17:00
 Raw Sample Received by: JA (CCI)
 Raw Sample Relinquished by: CR SM

Date/Time 10/20/21 18:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JA (CCI)