


LB116937

7470A

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 2.7396e-004

C= -2.5495e-002

Rho= 0.9989664

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.010	0.010	130	0.000	130					-
0.2	0.200	0.234	0.034	948	0.0 %	948					17
2.5	2.500	2.512	0.012	9262	0.0 %	9262					0
5.0	5.000	5.103	0.103	18718	0.0 %	18718					2
7.5	7.500	7.152	-0.348	26199	0.0 %	26199					-5
10.0	10.000	10.189	0.189	37285	0.0 %	37285					2

LB116937 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Oct 2021 10:09:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	58	130	-	0.0000	7470A	PPB	21 Oct 2021 10:14:58	S	Std
0.2 - 1	502	948	-	0.2000	7470A	PPB	21 Oct 2021 10:17:15	S	Std
2.5 - 1	52.5	9262	-	2.5000	7470A	PPB	21 Oct 2021 10:19:32	S	Std
5.0 - 1	55	18718	-	5.0000	7470A	PPB	21 Oct 2021 10:21:48	S	Std
7.5 - 1	57.5	26199	-	7.5000	7470A	PPB	21 Oct 2021 10:24:05	S	Std
10.0 - 1	510	37285	-	10.0000	7470A	PPB	21 Oct 2021 10:26:20	S	Std
ICV52 - 1	ICV52	14160	3.8538	-	7470A	PPB	21 Oct 2021 10:29:10	U	SMPL
ICB52 - 1	ICB52	35	-0.0159	-	7470A	PPB	21 Oct 2021 10:31:25	U	SMPL
CCV09 - 1	CCV09	19101	5.2074	-	7470A	PPB	21 Oct 2021 10:33:39	U	SMPL
CCB09 - 1	CCB09	39	-0.0148	-	7470A	PPB	21 Oct 2021 10:35:54	U	SMPL
CRA - 1	CRA	931	0.2296	-	7470A	PPB	21 Oct 2021 10:38:09	U	SMPL
HighStd - 1	HighStd	36861	10.0730	-	7470A	PPB	21 Oct 2021 10:40:24	U	SMPL
ChkStd - 1	ChkStd	25029	6.8315	-	7470A	PPB	21 Oct 2021 10:42:39	U	SMPL
PB140082BL - 1	PBW	137	0.0120	-	7470A	PPB	21 Oct 2021 10:44:56	U	SMPL
PB140082BS - 1	LCSW	14743	4.0135	-	7470A	PPB	21 Oct 2021 10:47:12	U	SMPL
M4285-02 - 1	MH-EA-(2)	143	0.0137	-	7470A	PPB	21 Oct 2021 10:49:29	U	SMPL
M4285-05 - 1	MH-EB-(2)	213	0.0329	-	7470A	PPB	21 Oct 2021 10:51:46	U	SMPL
M4285-08 - 1	MH-FA-(2)	201	0.0296	-	7470A	PPB	21 Oct 2021 10:54:03	U	SMPL
M4285-11 - 1	MH-GA-(2)	357	0.0723	-	7470A	PPB	21 Oct 2021 10:56:20	U	SMPL
M4285-14 - 1	MH-GB-(2)	342	0.0682	-	7470A	PPB	21 Oct 2021 10:58:35	U	SMPL
CCV10 - 1	CCV10	19191	5.2321	-	7470A	PPB	21 Oct 2021 11:00:50	U	SMPL
CCB10 - 1	CCB10	52	-0.0112	-	7470A	PPB	21 Oct 2021 11:03:05	U	SMPL
M4285-17 - 1	MH-FB-(2)	301	0.0570	-	7470A	PPB	21 Oct 2021 11:05:20	U	SMPL
M4287-03 - 1	EXC-SOIL-A	268	0.0479	-	7470A	PPB	21 Oct 2021 11:07:36	U	SMPL
M4287-07 - 1	EXC-SOIL-B	268	0.0479	-	7470A	PPB	21 Oct 2021 11:09:51	U	SMPL
M4287-07DUP - 1	EXC-SOIL-BDUP	265	0.0471	-	7470A	PPB	21 Oct 2021 11:12:06	U	SMPL
M4287-07MS - 1	EXC-SOIL-BMS	13980	3.8045	-	7470A	PPB	21 Oct 2021 11:14:22	U	SMPL
M4287-07MSD - 1	EXC-SOIL-BMSD	14493	3.9450	-	7470A	PPB	21 Oct 2021 11:16:38	U	SMPL
PB140082TB - 1	PB140039TB	180	0.0238	-	7470A	PPB	21 Oct 2021 11:18:55	U	SMPL
M4287-07LX5 - 1		163	0.0192	-	7470A	PPB	21 Oct 2021 11:21:11	U	SMPL
M4287-07A - 1		14399	3.9193	-	7470A	PPB	21 Oct 2021 11:23:29	U	SMPL
CCV11 - 1	CCV11	19577	5.3379	-	7470A	PPB	21 Oct 2021 11:25:44	U	SMPL
CCB11 - 1	CCB11	47	-0.0126	-	7470A	PPB	21 Oct 2021 11:28:00	U	SMPL