

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB116948

ANALYST: ketankumar
 SUPERVISOR REVIEW BY: apatel
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2843
IODINE SOLUTION .025N 1L	W2864
Starch Solution, 4L	W2703

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB140081BL		1	5.00	50	2.00	0.00	1.88	1.88	0.12	0.00	0.00	10/20/2021	17:32
2	M4280-01		1	5.05	50	2.00	1.88	3.72	1.84	0.16	0.04	3.17	10/20/2021	17:35
3	M4280-01DUP		1	5.09	50	2.00	3.72	5.58	1.86	0.14	0.02	1.57	10/20/2021	17:38
4	M4280-03		1	5.08	50	2.00	5.58	7.40	1.82	0.18	0.06	4.72	10/20/2021	17:40
5	M4280-05		1	5.07	50	2.00	7.40	9.20	1.80	0.20	0.08	6.31	10/20/2021	17:42
6	M4285-02		1	5.06	50	2.00	0.00	1.84	1.84	0.16	0.04	3.16	10/20/2021	17:45
7	M4285-05		1	5.05	50	2.00	1.84	3.70	1.86	0.14	0.02	1.58	10/20/2021	17:48
8	M4285-08		1	5.04	50	2.00	3.70	5.52	1.82	0.18	0.06	4.76	10/20/2021	17:50
9	M4285-11		1	5.03	50	2.00	5.52	7.38	1.86	0.14	0.02	1.59	10/20/2021	17:52
10	M4285-14		1	5.07	50	2.00	0.00	1.86	1.86	0.14	0.02	1.58	10/20/2021	17:55
11	M4285-17		1	5.09	50	2.00	1.86	3.70	1.84	0.16	0.04	3.14	10/20/2021	17:58
12	M4287-03		1	5.08	50	2.00	3.70	5.50	1.80	0.20	0.08	6.30	10/20/2021	18:00

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume

Analysis Method: 9034

ANALYST: ketankumar

Parameter: Reactive Sulfide

SUPERVISOR REVIEW BY: apatel

Run Number: LB116948

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2843
IODINE SOLUTION .025N 1L	W2864
Starch Solution, 4L	W2703

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
13	M4287-07		1	5.07	50	2.00	5.50	7.28	1.78	0.22	0.10	7.89	10/20/2021	18:02

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume