



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116952

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/21/2021	17:15
2	CAL2	1	Water	NA	NA	20.3	7.00	10/21/2021	17:16
3	CAL3	1	Water	NA	NA	20.4	10.03	10/21/2021	17:18
4	ICV	1	Water	NA	NA	20.3	6.99	10/21/2021	17:20
5	CCV1	1	Water	NA	NA	20.5	2.01	10/21/2021	17:22
6	M4312-04	1	Water	NA	NA	21.5	5.40	10/21/2021	18:10
7	M4312-04DUP	1	Water	NA	NA	21.5	5.41	10/21/2021	18:11
8	M4312-05	1	Water	NA	NA	21.3	4.96	10/21/2021	18:15
9	M4312-06	1	Water	NA	NA	21.9	7.60	10/21/2021	18:19
10	CCV2	1	Water	NA	NA	20.6	12.02	10/21/2021	18:20

VB 116952

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M4317 V

Department : Wet-Chemistry

WorkList ID : 153524

Date : 10-21-2021 17:04:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/21/2021	Water	M4312-04	Corrosivity	Cool 4 deg C	PSEG03	K41	5261	10/21/2021	9040C
10/21/2021	Water	M4312-05	Corrosivity	Cool 4 deg C	PSEG03	K41	comp	10/21/2021	9040C
10/21/2021	Water	M4312-06	Corrosivity	Cool 4 deg C	PSEG03	K41	5234	10/21/2021	9040C

Date/Time 10/21/21 17:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 10/21/21 18:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]