



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116953

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/21/2021	17:15
2	CAL2	1	Water	NA	NA	20.3	7.00	10/21/2021	17:16
3	CAL3	1	Water	NA	NA	20.4	10.03	10/21/2021	17:18
4	ICV	1	Water	NA	NA	20.3	6.99	10/21/2021	17:20
5	CCV1	1	Water	NA	NA	20.5	2.01	10/21/2021	17:22
6	M4313-07	1	Water	NA	NA	20.5	7.53	10/21/2021	17:28
7	M4313-07DUP	1	Water	NA	NA	20.6	7.54	10/21/2021	17:29
8	CCV2	1	Water	NA	NA	20.6	12.02	10/21/2021	17:30

VB 116953

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH W M4316 WorkList ID : 153525 Department : Wet-Chemistry Date : 10-21-2021 17:05:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/21/2021	Water	M4313-07	pH	Cool 4 deg C	PSEG03	D41	1-3-TOTES	10/21/2021	9040C

Date/Time 10/21/21 17:10
 Raw Sample Received by: JD 100C
 Raw Sample Relinquished by: Cal SM

Date/Time 10/21/21 18:30
 Raw Sample Received by: Cal SM
 Raw Sample Relinquished by: JD 100C