


LB116964

SFAM01.1

INSTRUMENT ID: LV1

 μ Abs.:

38496



Linear

A= 0.0000e+000

B= 2.5871e-004 slope

C= 1.9497e-002 y-intercept

Rho= 0.9998758

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.040	0.040	79	0.000	79					=
0.05	0.050					205	0				17
0.20	0.200	0.233	0.033	827	0.0 %	827					-4
2.50	2.500	2.401	-0.099	9206	0.0 %	9206					-1
5.00	5.000	4.967	-0.033	19124	0.0 %	19124					1
7.50	7.500	7.579	0.079	29221	0.0 %	29221					0
10.0	10.000	9.979	-0.021	38496	0.0 %	38496					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	79	-		0 SFAM01.1	PPB	10/22/2021 11:02	Std
	0.2 S01	827	-		0.2 SFAM01.1	PPB	10/22/2021 11:05	Std
	2.5 S02	9206	-		2.5 SFAM01.1	PPB	10/22/2021 11:07	Std
	5 S03	19124	-		5 SFAM01.1	PPB	10/22/2021 11:09	Std
	7.5 S04	29221	-		7.5 SFAM01.1	PPB	10/22/2021 11:16	Std
	10 S05	38496	-		10 SFAM01.1	PPB	10/22/2021 11:21	Std
ICV084	ICV084	16283	4.2321	-	SFAM01.1	PPB	10/22/2021 11:24	SMPL
ICB084	ICB084	-53	0.0058	-	SFAM01.1	PPB	10/22/2021 11:26	SMPL
CCV019	CCV019	18564	4.8222	-	SFAM01.1	PPB	10/22/2021 11:28	SMPL
CCB019	CCB019	-51	0.0063	-	SFAM01.1	PPB	10/22/2021 11:31	SMPL
PB140123BL	PBW123	22	0.0252	-	SFAM01.1	PPB	10/22/2021 11:33	SMPL
M4185-16	COB86	24	0.0257	-	SFAM01.1	PPB	10/22/2021 11:35	SMPL
PB140124BL	PBW124	60	0.035	-	SFAM01.1	PPB	10/22/2021 11:37	SMPL
M4201-02	J0AA4	32	0.0278	-	SFAM01.1	PPB	10/22/2021 11:40	SMPL
M4201-03	JDGH1	30	0.0273	-	SFAM01.1	PPB	10/22/2021 11:42	SMPL
M4201-04	JDGG8	48	0.0319	-	SFAM01.1	PPB	10/22/2021 11:44	SMPL
M4201-05	JDGG9	26	0.0262	-	SFAM01.1	PPB	10/22/2021 11:47	SMPL
M4201-06	JDGG3	24	0.0257	-	SFAM01.1	PPB	10/22/2021 11:49	SMPL
M4201-07	JDGG6	70	0.0376	-	SFAM01.1	PPB	10/22/2021 11:51	SMPL
M4201-08	JDGJ0	15	0.0234	-	SFAM01.1	PPB	10/22/2021 11:53	SMPL
M4201-09	JDGG6D	22	0.0252	-	SFAM01.1	PPB	10/22/2021 11:56	SMPL
M4201-10	JDGG6S	4214	1.1097	-	SFAM01.1	PPB	10/22/2021 11:58	SMPL
PB140125BL	PBW125	12	0.0226	-	SFAM01.1	PPB	10/22/2021 12:00	SMPL
M4203-01	J0AA3	2	0.02	-	SFAM01.1	PPB	10/22/2021 12:02	SMPL
M4203-02	JDGH2	-2	0.019	-	SFAM01.1	PPB	10/22/2021 12:05	SMPL
M4203-03	JDGG7	21	0.0249	-	SFAM01.1	PPB	10/22/2021 12:07	SMPL
M4203-04	JDGH0	25	0.026	-	SFAM01.1	PPB	10/22/2021 12:09	SMPL
M4203-05	JDGH9	3	0.0203	-	SFAM01.1	PPB	10/22/2021 12:11	SMPL
M4203-13	JDGH0D	6	0.021	-	SFAM01.1	PPB	10/22/2021 12:14	SMPL
M4203-14	JDGH0S	4466	1.1749	-	SFAM01.1	PPB	10/22/2021 12:16	SMPL
CCV020	CCV020	18155	4.7164	-	SFAM01.1	PPB	10/22/2021 12:18	SMPL
CCB020	CCB020	-31	0.0115	-	SFAM01.1	PPB	10/22/2021 12:21	SMPL
PB140126BL	PBW126	6	0.021	-	SFAM01.1	PPB	10/22/2021 12:23	SMPL
M4225-01	MH1S77	134	0.0542	-	SFAM01.1	PPB	10/22/2021 12:25	SMPL
M4225-02	MH1S78	508	0.1509	-	SFAM01.1	PPB	10/22/2021 12:27	SMPL
M4225-03	MH1S79	65	0.0363	-	SFAM01.1	PPB	10/22/2021 12:30	SMPL
M4225-04	MH1S80	-3	0.0187	-	SFAM01.1	PPB	10/22/2021 12:32	SMPL
M4225-05	MH1S81	-2	0.019	-	SFAM01.1	PPB	10/22/2021 12:34	SMPL
M4225-06	MH1S82	356	0.1116	-	SFAM01.1	PPB	10/22/2021 12:36	SMPL
M4225-07	MH1S82D	351	0.1103	-	SFAM01.1	PPB	10/22/2021 12:39	SMPL
M4225-08	MH1S82S	3929	1.036	-	SFAM01.1	PPB	10/22/2021 12:41	SMPL
M4225-09	MH1S83	-7	0.0177	-	SFAM01.1	PPB	10/22/2021 12:43	SMPL
M4225-10	MH1S84	-12	0.0164	-	SFAM01.1	PPB	10/22/2021 12:45	SMPL
M4225-11	MH1S85	7	0.0213	-	SFAM01.1	PPB	10/22/2021 12:48	SMPL
M4225-12	MH1S86	-14	0.0159	-	SFAM01.1	PPB	10/22/2021 12:50	SMPL
M4225-13	MH1S87	-10	0.0169	-	SFAM01.1	PPB	10/22/2021 12:52	SMPL

INSTRUMENT ID : CV1

M4225-14	MH1S88	28	0.0267 -	SFAM01.1	PPB	10/22/2021 12:55	SMPL
M4225-15	MH1S89	-16	0.0154 -	SFAM01.1	PPB	10/22/2021 12:57	SMPL
M4225-16	MH1S90	678	0.1949 -	SFAM01.1	PPB	10/22/2021 12:59	SMPL
M4225-17	MH1S91	-13	0.0161 -	SFAM01.1	PPB	10/22/2021 13:01	SMPL
M4225-18	MH1S92	14	0.0231 -	SFAM01.1	PPB	10/22/2021 13:04	SMPL
M4225-19	MH1S93	-6	0.0179 -	SFAM01.1	PPB	10/22/2021 13:06	SMPL
CCV021	CCV021	17946	4.6624 -	SFAM01.1	PPB	10/22/2021 13:08	SMPL
CCB021	CCB021	-21	0.0141 -	SFAM01.1	PPB	10/22/2021 13:10	SMPL
M4225-20	MH1S94	3	0.0203 -	SFAM01.1	PPB	10/22/2021 13:13	SMPL
M4225-21	MH1S95	42	0.0304 -	SFAM01.1	PPB	10/22/2021 13:15	SMPL
M4225-22	MH1S96	0	0.0195 -	SFAM01.1	PPB	10/22/2021 13:17	SMPL
PB140127BL	PBW127	0	0.0195 -	SFAM01.1	PPB	10/22/2021 13:19	SMPL
M4228-01	MH1S97	-29	0.012 -	SFAM01.1	PPB	10/22/2021 13:22	SMPL
M4228-02	MH1S98	-25	0.013 -	SFAM01.1	PPB	10/22/2021 13:24	SMPL
M4228-03	MH1T09	-30	0.0117 -	SFAM01.1	PPB	10/22/2021 13:26	SMPL
M4228-04	MH1T11	-25	0.013 -	SFAM01.1	PPB	10/22/2021 13:28	SMPL
M4228-05	MH1T13	-19	0.0146 -	SFAM01.1	PPB	10/22/2021 13:31	SMPL
M4228-06	MH1T15	-35	0.0104 -	SFAM01.1	PPB	10/22/2021 13:33	SMPL
M4228-07	MH1T17	-15	0.0156 -	SFAM01.1	PPB	10/22/2021 13:35	SMPL
M4228-08	MH1T19	-19	0.0146 -	SFAM01.1	PPB	10/22/2021 13:38	SMPL
M4228-09	MH1T21	-9	0.0172 -	SFAM01.1	PPB	10/22/2021 13:40	SMPL
M4228-10	MH1T23	-1	0.0192 -	SFAM01.1	PPB	10/22/2021 13:42	SMPL
M4228-11	MH1T25	-14	0.0159 -	SFAM01.1	PPB	10/22/2021 13:44	SMPL
M4228-12	MH1T27	-19	0.0146 -	SFAM01.1	PPB	10/22/2021 13:47	SMPL
M4228-13	MH1T29	-15	0.0156 -	SFAM01.1	PPB	10/22/2021 13:49	SMPL
M4228-14	MH1T31	-1	0.0192 -	SFAM01.1	PPB	10/22/2021 13:51	SMPL
M4228-15	MH1T33	86	0.0417 -	SFAM01.1	PPB	10/22/2021 13:53	SMPL
M4228-16	MH1T35	266	0.0883 -	SFAM01.1	PPB	10/22/2021 13:56	SMPL
CCV022	CCV022	18517	4.8101 -	SFAM01.1	PPB	10/22/2021 13:58	SMPL
CCB022	CCB022	-21	0.0141 -	SFAM01.1	PPB	10/22/2021 14:00	SMPL
M4228-17	MH1T37	137	0.0549 -	SFAM01.1	PPB	10/22/2021 14:03	SMPL
M4228-18	MH1T39	12	0.0226 -	SFAM01.1	PPB	10/22/2021 14:05	SMPL
M4228-19	MH1T41	188	0.0681 -	SFAM01.1	PPB	10/22/2021 14:07	SMPL
M4228-20	MH1T43	-8	0.0174 -	SFAM01.1	PPB	10/22/2021 14:09	SMPL
M4228-21	MH1T43D	-1	0.0192 -	SFAM01.1	PPB	10/22/2021 14:12	SMPL
M4228-22	MH1T43S	4525	1.1902 -	SFAM01.1	PPB	10/22/2021 14:14	SMPL
CCV023	CCV023	18104	4.7032 -	SFAM01.1	PPB	10/22/2021 14:19	SMPL
CCB023	CCB023	-3	0.0187 -	SFAM01.1	PPB	10/22/2021 14:21	SMPL