



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116970

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/22/2021	09:15
2	M4283-06		1	10	100.00	>212.0	>212.0	10/22/2021	09:40
3	M4283-06DUP		1	10	100.00	>212.0	>212.0	10/22/2021	10:05
4	M4284-05		1	12	100.00	>212.0	>212.0	10/22/2021	10:35
5	M4284-09		1	15	100.00	>212.0	>212.0	10/22/2021	10:58
6	M4284-14		1	14	100.00	>212.0	>212.0	10/22/2021	11:20
7	M4284-15		1	13	100.00	>212.0	>212.0	10/22/2021	11:40
8	M4284-16		1	12	100.00	>212.0	>212.0	10/22/2021	12:05
9	M4284-17		1	15	100.00	>212.0	>212.0	10/22/2021	12:32
10	M4299-06		1	11	100.00	>212.0	>212.0	10/22/2021	12:58
11	M4312-04		1	13	100.00	>212.0	>212.0	10/22/2021	13:20
12	M4312-05		1	12	100.00	>212.0	>212.0	10/22/2021	13:45
13	M4312-06		1	15	100.00	>212.0	>212.0	10/22/2021	14:10
14	M4313-07		1	12	100.00	>212.0	>212.0	10/22/2021	14:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB116970

WorkList Name : fp102021 WorkList ID : 153486 Department : Wet-Chemistry Date : 10-20-2021 17:06:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/26/2021	Water	M4283-06	Flash Point	Cool 4 deg C	PSEG03	K31	GREEN-TENT-TOTE	10/19/2021	1010B
10/26/2021	Water	M4284-05	Flash Point	Cool 4 deg C	PSEG03	K31	30851-WATER	10/19/2021	1010B
10/26/2021	Water	M4284-09	Flash Point	Cool 4 deg C	PSEG03	K31	76-78	10/19/2021	1010B
10/26/2021	Water	M4284-14	Flash Point	Cool 4 deg C	PSEG03	K31	53-56	10/19/2021	1010B
10/26/2021	Water	M4284-15	Flash Point	Cool 4 deg C	PSEG03	K31	30848	10/19/2021	1010B
10/26/2021	Water	M4284-16	Flash Point	Cool 4 deg C	PSEG03	K31	30849	10/19/2021	1010B
10/26/2021	Water	M4284-17	Flash Point	Cool 4 deg C	PSEG03	K31	20890	10/19/2021	1010B
11/02/2021	Water	M4299-06	Flash Point	Cool 4 deg C	PSEG03	K41	1694-1696	10/20/2021	1010B
10/28/2021	Water	M4312-04	Flash Point	Cool 4 deg C	PSEG03	K41	5261	10/21/2021	1010B
10/28/2021	Water	M4312-05	Flash Point	Cool 4 deg C	PSEG03	K41	comp	10/21/2021	1010B
10/28/2021	Water	M4312-06	Flash Point	Cool 4 deg C	PSEG03	K41	5234	10/21/2021	1010B
10/26/2021	Water	M4313-07	Flash Point	Cool 4 deg C	PSEG03	D41	1-3-TOTES	10/21/2021	1010B

Date/Time 10/22/21 08:45
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RP (G&L)

Date/Time 10/22/21 14:40
 Raw Sample Received by: RP (G&L)
 Raw Sample Relinquished by: KS (WC)