


LB116977

245.1

INSTRUMENT ID: CV4



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.016	-0.016	64	0.000	64					-6
0.05	0.050	0.047	-0.003	257	0.0 %	257	0				-7
0.2	0.200	0.187	-0.013	688	0.0 %	688					-5
2.5	2.500	2.366	-0.134	7380	0.0 %	7380					7
5.0	5.000	5.341	0.341	16520	0.0 %	16520					-1
7.5	7.500	7.405	-0.095	22861	0.0 %	22861					-1
10.0	10.000	9.920	-0.080	30587	0.0 %	30587					

LB116977 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 22 Oct 2021 14:39:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	64	-	0.0000	245.1	PPB	22 Oct 2021 14:40:02	S	Std
0.05 - 1	50.05	257	-	0.0500	245.1	PPB	22 Oct 2021 14:42:19	S	Std
0.2 - 1	50.2	688	-	0.2000	245.1	PPB	22 Oct 2021 14:44:37	S	Std
2.5 - 1	52.5	7380	-	2.5000	245.1	PPB	22 Oct 2021 14:46:54	S	Std
5.0 - 1	55	16520	-	5.0000	245.1	PPB	22 Oct 2021 14:52:55	S	Std
7.5 - 1	57.5	22861	-	7.5000	245.1	PPB	22 Oct 2021 14:57:44	S	Std
10.0 - 1	510	30587	-	10.0000	245.1	PPB	22 Oct 2021 15:02:36	S	Std
ICV54 - 1	ICV54	12410	4.0030	-245.1		PPB	22 Oct 2021 15:09:23	U	SMPL
ICB54 - 1	ICB54	23	-0.0291	-245.1		PPB	22 Oct 2021 15:11:38	U	SMPL
CCV24 - 1	CCV24	16570	5.3571	-245.1		PPB	22 Oct 2021 15:13:53	U	SMPL
CCB24 - 1	CCB24	17	-0.0310	-245.1		PPB	22 Oct 2021 15:16:08	U	SMPL
CRA - 1	CRA	245	0.0432	-245.1		PPB	22 Oct 2021 15:20:58	U	SMPL
HighStd - 1	HighStd	30965	10.0428	-245.1		PPB	22 Oct 2021 15:23:12	U	SMPL
ChkStd - 1	ChkStd	22432	7.2652	-245.1		PPB	22 Oct 2021 15:28:13	U	SMPL
PB140148BL - 1	PBW	-104	-0.0704	-245.1		PPB	22 Oct 2021 15:30:28	U	SMPL
M4245-02 - 1	EFF-WW	90	-0.0072	-245.1		PPB	22 Oct 2021 15:35:00	U	SMPL
M4270-01 - 1	32220008-3	24	-0.0287	-245.1		PPB	22 Oct 2021 15:37:17	U	SMPL
M4270-01DUP - 1	32220008-3DUP	45	-0.0219	-245.1		PPB	22 Oct 2021 15:39:34	U	SMPL
M4270-01MS - 1	32220008-3MS	14026	4.5290	-245.1		PPB	22 Oct 2021 15:41:50	U	SMPL
M4270-01MSD - 1	32220008-3MSD	14648	4.7315	-245.1		PPB	22 Oct 2021 15:44:06	U	SMPL
CCV25 - 1	CCV25	16156	5.2224	-245.1		PPB	22 Oct 2021 15:46:21	U	SMPL
CCB25 - 1	CCB25	-13	-0.0408	-245.1		PPB	22 Oct 2021 15:48:36	U	SMPL
PB14148BS - 1	LCSW	13719	4.4291	-245.1		PPB	22 Oct 2021 15:50:51	U	SMPL
M4270-01LX5 - 1		-4	-0.0378	-245.1		PPB	22 Oct 2021 15:53:07	U	SMPL
M4270-01A - 1		14203	4.5866	-245.1		PPB	22 Oct 2021 15:55:23	U	SMPL
CCV26 - 1	CCV26	16864	5.4528	-245.1		PPB	22 Oct 2021 15:57:39	U	SMPL
CCB26 - 1	CCB26	-16	-0.0417	-245.1		PPB	22 Oct 2021 15:59:56	U	SMPL