

LB116984


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

3743E



A= 0.0000e+000

B= 2.6685e-004

C= 2.8009e-002

Rho= 0.9996932

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.051	0.051	85	0.000	85					12
0.2	0.200	0.224	0.024	733	0.0 %	733					-6
2.5	2.500	2.343	-0.157	8674	0.0 %	8674					3
5.0	5.000	5.132	0.132	19127	0.0 %	19127					-1
7.5	7.500	7.433	-0.067	27750	0.0 %	27750					0
10.0	10.000	10.018	0.018	37438	0.0 %	37438					

LB116984 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Oct 2021 16:11:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	80	85	-	0.0000	7470A	PPB	22 Oct 2021 16:12:25	S	Std
0.2 - 1	50.2	733	-	0.2000	7470A	PPB	22 Oct 2021 16:14:42	S	Std
2.5 - 1	52.5	8674	-	2.5000	7470A	PPB	22 Oct 2021 16:16:59	S	Std
5.0 - 1	55	19127	-	5.0000	7470A	PPB	22 Oct 2021 16:19:16	S	Std
7.5 - 1	57.5	27750	-	7.5000	7470A	PPB	22 Oct 2021 16:21:32	S	Std
10.0 - 1	510	37438	-	10.0000	7470A	PPB	22 Oct 2021 16:23:48	S	Std
ICV55 - 1	ICV55	13453	3.6179	-7470A	PPB	22 Oct 2021 16:26:41	U	SMPL	
ICB55 - 1	ICB55	-33	0.0192	-7470A	PPB	22 Oct 2021 16:28:56	U	SMPL	
CCV27 - 1	CCV27	17892	4.8024	-7470A	PPB	22 Oct 2021 16:31:11	U	SMPL	
CCB27 - 1	CCB27	-38	0.0179	-7470A	PPB	22 Oct 2021 16:33:26	U	SMPL	
CRA - 1	CRA	698	0.2143	-7470A	PPB	22 Oct 2021 16:35:41	U	SMPL	
HighStd - 1	HighStd	36081	9.6560	-7470A	PPB	22 Oct 2021 16:37:56	U	SMPL	
ChkStd - 1	ChkStd	24431	6.5473	-7470A	PPB	22 Oct 2021 16:40:12	U	SMPL	
PB140111BL - 1	PBW	-105	-0.0000	-7470A	PPB	22 Oct 2021 16:42:28	U	SMPL	
PB140111BS - 1	LCSW	14335	3.8532	-7470A	PPB	22 Oct 2021 16:44:45	U	SMPL	
M4290-02 - 1	TP-1	-42	0.0168	-7470A	PPB	22 Oct 2021 16:47:01	U	SMPL	
M4290-05 - 1	TH-BS	27	0.0352	-7470A	PPB	22 Oct 2021 16:49:18	U	SMPL	
M4290-08 - 1	TH-BD	45	0.0400	-7470A	PPB	22 Oct 2021 16:51:35	U	SMPL	
M4290-11 - 1	TH-BBS	68	0.0462	-7470A	PPB	22 Oct 2021 16:53:52	U	SMPL	
M4290-14 - 1	TH-BBD	56	0.0430	-7470A	PPB	22 Oct 2021 16:56:07	U	SMPL	
CCV28 - 1	CCV28	18579	4.9857	-7470A	PPB	22 Oct 2021 16:58:23	U	SMPL	
CCB28 - 1	CCB28	-30	0.0200	-7470A	PPB	22 Oct 2021 17:00:38	U	SMPL	
M4294-02 - 1	WC-2-FS	55	0.0427	-7470A	PPB	22 Oct 2021 17:02:53	U	SMPL	
M4294-05 - 1	WC-2-ES	61	0.0443	-7470A	PPB	22 Oct 2021 17:05:08	U	SMPL	
M4294-08 - 1	WC-2-GS	71	0.0470	-7470A	PPB	22 Oct 2021 17:07:24	U	SMPL	
M4294-11 - 1	WC-2-GD	75	0.0480	-7470A	PPB	22 Oct 2021 17:09:40	U	SMPL	
M4294-14 - 1	WC-2-HS	77	0.0486	-7470A	PPB	22 Oct 2021 17:11:56	U	SMPL	
M4299-01 - 1	NWB-1697	71	0.0470	-7470A	PPB	22 Oct 2021 17:14:12	U	SMPL	
M4300-02 - 1	TH-CS	54	0.0424	-7470A	PPB	22 Oct 2021 17:16:28	U	SMPL	
M4300-05 - 1	TH-CD	66	0.0456	-7470A	PPB	22 Oct 2021 17:18:45	U	SMPL	
M4300-08 - 1	TH-CCS	57	0.0432	-7470A	PPB	22 Oct 2021 17:21:03	U	SMPL	
M4300-11 - 1	TH-CCD	62	0.0446	-7470A	PPB	22 Oct 2021 17:23:18	U	SMPL	
CCV29 - 1	CCV29	18615	4.9953	-7470A	PPB	22 Oct 2021 17:25:34	U	SMPL	
CCB29 - 1	CCB29	-27	0.0208	-7470A	PPB	22 Oct 2021 17:27:49	U	SMPL	
M4300-11DUP - 1	TH-CCDDUP	46	0.0403	-7470A	PPB	22 Oct 2021 17:30:04	U	SMPL	
M4300-11MS - 1	TH-CCDMS	14671	3.9429	-7470A	PPB	22 Oct 2021 17:32:20	U	SMPL	
M4300-11MSD - 1	TH-CCDMSD	15072	4.0499	-7470A	PPB	22 Oct 2021 17:34:35	U	SMPL	
CCV30 - 1	CCV30	18700	5.0180	-7470A	PPB	22 Oct 2021 17:52:43	U	SMPL	
CCB30 - 1	CCB30	9	0.0304	-7470A	PPB	22 Oct 2021 17:54:59	U	SMPL	
PB140111TB - 1	PB140086TB	27	0.0352	-7470A	PPB	22 Oct 2021 17:59:30	U	SMPL	
M4300-11LX5 - 1		-11	0.0251	-7470A	PPB	22 Oct 2021 18:01:45	U	SMPL	
M4300-11A - 1		14113	3.7940	-7470A	PPB	22 Oct 2021 18:04:01	U	SMPL	
CCV31 - 1	CCV31	18418	4.9428	-7470A	PPB	22 Oct 2021 18:10:50	U	SMPL	
CCB31 - 1	CCB31	-21	0.0224	-7470A	PPB	22 Oct 2021 18:13:07	U	SMPL	