



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116985

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/22/2021	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	10/22/2021	16:16
3	CAL3	1	Water	NA	NA	20.3	10.03	10/22/2021	16:19
4	ICV	1	Water	NA	NA	20.4	7.02	10/22/2021	16:20
5	CCV1	1	Water	NA	NA	20.4	2.01	10/22/2021	16:22
6	M4326-04	1	Water	NA	NA	22.7	6.40	10/22/2021	17:20
7	M4326-04DUP	1	Water	NA	NA	22.8	6.41	10/22/2021	17:21
8	M4326-05	1	Water	NA	NA	22.8	7.97	10/22/2021	17:23
9	CCV2	1	Water	NA	NA	20.8	12.03	10/22/2021	17:25

WORKLIST(Hardcopy Internal Chain)

VB 116985

WorkList Name : CORROSIONITY M4331 w WorkList ID : 153574 Department : Wet-Chemistry Date : 10-22-2021 15:57:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/22/2021	Water	M4326-04	Corrosivity	Cool 4 deg C	PSEG03	K51	17066	10/22/2021	9040C
10/22/2021	Water	M4326-05	Corrosivity	Cool 4 deg C	PSEG03	K51	17060	10/22/2021	9040C

Date/Time 10/22/21 16:00
Raw Sample Received by: JP (Cm)
Raw Sample Relinquished by: JP (Cm)

Date/Time 10/22/21 18:00
Raw Sample Received by: JP (Cm)
Raw Sample Relinquished by: JP (Cm)